

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
 Data File : BN007797.D
 Acq On : 21 Sep 2019 12:02
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 01:57:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.358	10.5	123	0.00
3	n-Nitrosodimethylamine	0.400	0.606	-51.5#	115	0.00
4 S	2-Fluorophenol	0.400	0.277	30.8#	92	0.00
5 S	Phenol-d6	0.400	0.274	31.5#	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	117	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.400	0.318	20.5	92	0.00
9	Naphthalene	0.400	0.398	0.5	113	-0.01
10	Hexachlorobutadiene	0.400	0.391	2.3	110	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	113	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	114	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.232	42.0#	72	-0.01
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	108	-0.01
16	Acenaphthylene	0.400	0.330	17.5	96	-0.01
17	Acenaphthene	0.400	0.355	11.3	101	-0.01
18	Fluorene	0.400	0.353	11.8	100	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	99	-0.01
21	Hexachlorobenzene	0.400	0.411	-2.7	103	-0.01
22	Pentachlorophenol	0.400	0.227	43.3#	64	-0.01
23	Phenanthrene	0.400	0.400	0.0	100	-0.01
24	Anthracene	0.400	0.360	10.0	92	-0.01
25 SURR	Fluoranthene-d10	0.400	0.373	6.8	93	-0.01
26	Fluoranthene	0.400	0.381	4.8	95	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	95	-0.01
28	Pvrene	0.400	0.386	3.5	95	-0.02
29 S	Terphenyl-d14	0.400	0.388	3.0	95	-0.01
30	Benzo(a)anthracene	0.400	0.339	15.3	83	-0.01
31	Chrysene	0.400	0.409	-2.2	99	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.194	51.5#	51	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	98	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	93	-0.02
35	Benzo(b)fluoranthene	0.400	0.369	7.8	91	-0.02
36	Benzo(k)fluoranthene	0.400	0.428	-7.0	104	-0.02
37 C	Benzo(a)pyrene	0.400	0.355	11.3	87	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	96	-0.03
39	Benzo(a,h,i)perylene	0.400	0.371	7.3	90	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
Data File : BN007797.D
Acq On : 21 Sep 2019 12:02
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 23 01:57:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				