

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
 Data File : BN007797.D
 Acq On : 21 Sep 2019 12:02
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 01:57:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.446	0.399	10.5	123	0.00
3	n-Nitrosodimethylamine	0.204	0.309	-51.5#	115	0.00
4 S	2-Fluorophenol	1.377	0.952	30.9#	92	0.00
5 S	Phenol-d6	1.962	1.184	39.7#	88	0.00
6	bis(2-Chloroethyl)ether	1.133	1.010	10.9	117	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.350	0.278	20.6	92	0.00
9	Naphthalene	1.123	1.117	0.5	113	-0.01
10	Hexachlorobutadiene	0.236	0.231	2.1	110	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.651	3.4	113	0.00
12	2-Methylnaphthalene	0.759	0.745	1.8	114	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.126	42.2#	72	-0.01
15 S	2-Fluorobiphenyl	1.647	1.589	3.5	108	-0.01
16	Acenaphthylene	2.121	1.753	17.4	96	-0.01
17	Acenaphthene	1.368	1.213	11.3	101	-0.01
18	Fluorene	1.753	1.547	11.8	100	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.243	0.238	2.1	99	-0.01
21	Hexachlorobenzene	0.265	0.273	-3.0	103	-0.01
22	Pentachlorophenol	0.098	0.055	43.9#	64	-0.01
23	Phenanthrene	1.246	1.246	0.0	100	-0.01
24	Anthracene	1.125	1.013	10.0	92	-0.01
25 SURR	Fluoranthene-d10	1.348	1.256	6.8	93	-0.01
26	Fluoranthene	1.549	1.477	4.6	95	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
28	Pvrene	1.364	1.316	3.5	95	-0.02
29 S	Terphenyl-d14	0.983	0.953	3.1	95	-0.01
30	Benzo(a)anthracene	1.473	1.247	15.3	83	-0.01
31	Chrysene	1.436	1.469	-2.3	99	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.367	51.6#	51	-0.02
33	Indeno(1,2,3-cd)pyrene	1.797	1.827	-1.7	98	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	93	-0.02
35	Benzo(b)fluoranthene	1.389	1.281	7.8	91	-0.02
36	Benzo(k)fluoranthene	1.373	1.470	-7.1	104	-0.02
37 C	Benzo(a)pyrene	1.305	1.157	11.3	87	-0.02
38	Dibenzo(a,h)anthracene	1.335	1.313	1.6	96	-0.03
39	Benzo(a,h,i)perylene	1.321	1.226	7.2	90	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
Data File : BN007797.D
Acq On : 21 Sep 2019 12:02
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 23 01:57:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			