

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
 Data File : BN007781.D
 Acq On : 21 Sep 2019 02:30
 Operator : HP/JU
 Sample : K4975-17
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP06-20190918

Quant Time: Sep 21 04:25:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	6412	0.40	ng	0.00
7) Naphthalene-d8	10.46	136	24278	0.40	ng	-0.01
13) Acenaphthene-d10	14.32	164	14642	0.40	ng	0.00
19) Phenanthrene-d10	17.06	188	33619	0.40	ng	-0.01
27) Chrysene-d12	21.25	240	35582	0.40	ng	-0.02
34) Perylene-d12	23.46	264	37170	0.40	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.31	112	2134	0.10	ng	0.00
5) Phenol-d6	6.87	99	1485	0.05	ng	0.00
8) Nitrobenzene-d5	8.83	82	5847	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.05	152	14044	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.80	330	1639	0.21	ng	-0.02
15) 2-Fluorobiphenyl	12.94	172	20432	0.34	ng	0.00
25) Fluoranthene-d10	19.09	212	38574	0.34	ng	-0.01
29) Terphenyl-d14	19.70	244	33918	0.39	ng	-0.01
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.19	149	2444	0.036	ng	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
Data File : BN007781.D
Acq On : 21 Sep 2019 02:30
Operator : HP/JU
Sample : K4975-17
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP06-20190918

Quant Time: Sep 21 04:25:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.69 min Scan# 604
Delta R.T. -0.01 min
Lab File: BN007781.D
Acq: 21 Sep 2019 02:30

Instrument :
BNA_N
ClientSampleId :
DUP06-20190918

Tot Ion:152 Resp: 6412
Ion Ratio Lower Upper
152 100
150 155.4 125.6 188.4
115 64.9 53.3 79.9




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#4


2-Fluorophenol
Concen: 0.097 ng
RT: 5.31 min Scan# 308
Delta R.T. -0.01 min
Lab File: BN007781.D
Acq: 21 Sep 2019 02:30

Tgt Ion:112 Resp: 2134
Ion Ratio Lower Upper
112 100
64 57.7 46.6 70.0
63 31.2 23.1 34.7



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BNC
Ion 63.00 (62.70 to 63.70): BNC





#5

Phenol-d6

Concen: 0.048 ng

RT: 6.87 min Scan# 502

Delta R.T. 0.00 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Instrument :

BNA_N

ClientSampleId :

DUP06-20190918


Tot Ion: 99 Resp: 1485

Ion Ratio Lower Upper

99 100

42 19.4 12.8 19.2#




71 42.5 27.6 41.4#



Abundance Ion 99.00 (98.70 to 99.70): BN007706.D (-493) (-)

Ion 42.00 (41.70 to 42.70): BN007706.D (-493) (-)

Ion 71.00 (70.70 to 71.70): BN007706.D (-493) (-)



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.46 min Scan# 850

Delta R.T. -0.01 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Tgt Ion: 136 Resp: 24278


Ion Ratio Lower Upper

136 100

137 11.2 9.2 13.8

54 8.3 5.8 8.8

68 8.8 6.5 9.7






Abundance Ion 136.00 (135.70 to 136.70): BN007706.D (-846) (-)

Ion 137.00 (136.70 to 137.70): BN007706.D (-846) (-)

Ion 54.00 (53.70 to 54.70): BN007706.D (-846) (-)

Ion 68.00 (67.70 to 68.70): BN007706.D (-846) (-)





#8

Nitrobenzene-d5

Concen: 0.275 ng

RT: 8.83 min Scan# 722

Delta R.T. 0.00 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Instrument :

BNA_N

ClientSampleId :

DUP06-20190918


Tot Ion: 82 Resp: 5847

Ion Ratio Lower Upper

82 100

128 36.8 24.5 36.7#




54 39.7 35.2 52.8



Abundance Ion 82.00 (81.70 to 82.70): BN007706.D (-719) (-)

Ion 128.00 (127.70 to 128.70): BN007706.D (-719) (-)

Ion 54.00 (53.70 to 54.70): BN007706.D (-719) (-)



#11

2-Methylnaphthalene-d10

Concen: 0.343 ng

RT: 12.05 min Scan# 1056

Delta R.T. -0.01 min

Lab File: BN007781.D


Acq: 21 Sep 2019 02:30

Tgt Ion: 152 Resp: 14044

Ion Ratio Lower Upper



152 100


151 19.6 15.4 23.2



Abundance Ion 152.00 (151.70 to 152.70): BN007706.D (-1046) (-)

Ion 151.00 (150.70 to 151.70): BN007706.D (-1046) (-)





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.32 min Scan# 1320

Delta R.T. -0.00 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Instrument :

BNA_N

ClientSampleId :

DUP06-20190918


Tot Ion:164 Resp: 14642

Ion Ratio Lower Upper

164 100

162 99.2 81.7 122.5




160 44.8 37.0 55.4



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.206 ng

RT: 15.80 min Scan# 1453

Delta R.T. -0.02 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30


Tgt Ion:330 Resp: 1639

Ion Ratio Lower Upper

330 100

332 0.0 0.0 0.0



141 39.0 28.7 43.1




Abundance Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.339 ng
RT: 12.94 min Scan# 1196
Delta R.T. -0.00 min
Lab File: BN007781.D
Acq: 21 Sep 2019 02:30




Instrument :
BNA_N
ClientSampleId :
DUP06-20190918


Tot Ion:172 Resp: 20432
Ion Ratio Lower Upper
172 100
171 34.8 28.8 43.2
170 22.6 19.5 29.3



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.06 min Scan# 1570
Delta R.T. -0.01 min
Lab File: BN007781.D
Acq: 21 Sep 2019 02:30

Tgt Ion:188 Resp: 33619
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 7.4 11.2#









#25
 Fluoranthene-d10
 Concen: 0.340 ng
 RT: 19.09 min Scan# 1878
 Delta R.T. -0.01 min
 Lab File: BN007781.D
 Acq: 21 Sep 2019 02:30




Instrument :
 BNA_N
 ClientSampleId :
 DUP06-20190918


Tot Ion:212 Resp: 38574
 Ion Ratio Lower Upper
 212 100
 106 11.7 9.4 14.0
 104 6.5 5.3 7.9



#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.25 min Scan# 2179
 Delta R.T. -0.02 min
 Lab File: BN007781.D
 Acq: 21 Sep 2019 02:30

Tgt Ion:240 Resp: 35582
 Ion Ratio Lower Upper
 240 100
 120 8.2 5.4 8.2#
 236 26.5 21.0 31.4





#29

Terphenyl-d14

Concen: 0.388 ng

RT: 19.70 min Scan# 2000

Delta R.T. -0.01 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Instrument :

BNA_N

ClientSampleId :

DUP06-20190918


Tot Ion:244 Resp: 33918

Ion Ratio Lower Upper




244 100

212 7.2 5.9 8.9

122 9.1 7.4 11.2



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.036 ng

RT: 21.19 min Scan# 2173

Delta R.T. -0.02 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30


Tgt Ion:149 Resp: 2444

Ion Ratio Lower Upper



149 100




167 24.0 23.7 35.5

279 4.7 4.2 6.2



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E





#34

Pervlene-d12

Concen: 0.400 ng

RT: 23.46 min Scan# 2596

Delta R.T. -0.03 min

Lab File: BN007781.D

Acq: 21 Sep 2019 02:30

Instrument :

BNA_N

ClientSampleId :

DUP06-20190918

Tot Ion:264 Resp: 37170

Ion Ratio Lower Upper

264 100

260 24.2 19.3 28.9

265 32.1 26.9 40.3

