

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
 Data File : BN016367.D
 Acq On : 21 Sep 2021 06:09
 Operator : CG/JU
 Sample : M3770-05MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-

Quant Time: Sep 21 06:38:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 20 11:09:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	26831	0.400	ng	0.00
7) Naphthalene-d8	10.457	136	111253	0.400	ng	# 0.00
13) Acenaphthene-d10	14.307	164	74580	0.400	ng	0.00
19) Phenanthrene-d10	17.054	188	141926	0.400	ng	# 0.00
29) Chrysene-d12	21.259	240	146261	0.400	ng	# 0.00
35) Perylene-d12	23.522	264	160767	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.328	112	17977	0.267	ng	0.02
5) Phenol-d6	6.868	99	32841	0.402	ng	0.00
8) Nitrobenzene-d5	8.670	82	15260	0.231	ng	-0.15
11) 2-Methylnaphthalene-d10	12.043	152	48716	0.278	ng	0.00
14) 2,4,6-Tribromophenol	15.804	330	7116	0.325	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	67991	0.233	ng	-0.01
27) Fluoranthene-d10	19.093	212	99301	0.242	ng	0.00
31) Terphenyl-d14	19.708	244	89035	0.263	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.327	88	5745	0.515	ng	# 86
3) n-Nitrosodimethylamine	3.631	42	8382	0.328	ng	# 82
6) bis(2-Chloroethyl)ether	7.126	93	29914	0.412	ng	# 19
9) Naphthalene	10.495	128	2009608	6.704	ng	98
10) Hexachlorobutadiene	10.787	225	28107	0.493	ng	# 99
12) 2-Methylnaphthalene	12.114	142	244332	1.230	ng	98
16) Acenaphthylene	14.015	152	103916	0.299	ng	# 93
17) Acenaphthene	14.370	154	80332	0.373	ng	96
18) Fluorene	15.360	166	112339	0.421	ng	# 94
20) 4,6-Dinitro-2-methylph...	15.411	198	482	0.215	ng	# 1
21) 4-Bromophenyl-phenylether	16.251	248	23207	0.279	ng	# 56
22) Hexachlorobenzene	16.360	284	24187	0.308	ng	# 86
23) Atrazine	16.543	200	29462	0.388	ng	# 86
24) Pentachlorophenol	16.713	266	18300	0.693	ng	96
25) Phenanthrene	17.091	178	776003	1.935	ng	98
26) Anthracene	17.188	178	263672	0.679	ng	96
28) Fluoranthene	19.128	202	1572977	3.405	ng	99
30) Pyrene	19.490	202	1409585	3.047	ng	# 94
32) Benzo(a)anthracene	21.238	228	1046144	2.226	ng	97
33) Chrysene	21.291	228	907071	1.995	ng	96
34) Bis(2-ethylhexyl)phtha...	21.196	149	2857174	11.274	ng	99
36) Indeno(1,2,3-cd)pyrene	25.815	276	738172	1.422	ng	97
37) Benzo(b)fluoranthene	22.841	252	1539724	3.016	ng	# 98
39) Benzo(a)pyrene	23.419	252	1081032	2.341	ng	98
40) Dibenzo(a,h)anthracene	25.836	278	280415	0.636	ng	# 92
41) Benzo(g,h,i)perylene	26.521	276	640802	1.461	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
Data File : BN016367.D
Acq On : 21 Sep 2021 06:09
Operator : CG/JU
Sample : M3770-05MSD
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-

Quant Time: Sep 21 06:38:31 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 20 11:09:23 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.689 min Scan# 500
 Delta R.T. 0.001 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09




Instrument :
 BNA_N
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion: 152 Resp: 26831
 Ion Ratio Lower Upper
 152 100
 150 147.9 125.6 188.4
 115 55.6 50.2 75.2



#2
 1,4-Dioxane
 Concen: 0.515 ng
 RT: 3.327 min Scan# 27
 Delta R.T. 0.028 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09

Tgt Ion: 88 Resp: 5745
 Ion Ratio Lower Upper
 88 100
 43 46.7 20.3 30.5#
 58 81.6 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 0.328 ng
RT: 3.631 min Scan# 60
Delta R.T. -0.009 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion: 42 Resp: 8382
Ion Ratio Lower Upper
42 100
74 119.5 113.6 170.4
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.267 ng
RT: 5.328 min Scan# 244
Delta R.T. 0.019 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion: 112 Resp: 17977
Ion Ratio Lower Upper
112 100
64 60.2 47.3 70.9
63 27.6 23.4 35.0









#5
Phenol-d6
Concen: 0.402 ng
RT: 6.868 min Scan# 411
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion: 99 Resp: 32841
Ion Ratio Lower Upper
99 100
42 193.2 14.0 21.0#
71 938.7 23.8 35.6#



#6
bis(2-Chloroethyl)ether
Concen: 0.412 ng
RT: 7.126 min Scan# 439
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion: 93 Resp: 29914
Ion Ratio Lower Upper
93 100
63 0.0 63.4 95.0#
95 0.0 26.8 40.2#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.457 min Scan# 737
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:136 Resp: 111253
Ion Ratio Lower Upper
136 100
137 15.5 8.8 13.2#
54 73.1 5.3 7.9#
68 46.6 3.9 5.9#



#8
Nitrobenzene-d5
Concen: 0.231 ng
RT: 8.670 min Scan# 596
Delta R.T. -0.152 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion: 82 Resp: 15260
Ion Ratio Lower Upper
82 100
128 41.6 31.7 47.5
54 65.6 46.4 69.6









#9
Naphthalene
Concen: 6.704 ng
RT: 10.495 min Scan# 740
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:128 Resp: 2009608
Ion Ratio Lower Upper
128 100
129 11.8 8.9 13.3
127 13.7 10.1 15.1



#10
Hexachlorobutadiene
Concen: 0.493 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:225 Resp: 28107
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.278 ng
RT: 12.043 min Scan# 940
Delta R.T. -0.004 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:152 Resp: 48716
Ion Ratio Lower Upper
152 100
151 24.3 16.1 24.1#



#12
2-Methylnaphthalene
Concen: 1.230 ng
RT: 12.114 min Scan# 956
Delta R.T. -0.004 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:142 Resp: 244332
Ion Ratio Lower Upper
142 100
141 88.6 69.8 104.8
115 33.2 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.307 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:164 Resp: 74580
Ion Ratio Lower Upper
164 100
162 98.7 79.0 118.4
160 43.7 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.325 ng
RT: 15.804 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:330 Resp: 7116
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 143.7 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.233 ng
RT: 12.924 min Scan# 1077
Delta R.T. -0.012 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:172	Resp: 67991
Ion Ratio	Lower Upper
172 100	
171 35.8	27.8 41.6
170 24.2	18.2 27.2



#16
Acenaphthylene
Concen: 0.299 ng
RT: 14.015 min Scan# 1163
Delta R.T. -0.012 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:152	Resp: 103916
Ion Ratio	Lower Upper
152 100	
151 21.5	15.2 22.8
153 16.4	10.5 15.7#









#17
Acenaphthene
Concen: 0.373 ng
RT: 14.370 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:154 Resp: 80332
Ion Ratio Lower Upper
154 100
153 114.6 89.3 133.9
152 56.4 42.2 63.4



#18
Fluorene
Concen: 0.421 ng
RT: 15.360 min Scan# 1269
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09


Tgt Ion:166 Resp: 112339
Ion Ratio Lower Upper
166 100
165 101.9 76.7 115.1
167 16.5 10.9 16.3#








#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.054 min Scan# 1406
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09


Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-





Tgt Ion:188 Resp: 141926
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.5 7.0 10.6#




#20
4,6-Dinitro-2-methylphenol
Concen: 0.215 ng
RT: 15.411 min Scan# 1273
Delta R.T. -0.025 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Tgt Ion:198 Resp: 482
Ion Ratio Lower Upper
198 100
182 1611.8 26.6 39.8#
77 0.0 0.0 0.0








#21
4-Bromophenyl-phenylether
Concen: 0.279 ng
RT: 16.251 min Scan# 1340
Delta R.T. -0.012 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-







Tgt Ion: 248 Resp: 23207
Ion Ratio Lower Upper
248 100
250 91.5 80.5 120.7
141 123.9 34.8 52.2#



#22
Hexachlorobenzene
Concen: 0.308 ng
RT: 16.360 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion: 284 Resp: 24187
Ion Ratio Lower Upper
284 100
142 42.7 22.6 34.0#
249 31.7 24.4 36.6









#23
 Atrazine
 Concen: 0.388 ng
 RT: 16.543 min Scan# 1364
 Delta R.T. 0.000 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09




Instrument :
 BNA_N
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:200 Resp: 29462
 Ion Ratio Lower Upper
 200 100
 173 38.0 18.7 28.1#
 215 57.3 41.6 62.4



#24
 Pentachlorophenol
 Concen: 0.693 ng
 RT: 16.713 min Scan# 1378
 Delta R.T. 0.000 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09

Tgt Ion:266 Resp: 18300
 Ion Ratio Lower Upper
 266 100
 264 62.2 50.6 75.8
 268 61.3 52.6 79.0









#25
Phenanthrene
Concen: 1.935 ng
RT: 17.091 min Scan# 1409
Delta R.T. -0.012 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:178 Resp: 776003
Ion Ratio Lower Upper
178 100
176 18.5 14.4 21.6
179 16.1 12.2 18.4



#26
Anthracene
Concen: 0.679 ng
RT: 17.188 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:178 Resp: 263672
Ion Ratio Lower Upper
178 100
176 18.4 13.9 20.9
179 17.6 12.2 18.4





#27

Fluoranthene-d10

Concen: 0.242 ng





RT: 19.093 min Scan# 1702

Delta R.T. -0.005 min

Lab File: BN016367.D

Acq: 21 Sep 2021 06:09

Tgt Ion:212	Resp:	99301
Ion Ratio	Lower	Upper
212	100	
106	13.5	9.8 14.6
104	8.0	5.4 8.2



#28

Fluoranthene

Concen: 3.405 ng




RT: 19.128 min Scan# 1709


Delta R.T. 0.000 min

Lab File: BN016367.D

Acq: 21 Sep 2021 06:09

Tgt Ion:202	Resp:	1572977
Ion Ratio	Lower	Upper
202	100	
101	10.6	8.5 12.7
203	17.7	13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.259 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:240 Resp: 146261
Ion Ratio Lower Upper
240 100
120 13.7 5.3 7.9#
236 27.5 19.8 29.8



#30
Pyrene
Concen: 3.047 ng
RT: 19.490 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:202 Resp: 1409585
Ion Ratio Lower Upper
202 100
200 20.6 15.8 23.6
203 21.7 13.9 20.9#









#31
 Terphenyl-d14
 Concen: 0.263 ng
 RT: 19.708 min Scan# 1826
 Delta R.T. 0.000 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09




Instrument :
 BNA_N
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:244 Resp: 89035
 Ion Ratio Lower Upper
 244 100
 212 10.1 5.5 8.3#
 122 15.9 9.0 13.4#



#32
 Benzo(a)anthracene
 Concen: 2.226 ng
 RT: 21.238 min Scan# 2002
 Delta R.T. -0.010 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09

Tgt Ion:228 Resp: 1046144
 Ion Ratio Lower Upper
 228 100
 226 28.1 20.8 31.2
 229 20.3 15.8 23.6









#33
Chrysene
Concen: 1.995 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:	228	Resp:	907071
Ion Ratio	Lower	Upper	
228	100		
226	29.8	23.0	34.4
229	22.1	15.6	23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 11.274 ng
RT: 21.196 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09


Tgt Ion:	149	Resp:	2857174
Ion Ratio	Lower	Upper	
149	100		
167	29.6	23.4	35.0
279	4.5	3.9	5.9








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.522 min Scan# 2438
 Delta R.T. -0.003 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09


Instrument :
 BNA_N
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-





Tgt Ion:264 Resp: 160767
 Ion Ratio Lower Upper
 264 100
 260 25.5 18.9 28.3
 265 30.1 23.7 35.5




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.422 ng
 RT: 25.815 min Scan# 3108
 Delta R.T. -0.003 min
 Lab File: BN016367.D
 Acq: 21 Sep 2021 06:09



Tgt Ion:276 Resp: 738172
 Ion Ratio Lower Upper
 276 100
 138 21.3 19.3 28.9
 277 24.8 20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 3.016 ng
RT: 22.841 min Scan# 2239
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09




Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-


Tgt Ion:252 Resp: 1539724
Ion Ratio Lower Upper
252 100
253 23.1 18.2 27.2
125 12.7 8.4 12.6#



#39
Benzo(a)pyrene
Concen: 2.341 ng
RT: 23.419 min Scan# 2408
Delta R.T. -0.003 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:252 Resp: 1081032
Ion Ratio Lower Upper
252 100
253 22.5 18.7 28.1
125 13.2 9.9 14.9









#40
Dibenzo(a,h)anthracene
Concen: 0.636 ng
RT: 25.836 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Instrument :
BNA_N
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-

Tgt Ion:278 Resp: 280415
Ion Ratio Lower Upper
278 100
139 24.0 14.2 21.2#
279 24.9 21.3 31.9



#41
Benzo(g,h,i)perylene
Concen: 1.461 ng
RT: 26.521 min Scan# 3314
Delta R.T. 0.000 min
Lab File: BN016367.D
Acq: 21 Sep 2021 06:09

Tgt Ion:276 Resp: 640802
Ion Ratio Lower Upper
276 100
277 24.2 20.5 30.7
138 21.0 18.2 27.4

