

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
 Data File : BN016350.D
 Acq On : 20 Sep 2021 18:46
 Operator : CG/JU
 Sample : M3817-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SW-CB-04-(1.5)

Manual Integrations
 APPROVED

mohammad
 9/21/2021 11:30:25 AM

Quant Time: Sep 21 01:34:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 20 11:09:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	25700	0.40	ng	0.00
7) Naphthalene-d8	10.457	136	114509	0.40	ng	0.00
13) Acenaphthene-d10	14.307	164	67680	0.40	ng	0.00
19) Phenanthrene-d10	17.054	188	137457	0.40	ng	# 0.00
29) Chrysene-d12	21.259	240	146006	0.40	ng	# 0.00
35) Perylene-d12	23.522	264	147961	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	4139	0.06	ng	0.00
5) Phenol-d6	6.868	99	2821	0.04	ng	0.00
8) Nitrobenzene-d5	8.822	82	11975	0.18	ng	0.00
11) 2-Methylnaphthalene-d10	12.047	152	28135	0.16	ng	0.00
14) 2,4,6-Tribromophenol	15.804	330	5546	0.28	ng	0.00
15) 2-Fluorobiphenyl	12.936	172	36618	0.14	ng	0.00
27) Fluoranthene-d10	19.097	212	84368	0.21	ng	0.00
31) Terphenyl-d14	19.708	244	66686	0.20	ng	0.00
Target Compounds						
9) Naphthalene	10.508	128	548495	1.78	ng	100
34) Bis(2-ethylhexyl)phtha...	21.195	149	27613	0.11	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
Data File : BN016350.D
Acq On : 20 Sep 2021 18:46
Operator : CG/JU
Sample : M3817-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SW-CB-04-(1.5)

Manual Integrations
APPROVED

mohammad
9/21/2021 11:30:25 AM

Quant Time: Sep 21 01:34:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 20 11:09:23 2021
Response via : Initial Calibration

