

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
 Data File : BN016358.D
 Acq On : 21 Sep 2021 00:47
 Operator : CG/JU
 Sample : M3781-15
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D1-20210916

Quant Time: Sep 21 02:30:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 20 11:09:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.679	152	24783	0.400	ng	0.00
7) Naphthalene-d8	10.445	136	110507	0.400	ng	#-0.01
13) Acenaphthene-d10	14.307	164	58495	0.400	ng	0.00
19) Phenanthrene-d10	17.054	188	119399	0.400	ng	0.00
29) Chrysene-d12	21.259	240	125332	0.400	ng	# 0.00
35) Perylene-d12	23.515	264	106844	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	8115	0.130	ng	0.00
5) Phenol-d6	6.868	99	5585	0.074	ng	0.00
8) Nitrobenzene-d5	8.822	82	23320	0.355	ng	0.00
11) 2-Methylnaphthalene-d10	12.043	152	58426	0.336	ng	0.00
14) 2,4,6-Tribromophenol	15.804	330	6230	0.363	ng	0.00
15) 2-Fluorobiphenyl	12.937	172	83495	0.365	ng	0.00
27) Fluoranthene-d10	19.093	212	148916	0.431	ng	0.00
31) Terphenyl-d14	19.703	244	135498	0.467	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.299	88	6160	0.598	ng	# 81
34) Bis(2-ethylhexyl)phtha...	21.196	149	22256	0.102	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092021\
Data File : BN016358.D
Acq On : 21 Sep 2021 00:47
Operator : CG/JU
Sample : M3781-15
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D1-20210916

Quant Time: Sep 21 02:30:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 20 11:09:23 2021
Response via : Initial Calibration

