

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
 Data File : BN016381.D
 Acq On : 21 Sep 2021 17:10
 Operator : CG/JU
 Sample : PB139259BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139259BL

Quant Time: Sep 21 17:41:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 13:21:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	5586	0.400	ng	0.02
7) Naphthalene-d8	10.445	136	24008	0.400	ng	# 0.00
13) Acenaphthene-d10	14.294	164	11067	0.400	ng	#-0.01
19) Phenanthrene-d10	17.054	188	18286	0.400	ng	# 0.00
29) Chrysene-d12	21.259	240	8718	0.400	ng	# 0.00
35) Perylene-d12	23.519	264	6028	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.347	112	4498	0.321	ng	0.07
5) Phenol-d6	6.877	99	4963	0.292	ng	0.03
8) Nitrobenzene-d5	8.823	82	6853	0.480	ng	0.01
11) 2-Methylnaphthalene-d10	12.038	152	12981	0.343	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	572	0.176	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	16483	0.381	ng	0.00
27) Fluoranthene-d10	19.093	212	13670	0.258	ng	0.00
31) Terphenyl-d14	19.703	244	9234	0.457	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
Data File : BN016381.D
Acq On : 21 Sep 2021 17:10
Operator : CG/JU
Sample : PB139259BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB139259BL

Quant Time: Sep 21 17:41:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:21:55 2021
Response via : Initial Calibration

