

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
 Data File : BN016374.D
 Acq On : 21 Sep 2021 12:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 21 13:22:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 13:21:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.395	1.3	73	0.00
3	n-Nitrosodimethylamine	0.400	0.383	4.3	77	0.00
4 S	2-Fluorophenol	0.400	0.353	11.8	72	0.00
5 S	Phenol-d6	0.400	0.348	13.0	71	0.00
6	bis(2-Chloroethyl)ether	0.400	0.411	-2.7	84	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.526	-31.5#	113	0.00
9	Naphthalene	0.400	0.386	3.5	75	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	83	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	73	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	73	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	66	0.00
15 S	2-Fluorobiphenyl	0.400	0.415	-3.7	75	0.00
16	Acenaphthylene	0.400	0.369	7.8	67	0.00
17	Acenaphthene	0.400	0.393	1.8	71	0.00
18	Fluorene	0.400	0.379	5.3	68	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.840	-110.0#	287	0.00
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	72	0.00
22	Hexachlorobenzene	0.400	0.423	-5.7	77	0.00
23	Atrazine	0.400	0.272	32.0#	53	0.00
24	Pentachlorophenol	0.400	0.329	17.8	73	0.00
25	Phenanthrene	0.400	0.403	-0.8	73	0.00
26	Anthracene	0.400	0.354	11.5	64	0.00
27 SURR	Fluoranthene-d10	0.400	0.367	8.3	67	0.00
28	Fluoranthene	0.400	0.396	1.0	72	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	74	0.00
30	Pyrene	0.400	0.402	-0.5	68	0.00
31 S	Terphenyl-d14	0.400	0.398	0.5	69	0.00
32	Benzo(a)anthracene	0.400	0.382	4.5	66	0.00
33	Chrysene	0.400	0.429	-7.2	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.226	43.5#	43	0.00
35 I	Perylene-d12	0.400	0.400	0.0	68	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.443	-10.7	72	0.00
37	Benzo(b)fluoranthene	0.400	0.434	-8.5	67	0.00
38	Benzo(k)fluoranthene	0.400	0.432	-8.0	69	0.00
39 C	Benzo(a)pyrene	0.400	0.403	-0.8	64	0.00
40	Dibenzo(a,h)anthracene	0.400	0.446	-11.5	73	0.00
41	Benzo(g,h,i)perylene	0.400	0.454	-13.5	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4