

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
 Data File : BN016390.D
 Acq On : 21 Sep 2021 23:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 22 05:02:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 13:21:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.424	-6.0	96	0.00
3	n-Nitrosodimethylamine	0.400	0.432	-8.0	106	0.03
4 S	2-Fluorophenol	0.400	0.392	2.0	97	0.00
5 S	Phenol-d6	0.400	0.382	4.5	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.411	-2.7	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.444	-11.0	117	0.00
9	Naphthalene	0.400	0.387	3.3	92	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	90	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	91	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	92	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.405	-1.3	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	95	0.00
16	Acenaphthylene	0.400	0.341	14.8	78	0.00
17	Acenaphthene	0.400	0.397	0.8	91	0.00
18	Fluorene	0.400	0.384	4.0	88	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.765	-91.3#	329	0.00
21	4-Bromophenyl-phenylether	0.400	0.382	4.5	92	0.00
22	Hexachlorobenzene	0.400	0.402	-0.5	95	0.00
23	Atrazine	0.400	0.306	23.5	77	0.00
24	Pentachlorophenol	0.400	0.421	-5.2	123	0.00
25	Phenanthrene	0.400	0.405	-1.3	97	0.00
26	Anthracene	0.400	0.348	13.0	83	0.00
27 SURR	Fluoranthene-d10	0.400	0.391	2.3	94	0.00
28	Fluoranthene	0.400	0.413	-3.2	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	93	-0.01
30	Pyrene	0.400	0.443	-10.7	93	0.00
31 S	Terphenyl-d14	0.400	0.439	-9.7	94	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	83	0.00
33	Chrysene	0.400	0.422	-5.5	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.284	29.0#	68	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.436	-9.0	91	0.00
37	Benzo(b)fluoranthene	0.400	0.403	-0.8	80	0.00
38	Benzo(k)fluoranthene	0.400	0.439	-9.7	89	0.00
39 C	Benzo(a)pyrene	0.400	0.397	0.8	81	0.00
40	Dibenzo(a,h)anthracene	0.400	0.442	-10.5	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.434	-8.5	91	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4