

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
 Data File : BN016374.D
 Acq On : 21 Sep 2021 12:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 21 13:22:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 13:21:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	19228	0.400	ng	0.00
7) Naphthalene-d8	10.444	136	82045	0.400	ng	0.00
13) Acenaphthene-d10	14.306	164	42171	0.400	ng	0.00
19) Phenanthrene-d10	17.054	188	80651	0.400	ng	# 0.00
29) Chrysene-d12	21.259	240	79432	0.400	ng	0.00
35) Perylene-d12	23.522	264	73246	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	17055	0.353	ng	0.00
5) Phenol-d6	6.849	99	20334	0.348	ng	0.00
8) Nitrobenzene-d5	8.809	82	25651	0.526	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	48370	0.374	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	4078	0.329	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	68425	0.415	ng	0.00
27) Fluoranthene-d10	19.092	212	85624	0.367	ng	0.00
31) Terphenyl-d14	19.703	244	73300	0.398	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	3158	0.395	ng	# 85
3) n-Nitrosodimethylamine	3.603	42	7012	0.383	ng	# 75
6) bis(2-Chloroethyl)ether	7.116	93	21408	0.411	ng	94
9) Naphthalene	10.495	128	85396	0.386	ng	100
10) Hexachlorobutadiene	10.787	225	17866	0.425	ng	# 99
12) 2-Methylnaphthalene	12.114	142	55576	0.379	ng	99
16) Acenaphthylene	14.015	152	72435	0.369	ng	99
17) Acenaphthene	14.357	154	47788	0.393	ng	98
18) Fluorene	15.360	166	57162	0.379	ng	99
20) 4,6-Dinitro-2-methylph...	15.436	198	3220	0.840	ng	# 65
21) 4-Bromophenyl-phenylether	16.250	248	18501	0.392	ng	# 84
22) Hexachlorobenzene	16.360	284	18865	0.423	ng	# 91
23) Atrazine	16.530	200	11738	0.272	ng	97
24) Pentachlorophenol	16.701	266	4940	0.329	ng	98
25) Phenanthrene	17.090	178	91918	0.403	ng	99
26) Anthracene	17.188	178	78172	0.354	ng	99
28) Fluoranthene	19.122	202	103983	0.396	ng	99
30) Pyrene	19.485	202	100994	0.402	ng	99
32) Benzo(a)anthracene	21.238	228	97589	0.382	ng	97
33) Chrysene	21.291	228	105882	0.429	ng	98
34) Bis(2-ethylhexyl)phtha...	21.195	149	31089	0.226	ng	97
36) Indeno(1,2,3-cd)pyrene	25.818	276	104686	0.443	ng	98
37) Benzo(b)fluoranthene	22.841	252	100851	0.434	ng	99
38) Benzo(k)fluoranthene	22.885	252	96775	0.432	ng	99
39) Benzo(a)pyrene	23.422	252	84814	0.403	ng	98
40) Dibenzo(a,h)anthracene	25.839	278	89610	0.446	ng	94
41) Benzo(g,h,i)perylene	26.520	276	90711	0.454	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
Data File : BN016374.D
Acq On : 21 Sep 2021 12:51
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Sep 21 13:22:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:21:55 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.670 min Scan# 498
Delta R.T. -0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 19228
Ion Ratio Lower Upper
152 100
150 156.1 125.6 188.4
115 59.2 50.2 75.2



#2
1,4-Dioxane
Concen: 0.395 ng
RT: 3.281 min Scan# 22
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion: 88 Resp: 3158
Ion Ratio Lower Upper
88 100
43 58.7 20.3 30.5#
58 86.7 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 0.383 ng
RT: 3.603 min Scan# 57
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 7012
Ion Ratio Lower Upper
42 100
74 110.9 113.6 170.4#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.353 ng
RT: 5.282 min Scan# 239
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion: 112 Resp: 17055
Ion Ratio Lower Upper
112 100
64 62.4 47.3 70.9
63 32.3 23.4 35.0









#5
Phenol-d6
Concen: 0.348 ng
RT: 6.849 min Scan# 409
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 20334
Ion Ratio Lower Upper
99 100
42 25.1 14.0 21.0#
71 31.8 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.411 ng
RT: 7.116 min Scan# 438
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion: 93 Resp: 21408
Ion Ratio Lower Upper
93 100
63 86.8 63.4 95.0
95 32.8 26.8 40.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.444 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	82045
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	7.6	5.3	7.9
68	4.7	3.9	5.9



#8
Nitrobenzene-d5
Concen: 0.526 ng
RT: 8.809 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:	82	Resp:	25651
Ion Ratio	Lower	Upper	
82	100		
128	36.8	31.7	47.5
54	71.0	46.4	69.6









#9
Naphthalene
Concen: 0.386 ng
RT: 10.495 min Scan# 740
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	85396
Ion Ratio	Lower	Upper	
128	100		
129	11.1	8.9	13.3
127	12.8	10.1	15.1



#10
Hexachlorobutadiene
Concen: 0.425 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:	225	Resp:	17866
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.8	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.374 ng
RT: 12.038 min Scan# 939
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 48370
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.379 ng
RT: 12.114 min Scan# 956
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:142 Resp: 55576
Ion Ratio Lower Upper
142 100
141 86.8 69.8 104.8
115 32.5 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.306 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 42171
Ion Ratio Lower Upper
164 100
162 98.0 79.0 118.4
160 44.6 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.329 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:330 Resp: 4078
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.8 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.415 ng
RT: 12.923 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	68425
Ion Ratio	Lower	Upper	
172	100		
171	34.6	27.8	41.6
170	23.0	18.2	27.2



#16
Acenaphthylene
Concen: 0.369 ng
RT: 14.015 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:	152	Resp:	72435
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.2	22.8
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.393 ng
RT: 14.357 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 47788
Ion Ratio Lower Upper
154 100
153 110.7 89.3 133.9
152 55.3 42.2 63.4



#18
Fluorene
Concen: 0.379 ng
RT: 15.360 min Scan# 1269
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:166 Resp: 57162
Ion Ratio Lower Upper
166 100
165 97.4 76.7 115.1
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.054 min Scan# 1406
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 80651
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.4 7.0 10.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.840 ng
RT: 15.436 min Scan# 1275
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:198 Resp: 3220
Ion Ratio Lower Upper
198 100
182 13.1 26.6 39.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.392 ng
RT: 16.250 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 248 Resp: 18501
Ion Ratio Lower Upper
248 100
250 91.9 80.5 120.7
141 64.4 34.8 52.2#



#22
Hexachlorobenzene
Concen: 0.423 ng
RT: 16.360 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion: 284 Resp: 18865
Ion Ratio Lower Upper
284 100
142 36.8 22.6 34.0#
249 32.2 24.4 36.6









#23
Atrazine
Concen: 0.272 ng
RT: 16.530 min Scan# 1363
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 11738
Ion Ratio Lower Upper
200 100
173 24.8 18.7 28.1
215 49.3 41.6 62.4



#24
Pentachlorophenol
Concen: 0.329 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:266 Resp: 4940
Ion Ratio Lower Upper
266 100
264 62.4 50.6 75.8
268 63.3 52.6 79.0









#25
Phenanthrene
Concen: 0.403 ng
RT: 17.090 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 91918
Ion Ratio Lower Upper
178 100
176 18.8 14.4 21.6
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.354 ng
RT: 17.188 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:178 Resp: 78172
Ion Ratio Lower Upper
178 100
176 18.0 13.9 20.9
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.367 ng
RT: 19.092 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 85624
Ion Ratio Lower Upper
212 100
106 11.1 9.8 14.6
104 6.1 5.4 8.2



#28
Fluoranthene
Concen: 0.396 ng
RT: 19.122 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:202 Resp: 103983
Ion Ratio Lower Upper
202 100
101 9.6 8.5 12.7
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.259 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 79432
Ion Ratio Lower Upper
240 100
120 7.3 5.3 7.9
236 26.2 19.8 29.8



#30
Pyrene
Concen: 0.402 ng
RT: 19.485 min Scan# 1781
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:202 Resp: 100994
Ion Ratio Lower Upper
202 100
200 20.5 15.8 23.6
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.398 ng
RT: 19.703 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 73300
Ion Ratio Lower Upper
244 100
212 7.4 5.5 8.3
122 10.6 9.0 13.4



#32
Benzo(a)anthracene
Concen: 0.382 ng
RT: 21.238 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:228 Resp: 97589
Ion Ratio Lower Upper
228 100
226 27.7 20.8 31.2
229 19.1 15.8 23.6









#33
Chrysene
Concen: 0.429 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 105882
Ion Ratio Lower Upper
228 100
226 30.2 23.0 34.4
229 19.2 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.226 ng
RT: 21.195 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51


Tgt Ion:149 Resp: 31089
Ion Ratio Lower Upper
149 100
167 27.6 23.4 35.0
279 4.4 3.9 5.9








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.522 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:264 Resp: 73246
Ion Ratio Lower Upper
264 100
260 24.7 18.9 28.3
265 29.1 23.7 35.5



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.443 ng
RT: 25.818 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:276 Resp: 104686
Ion Ratio Lower Upper
276 100
138 22.6 19.3 28.9
277 25.0 20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 0.434 ng
RT: 22.841 min Scan# 2239
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 100851
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.2
125 9.3 8.4 12.6



#38
Benzo(k)fluoranthene
Concen: 0.432 ng
RT: 22.885 min Scan# 2252
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:252 Resp: 96775
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.2
125 10.0 8.2 12.4









#39
Benzo(a)pyrene
Concen: 0.403 ng
RT: 23.422 min Scan# 2409
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 84814
Ion Ratio Lower Upper
252 100
253 22.8 18.7 28.1
125 10.5 9.9 14.9



#40
Dibenzo(a,h)anthracene
Concen: 0.446 ng
RT: 25.839 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BN016374.D
Acq: 21 Sep 2021 12:51

Tgt Ion:278 Resp: 89610
Ion Ratio Lower Upper
278 100
139 15.1 14.2 21.2
279 23.8 21.3 31.9





#41
 Benzo(g,h,i)perylene
 Concen: 0.454 ng
 RT: 26.520 min Scan# 3314
 Delta R.T. 0.000 min
 Lab File: BN016374.D
 Acq: 21 Sep 2021 12:51

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	90711
Ion	Ratio	Lower	Upper
276	100		
277	24.0	20.5	30.7
138	19.2	18.2	27.4

