



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
 Data File : BN016373.D
 Acq On : 21 Sep 2021 12:08
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 21 13:21:55 2021






AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.3	96964	PASS
68	69	0.00	2	1.7	1660	PASS
69	198	0.00	100	36.0	98923	PASS
70	69	0.00	2	0.7	701	PASS
127	198	10	80	46.7	128440	PASS
197	198	0.00	2	0.2	619	PASS
198	198	100	100	100.0	274752	PASS
199	198	5	9	6.9	19050	PASS
275	198	10	60	25.6	70323	PASS
365	198	1	100	3.0	8130	PASS
441	198	0.01	100	11.2	30843	PASS
442	442	50	100	100.0	193771	PASS
443	442	15	24	19.7	38168	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
Data File : BN016373.D
Acq On : 21 Sep 2021 12:08
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 21 13:05:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:04:53 2021
Response via : Initial Calibration



TIC: BN016373.D\data.ms

(70) Pentachlorophenol (C)

16.704min (0.000) 23347.05 ng




response 292422

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.91
264.00	61.60	60.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092121\
Data File : BN016373.D
Acq On : 21 Sep 2021 12:08
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 21 13:05:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:04:53 2021
Response via : Initial Calibration



TIC: BN016373.D\data.ms

(77) Benzidine

19.322min (0.000) 0.00 ng

response 2089593

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.46
183.00	13.20	11.54
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/21/2021	BNA_N	<u>BN016373.D</u>
Compound Name	Response	Retention Time
DDT	915307	20.563
DDD	3536	20.169
DDE	0	19.616
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3536	918843	0.38