

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092122\
 Data File : BN021854.D
 Acq On : 21 Sep 2022 23:04
 Operator : CG/JU
 Sample : N4649-15MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXZ32MSD

Quant Time: Sep 22 00:47:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	5028	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	16580	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.703	164	9149	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.455	188	16530	0.400	ng/ul	0.00
17) Chrysene-d12	21.644	240	9861	0.400	ng/ul	# 0.00
23) Perylene-d12	24.146	264	7936	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	1222	0.198	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.462	152	7715	0.294	ng/ul	-0.01
18) Fluoranthene-d10	19.483	212	14527	0.396	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.397	88	6556	1.033	ng/ul	89
5) Naphthalene	10.927	128	18436	0.394	ng/ul	98
7) 2-Methylnaphthalene	12.539	142	9104	0.300	ng/ul	98
8) 1-Methylnaphthalene	12.753	142	9603	0.308	ng/ul	99
10) Acenaphthylene	14.425	152	12336	0.321	ng/ul	99
11) Acenaphthene	14.768	153	9687	0.298	ng/ul	99
12) Fluorene	15.753	166	10951	0.300	ng/ul	99
14) Pentachlorophenol	17.105	266	2478	0.774	ng/ul	99
15) Phenanthrene	17.498	178	17729	0.340	ng/ul	99
16) Anthracene	17.591	178	14801	0.342	ng/ul	99
19) Fluoranthene	19.511	202	18097	0.393	ng/ul	96
20) Pyrene	19.873	202	17883	0.394	ng/ul#	94
21) Benzo(a)anthracene	21.627	228	13105	0.397	ng/ul	99
22) Chrysene	21.682	228	13826	0.384	ng/ul	99
24) Benzo(b)fluoranthene	23.375	252	12813	0.333	ng/ul	91
25) Benzo(k)fluoranthene	23.424	252	12306	0.330	ng/ul#	91
26) Benzo(a)pyrene	24.032	252	11558	0.396	ng/ul#	90
27) Indeno(1,2,3-cd)pyrene	26.787	276	13189	0.402	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.811	278	10807	0.388	ng/ul	96
29) Benzo(g,h,i)perylene	27.599	276	11741	0.388	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092122\
Data File : BN021854.D
Acq On : 21 Sep 2022 23:04
Operator : CG/JU
Sample : N4649-15MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXZ32MSD

Quant Time: Sep 22 00:47:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

