

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092122\
 Data File : BN021906.D
 Acq On : 23 Sep 2022 18:35
 Operator : CG/JU
 Sample : N4706-06DL 2X
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB8LODL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/26/2022
 Supervised By :mohammad ahmed 09/26/2022

Quant Time: Sep 24 01:02:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.042	152	5083	0.400	ng/ul	0.00
4) Naphthalene-d8	10.872	136	16659	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.703	164	9673	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.451	188	18834	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.647	240	11206	0.400	ng/ul	# 0.00
23) Perylene-d12	24.155	264	8760	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.359	96	6024	0.963	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.462	152	3106	0.118	ng/ul	-0.01
18) Fluoranthene-d10	19.483	212	6369	0.153	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.922	128	6405	0.136	ng/ul#	87
7) 2-Methylnaphthalene	12.533	142	3864	0.127	ng/ul	98
8) 1-Methylnaphthalene	12.753	142	4150	0.133	ng/ul	97
10) Acenaphthylene	14.430	152	1176	0.029	ng/ul#	5
11) Acenaphthene	14.763	153	35913	1.043	ng/ul	98
12) Fluorene	15.748	166	26420	0.684	ng/ul	99
15) Phenanthrene	17.493	178	37424	0.630	ng/ul	100
16) Anthracene	17.586	178	5179	0.105	ng/ul#	83
19) Fluoranthene	19.511	202	5973	0.114	ng/ul#	1
20) Pyrene	19.873	202	3786m	0.073	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092122\
Data File : BN021906.D
Acq On : 23 Sep 2022 18:35
Operator : CG/JU
Sample : N4706-06DL 2X
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB8L0DL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/26/2022
Supervised By :mohammad ahmed 09/26/2022

Quant Time: Sep 24 01:02:50 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

