

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
 Data File : BN011850.D
 Acq On : 22 Sep 2020 13:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 22 14:57:44 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 14:36:55 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	3239	0.40	ng	# 0.00
7) Naphthalene-d8	10.478	136	13595	0.40	ng	# 0.00
13) Acenaphthene-d10	14.332	164	7052	0.40	ng	0.00
19) Phenanthrene-d10	17.078	188	14642	0.40	ng	0.00
29) Chrysene-d12	21.269	240	12849	0.40	ng	# 0.00
36) Perylene-d12	23.491	264	12778	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	5076	0.53	ng	0.00
5) Phenol-d6	6.870	99	6509	0.55	ng	0.00
8) Nitrobenzene-d5	8.843	82	5930	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	12.064	152	9019	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.816	330	1353	0.38	ng	0.00
15) 2-Fluorobiphenyl	12.949	172	11615	0.38	ng	0.00
27) Fluoranthene-d10	19.107	212	17420	0.38	ng	0.00
31) Terphenyl-d14	19.713	244	12969	0.44	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.270	88	2272	0.47	ng	96
3) n-Nitrosodimethylamine	3.576	42	3318	0.52	ng	# 82
6) bis(2-Chloroethyl)ether	7.127	93	5515	0.53	ng	95
9) Naphthalene	10.516	128	16342	0.43	ng	99
10) Hexachlorobutadiene	10.820	225	2636	0.30	ng	# 98
12) 2-Methylnaphthalene	12.140	142	10290	0.39	ng	98
16) Acenaphthylene	14.040	152	13501	0.40	ng	99
17) Acenaphthene	14.395	154	9898	0.42	ng	94
18) Fluorene	15.385	166	11881	0.40	ng	100
20) 4,6-Dinitro-2-methylph...	15.461	198	886	0.35	ng	# 12
21) 4-Bromophenyl-phenylether	16.275	248	3217	0.37	ng	# 82
22) Hexachlorobenzene	16.384	284	3809	0.35	ng	# 87
23) Atrazine	16.542	200	2832	0.36	ng	# 94
24) Pentachlorophenol	16.725	266	1639	0.38	ng	98
25) Phenanthrene	17.114	178	18675	0.40	ng	99
26) Anthracene	17.212	178	15926	0.40	ng	99
28) Fluoranthene	19.137	202	20156	0.37	ng	# 95
30) Pyrene	19.504	202	20495	0.46	ng	99
32) Benzo(a)anthracene	21.248	228	17192	0.40	ng	98
33) Chrysene	21.301	228	18550	0.42	ng	98
34) Bis(2-ethylhexyl)phtha...	21.184	149	10485	0.47	ng	# 91
35) Indeno(1,2,3-cd)pyrene	25.719	276	21721	0.39	ng	# 91
37) Benzo(b)fluoranthene	22.823	252	18618	0.39	ng	# 85
38) Benzo(k)fluoranthene	22.868	252	19432	0.39	ng	# 87
39) Benzo(a)pyrene	23.391	252	16761	0.40	ng	# 75
40) Dibenzo(a,h)anthracene	25.733	278	17536	0.37	ng	# 87
41) Benzo(g,h,i)perylene	26.393	276	19065	0.38	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
Data File : BN011850.D
Acq On : 22 Sep 2020 13:41
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Sep 22 14:57:44 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 22 14:36:55 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.699 min Scan# 586
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 3239
Ion Ratio Lower Upper
152 100
150 157.5 118.3 177.5
115 76.9 51.3 76.9#



#2
1,4-Dioxane
Concen: 0.47 ng
RT: 3.270 min Scan# 36
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion: 88 Resp: 2272
Ion Ratio Lower Upper
88 100
58 83.0 63.3 94.9
43 39.6 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.52 ng
RT: 3.576 min Scan# 74
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 3318
Ion Ratio Lower Upper
42 100
74 97.7 64.4 96.6#
44 10.9 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.53 ng
RT: 5.307 min Scan# 289
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion: 112 Resp: 5076
Ion Ratio Lower Upper
112 100
64 67.8 49.0 73.6
63 39.5 31.8 47.6









#5
Phenol-d6
Concen: 0.55 ng
RT: 6.870 min Scan# 483
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 6509
Ion Ratio Lower Upper
99 100
42 26.7 23.4 35.0
71 42.8 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.53 ng
RT: 7.127 min Scan# 515
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion: 93 Resp: 5515
Ion Ratio Lower Upper
93 100
63 81.5 61.1 91.7
95 30.6 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.478 min Scan# 833
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	13595
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.5	14.3
54	6.7	5.6	8.4
68	5.0	3.2	4.8



#8
Nitrobenzene-d5
Concen: 0.43 ng
RT: 8.843 min Scan# 704
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:	82	Resp:	5930
Ion Ratio	Lower	Upper	
82	100		
128	36.7	34.2	51.2
54	54.8	43.1	64.7









#9
Naphthalene
Concen: 0.43 ng
RT: 10.516 min Scan# 836
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	16342
Ion Ratio	Lower	Upper	
128	100		
129	11.7	9.7	14.5
127	13.8	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.30 ng
RT: 10.820 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:	225	Resp:	2636
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.064 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 9019
Ion Ratio Lower Upper
152 100
151 20.2 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.140 min Scan# 1055
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:142 Resp: 10290
Ion Ratio Lower Upper
142 100
141 89.4 70.0 105.0
115 45.4 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.332 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	164	Resp:	7052
Ion	Ratio	Lower	Upper
164	100		
162	101.1	83.6	125.4
160	46.6	41.2	61.8



#14
2,4,6-Tribromophenol
Concen: 0.38 ng
RT: 15.816 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:	330	Resp:	1353
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	49.4	33.7	50.5









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.949 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	11615
Ion Ratio	Lower	Upper	
172	100		
171	34.8	27.4	41.2
170	24.5	17.8	26.6



#16
Acenaphthylene
Concen: 0.40 ng
RT: 14.040 min Scan# 1258
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:	152	Resp:	13501
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.9	23.9
153	12.7	10.5	15.7









#17
Acenaphthene
Concen: 0.42 ng
RT: 14.395 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 9898
Ion Ratio Lower Upper
154 100
153 112.3 83.4 125.2
152 56.1 42.8 64.2



#18
Fluorene
Concen: 0.40 ng
RT: 15.385 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:166 Resp: 11881
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.7 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.078 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 14642
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.6 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.35 ng
RT: 15.461 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:198 Resp: 886
Ion Ratio Lower Upper
198 100
51 144.3 45.0 67.4#
105 62.4 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.275 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 3217
Ion Ratio Lower Upper
248 100
250 99.6 82.7 124.1
141 75.6 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.35 ng
RT: 16.384 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:284 Resp: 3809
Ion Ratio Lower Upper
284 100
142 54.1 32.4 48.6#
249 31.9 26.8 40.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.542 min Scan# 1457
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 2832
Ion Ratio Lower Upper
200 100
173 28.4 18.7 28.1#
215 47.5 40.2 60.4



#24
Pentachlorophenol
Concen: 0.38 ng
RT: 16.725 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:266 Resp: 1639
Ion Ratio Lower Upper
266 100
264 65.8 51.2 76.8
268 67.1 52.4 78.6





#25

Phenanthrene

Concen: 0.40 ng

RT: 17.114 min Scan# 1504

Delta R.T. 0.000 min

Lab File: BN011850.D

Acq: 22 Sep 2020 13:41

Tgt Ion:178 Resp: 18675

Ion Ratio Lower Upper

178 100

176 19.5 15.1 22.7





179 15.6 12.3 18.5

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#26

Anthracene

Concen: 0.40 ng

RT: 17.212 min Scan# 1512

Delta R.T. 0.000 min

Lab File: BN011850.D

Acq: 22 Sep 2020 13:41




Tgt Ion:178 Resp: 15926


Ion Ratio Lower Upper

178 100

176 18.7 14.6 21.8

179 14.9 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.107 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 17420
Ion Ratio Lower Upper
212 100
106 14.1 6.4 9.6#
104 8.3 3.9 5.9#



#28
Fluoranthene
Concen: 0.37 ng
RT: 19.137 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:202 Resp: 20156
Ion Ratio Lower Upper
202 100
101 12.4 5.4 8.2#
203 17.0 13.8 20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.269 min Scan# 2098

Delta R.T. 0.000 min

Lab File: BN011850.D

Acq: 22 Sep 2020 13:41





Tgt Ion:	240	Resp:	12849
Ion Ratio	Lower	Upper	
240	100		
120	12.0	3.5	5.3#
236	26.2	20.5	30.7

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#30

Pyrene

Concen: 0.46 ng




RT: 19.504 min Scan# 1878


Delta R.T. 0.000 min

Lab File: BN011850.D

Acq: 22 Sep 2020 13:41

Tgt Ion:	202	Resp:	20495
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.5	24.7
203	17.5	13.9	20.9









#31
 Terphenyl-d14
 Concen: 0.44 ng
 RT: 19.713 min Scan# 1920
 Delta R.T. 0.000 min
 Lab File: BN011850.D
 Acq: 22 Sep 2020 13:41




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 12969
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.3 10.9
 122 14.6 7.8 11.6#



#32
 Benzo(a)anthracene
 Concen: 0.40 ng
 RT: 21.248 min Scan# 2096
 Delta R.T. 0.000 min
 Lab File: BN011850.D
 Acq: 22 Sep 2020 13:41

Tgt Ion:228 Resp: 17192
 Ion Ratio Lower Upper
 228 100
 226 27.9 21.0 31.4
 229 19.6 16.0 24.0









#33
Chrysene
Concen: 0.42 ng
RT: 21.301 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	18550
Ion Ratio	Lower	Upper	
228	100		
226	30.4	23.3	34.9
229	19.9	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.47 ng
RT: 21.184 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:	149	Resp:	10485
Ion Ratio	Lower	Upper	
149	100		
167	24.4	23.6	35.4
279	3.6	4.0	6.0#









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.719 min Scan# 3173
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 21721
Ion Ratio Lower Upper
276 100
138 24.9 13.0 19.4#
277 24.2 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.491 min Scan# 2522
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41


Tgt Ion:264 Resp: 12778
Ion Ratio Lower Upper
264 100
260 23.5 19.6 29.4
265 73.8 38.6 58.0#







#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 22.823 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 18618
Ion Ratio Lower Upper
252 100
253 27.7 19.1 28.7
125 20.6 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 22.868 min Scan# 2340
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:252 Resp: 19432
Ion Ratio Lower Upper
252 100
253 28.0 19.3 28.9
125 18.2 6.0 9.0#









#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.391 min Scan# 2493
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 16761
Ion Ratio Lower Upper
252 100
253 31.4 19.8 29.6#
125 29.0 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.733 min Scan# 3177
Delta R.T. 0.000 min
Lab File: BN011850.D
Acq: 22 Sep 2020 13:41

Tgt Ion:278 Resp: 17536
Ion Ratio Lower Upper
278 100
139 20.3 8.6 13.0#
279 28.8 19.8 29.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.38 ng
 RT: 26.393 min Scan# 3370
 Delta R.T. 0.000 min
 Lab File: BN011850.D
 Acq: 22 Sep 2020 13:41

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	19065
Ion Ratio	Lower	Upper	
276	100		
277	25.2	19.4	29.0
138	26.3	11.3	16.9#

