

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
 Data File : BN011869.D
 Acq On : 23 Sep 2020 02:29
 Operator : CG/JU
 Sample : PB131762BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB131762BL

Quant Time: Sep 23 08:27:54 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 14:36:55 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	3780	0.40	ng	0.00
7) Naphthalene-d8	10.478	136	14928	0.40	ng	0.00
13) Acenaphthene-d10	14.331	164	7171	0.40	ng	0.00
19) Phenanthrene-d10	17.078	188	14539	0.40	ng	0.00
29) Chrysene-d12	21.269	240	13663	0.40	ng	# 0.00
36) Perylene-d12	23.487	264	13877	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	4791	0.43	ng	0.00
5) Phenol-d6	6.869	99	5718	0.42	ng	0.00
8) Nitrobenzene-d5	8.843	82	4512	0.30	ng	0.00
11) 2-Methylnaphthalene-d10	12.064	152	8342	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.829	330	1242	0.34	ng	0.01
15) 2-Fluorobiphenyl	12.948	172	9714	0.31	ng	0.00
27) Fluoranthene-d10	19.107	212	12442	0.27	ng	0.00
31) Terphenyl-d14	19.712	244	12339	0.40	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.269	88	1733	0.31	ng	# 71
3) n-Nitrosodimethylamine	3.575	42	2344	0.31	ng	# 83
6) bis(2-Chloroethyl)ether	7.127	93	4258	0.35	ng	95
9) Naphthalene	10.516	128	12350	0.29	ng	99
10) Hexachlorobutadiene	10.808	225	1923	0.20	ng	# 99
12) 2-Methylnaphthalene	12.140	142	7828	0.27	ng	99
16) Acenaphthylene	14.040	152	9883	0.29	ng	100
17) Acenaphthene	14.395	154	6801	0.28	ng	94
18) Fluorene	15.384	166	8242	0.27	ng	99
20) 4,6-Dinitro-2-methylph...	15.461	198	617	0.25	ng	# 1
21) 4-Bromophenyl-phenylether	16.274	248	2062	0.24	ng	# 82
22) Hexachlorobenzene	16.384	284	2448	0.23	ng	# 88
23) Atrazine	16.542	200	1937	0.25	ng	# 91
24) Pentachlorophenol	16.725	266	2729	0.63	ng	99
25) Phenanthrene	17.114	178	12987	0.28	ng	99
26) Anthracene	17.211	178	11474	0.29	ng	99
28) Fluoranthene	19.137	202	14813	0.28	ng	# 95
30) Pyrene	19.499	202	15116	0.32	ng	99
32) Benzo(a)anthracene	21.248	228	13371	0.29	ng	99
33) Chrysene	21.301	228	13894	0.29	ng	99
34) Bis(2-ethylhexyl)phtha...	21.184	149	7569	0.32	ng	# 91
35) Indeno(1,2,3-cd)pyrene	25.715	276	16718	0.28	ng	# 89
37) Benzo(b)fluoranthene	22.820	252	13948	0.27	ng	# 81
38) Benzo(k)fluoranthene	22.864	252	15154	0.28	ng	# 81
39) Benzo(a)pyrene	23.388	252	12535	0.27	ng	# 73
40) Dibenzo(a,h)anthracene	25.732	278	13448	0.26	ng	# 82
41) Benzo(g,h,i)perylene	26.393	276	15598	0.29	ng	# 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
Data File : BN011869.D
Acq On : 23 Sep 2020 02:29
Operator : CG/JU
Sample : PB131762BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB131762BL

Quant Time: Sep 23 08:27:54 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 22 14:36:55 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.699 min Scan# 586
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:152 Resp: 3780
Ion Ratio Lower Upper
152 100
150 154.3 118.3 177.5
115 72.1 51.3 76.9



#2
1,4-Dioxane
Concen: 0.31 ng
RT: 3.269 min Scan# 36
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion: 88 Resp: 1733
Ion Ratio Lower Upper
88 100
58 85.1 63.3 94.9
43 85.3 33.0 49.4#









#3
n-Nitrosodimethylamine
Concen: 0.31 ng
RT: 3.575 min Scan# 74
Delta R.T. -0.001 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Instrument :
BNA_N
ClientSampleId :
PB131762BL

Tgt Ion: 42 Resp: 2344




Ion	Ratio	Lower	Upper
42	100		
74	96.3	64.4	96.6
44	4.4	6.7	10.1#




#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.307 min Scan# 289
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion: 112 Resp: 4791

Ion	Ratio	Lower	Upper
112	100		
64	67.6	49.0	73.6
63	39.0	31.8	47.6









#5
Phenol-d6
Concen: 0.42 ng
RT: 6.869 min Scan# 483
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion: 99 Resp: 5718
Ion Ratio Lower Upper
99 100
42 29.9 23.4 35.0
71 42.3 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.35 ng
RT: 7.127 min Scan# 515
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion: 93 Resp: 4258
Ion Ratio Lower Upper
93 100
63 82.6 61.1 91.7
95 31.5 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.478 min Scan# 833
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Instrument :
BNA_N
ClientSampleId :
PB131762BL

Tgt Ion: 136 Resp: 14928




Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.5	14.3
54	6.4	5.6	8.4
68	4.4	3.2	4.8




#8
Nitrobenzene-d5
Concen: 0.30 ng
RT: 8.843 min Scan# 704
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion: 82 Resp: 4512

Ion	Ratio	Lower	Upper
82	100		
128	39.4	34.2	51.2
54	53.4	43.1	64.7









#9
Naphthalene
Concen: 0.29 ng
RT: 10.516 min Scan# 836
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:128 Resp: 12350
Ion Ratio Lower Upper
128 100
129 11.5 9.7 14.5
127 13.8 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.20 ng
RT: 10.808 min Scan# 859
Delta R.T. -0.013 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:225 Resp: 1923
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.31 ng
RT: 12.064 min Scan# 1038
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:152 Resp: 8342
Ion Ratio Lower Upper
152 100
151 15.8 16.7 25.1#



#12
2-Methylnaphthalene
Concen: 0.27 ng
RT: 12.140 min Scan# 1055
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:142 Resp: 7828
Ion Ratio Lower Upper
142 100
141 88.4 70.0 105.0
115 45.2 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.331 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:	164	Resp:	7171
Ion	Ratio	Lower	Upper
164	100		
162	101.1	83.6	125.4
160	46.6	41.2	61.8



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.829 min Scan# 1399
Delta R.T. 0.013 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:	330	Resp:	1242
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	53.6	33.7	50.5









#15
2-Fluorobiphenyl
Concen: 0.31 ng
RT: 12.948 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:	172	Resp:	9714
Ion	Ratio	Lower	Upper
172	100		
171	35.4	27.4	41.2
170	24.3	17.8	26.6



#16
Acenaphthylene
Concen: 0.29 ng
RT: 14.040 min Scan# 1258
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:	152	Resp:	9883
Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.9	23.9
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.28 ng
RT: 14.395 min Scan# 1286
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:154 Resp: 6801
Ion Ratio Lower Upper
154 100
153 112.5 83.4 125.2
152 55.7 42.8 64.2



#18
Fluorene
Concen: 0.27 ng
RT: 15.384 min Scan# 1364
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:166 Resp: 8242
Ion Ratio Lower Upper
166 100
165 99.5 78.8 118.2
167 13.7 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.078 min Scan# 1501
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:188 Resp: 14539
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.9 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.25 ng
RT: 15.461 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:198 Resp: 617
Ion Ratio Lower Upper
198 100
51 189.2 45.0 67.4#
105 85.2 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.24 ng
RT: 16.274 min Scan# 1435
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:248 Resp: 2062
Ion Ratio Lower Upper
248 100
250 100.5 82.7 124.1
141 76.4 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.23 ng
RT: 16.384 min Scan# 1444
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:284 Resp: 2448
Ion Ratio Lower Upper
284 100
142 54.1 32.4 48.6#
249 33.0 26.8 40.2









#23
Atrazine
Concen: 0.25 ng
RT: 16.542 min Scan# 1457
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:200 Resp: 1937
Ion Ratio Lower Upper
200 100
173 31.1 18.7 28.1#
215 46.6 40.2 60.4



#24
Pentachlorophenol
Concen: 0.63 ng
RT: 16.725 min Scan# 1472
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:266 Resp: 2729
Ion Ratio Lower Upper
266 100
264 63.7 51.2 76.8
268 63.8 52.4 78.6









#25
Phenanthrene
Concen: 0.28 ng
RT: 17.114 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:178 Resp: 12987
Ion Ratio Lower Upper
178 100
176 19.7 15.1 22.7
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.29 ng
RT: 17.211 min Scan# 1512
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:178 Resp: 11474
Ion Ratio Lower Upper
178 100
176 18.7 14.6 21.8
179 15.3 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.27 ng
RT: 19.107 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:	212	Resp:	12442
Ion	Ratio	Lower	Upper
212	100		
106	14.0	6.4	9.6#
104	8.8	3.9	5.9#



#28
Fluoranthene
Concen: 0.28 ng
RT: 19.137 min Scan# 1804
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:	202	Resp:	14813
Ion	Ratio	Lower	Upper
202	100		
101	12.0	5.4	8.2#
203	16.6	13.8	20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.269 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:240 Resp: 13663
Ion Ratio Lower Upper
240 100
120 10.3 3.5 5.3#
236 26.8 20.5 30.7



#30
Pyrene
Concen: 0.32 ng
RT: 19.499 min Scan# 1877
Delta R.T. -0.005 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:202 Resp: 15116
Ion Ratio Lower Upper
202 100
200 20.9 16.5 24.7
203 17.8 13.9 20.9









#31
Terphenyl-d14
Concen: 0.40 ng
RT: 19.712 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:244 Resp: 12339
Ion Ratio Lower Upper
244 100
212 8.6 7.3 10.9
122 13.8 7.8 11.6#



#32
Benzo(a)anthracene
Concen: 0.29 ng
RT: 21.248 min Scan# 2096
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:228 Resp: 13371
Ion Ratio Lower Upper
228 100
226 26.7 21.0 31.4
229 19.7 16.0 24.0









#33
Chrysene
Concen: 0.29 ng
RT: 21.301 min Scan# 2101
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:228 Resp: 13894
Ion Ratio Lower Upper
228 100
226 29.9 23.3 34.9
229 20.1 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.32 ng
RT: 21.184 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29


Tgt Ion:149 Resp: 7569
Ion Ratio Lower Upper
149 100
167 24.3 23.6 35.4
279 3.5 4.0 6.0#








#35
Indeno(1,2,3-cd)pyrene
Concen: 0.28 ng
RT: 25.715 min Scan# 3172
Delta R.T. -0.004 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29


Instrument :
BNA_N
ClientSampleId :
PB131762BL





Tgt Ion:276 Resp: 16718
Ion Ratio Lower Upper
276 100
138 26.6 13.0 19.4#
277 24.2 20.0 30.0




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.487 min Scan# 2521
Delta R.T. -0.004 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29



Tgt Ion:264 Resp: 13877
Ion Ratio Lower Upper
264 100
260 24.7 19.6 29.4
265 72.1 38.6 58.0#









#37
Benzo(b)fluoranthene
Concen: 0.27 ng
RT: 22.820 min Scan# 2326
Delta R.T. -0.004 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:252 Resp: 13948
Ion Ratio Lower Upper
252 100
253 29.8 19.1 28.7#
125 22.2 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.28 ng
RT: 22.864 min Scan# 2339
Delta R.T. -0.004 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:252 Resp: 15154
Ion Ratio Lower Upper
252 100
253 30.5 19.3 28.9#
125 20.3 6.0 9.0#









#39
Benzo(a)pyrene
Concen: 0.27 ng
RT: 23.388 min Scan# 2492
Delta R.T. -0.004 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29




Instrument :
BNA_N
ClientSampleId :
PB131762BL


Tgt Ion:252 Resp: 12535
Ion Ratio Lower Upper
252 100
253 33.3 19.8 29.6#
125 28.2 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.26 ng
RT: 25.732 min Scan# 3177
Delta R.T. -0.000 min
Lab File: BN011869.D
Acq: 23 Sep 2020 02:29

Tgt Ion:278 Resp: 13448
Ion Ratio Lower Upper
278 100
139 21.0 8.6 13.0#
279 31.6 19.8 29.8#





#41

Benzo(g,h,i)perylene


Concen: 0.29 ng

RT: 26.393 min Scan# 3370

Delta R.T. -0.000 min

Lab File: BN011869.D

Acq: 23 Sep 2020 02:29





Tgt Ion:276 Resp: 15598

Ion Ratio Lower Upper

276 100

277 25.1 19.4 29.0

138 26.7 11.3 16.9#



Instrument :

BNA_N

ClientSampleId :

PB131762BL