

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
 Data File : BN011866.D
 Acq On : 22 Sep 2020 23:56
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 07:31:18 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 14:36:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	193	0.00
2	1,4-Dioxane	0.400	0.452	-13.0	216	0.00
3	n-Nitrosodimethylamine	0.400	0.454	-13.5	204	0.00
4 S	2-Fluorophenol	0.400	0.534	-33.5#	233	0.00
5 S	Phenol-d6	0.400	0.558	-39.5#	251	0.00
6	bis(2-Chloroethyl)ether	0.400	0.528	-32.0#	244	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	216	0.00
8 S	Nitrobenzene-d5	0.400	0.442	-10.5	228	0.00
9	Naphthalene	0.400	0.432	-8.0	225	0.00
10	Hexachlorobutadiene	0.400	0.302	24.5	156	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	198	0.00
12	2-Methylnaphthalene	0.400	0.399	0.3	209	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	195	0.00
14 S	2,4,6-Tribromophenol	0.400	0.382	4.5	181	0.00
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	175	0.00
16	Acenaphthylene	0.400	0.401	-0.3	193	0.00
17	Acenaphthene	0.400	0.416	-4.0	199	0.00
18	Fluorene	0.400	0.404	-1.0	190	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	189	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.341	14.8	165	0.00
21	4-Bromophenyl-phenylether	0.400	0.358	10.5	166	0.00
22	Hexachlorobenzene	0.400	0.350	12.5	161	0.00
23	Atrazine	0.400	0.376	6.0	179	0.00
24	Pentachlorophenol	0.400	0.396	1.0	189	0.00
25	Phenanthrene	0.400	0.404	-1.0	187	0.00
26	Anthracene	0.400	0.401	-0.3	190	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	177	0.00
28	Fluoranthene	0.400	0.375	6.3	177	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	160	0.00
30	Pyrene	0.400	0.460	-15.0	179	0.00
31 S	Terphenyl-d14	0.400	0.437	-9.2	167	0.00
32	Benzo(a)anthracene	0.400	0.396	1.0	154	0.00
33	Chrysene	0.400	0.417	-4.2	164	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.501	-25.2#	207	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	151	0.00
36 I	Perylene-d12	0.400	0.400	0.0	166	0.00
37	Benzo(b)fluoranthene	0.400	0.382	4.5	156	0.00
38	Benzo(k)fluoranthene	0.400	0.405	-1.3	159	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	166	0.00
40	Dibenzo(a,h)anthracene	0.400	0.377	5.8	150	0.00
41	Benzo(g,h,i)perylene	0.400	0.382	4.5	149	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4