

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092220\
 Data File : BN011880.D
 Acq On : 23 Sep 2020 10:20
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 11:38:04 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 11:18:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	162	0.00
2	1,4-Dioxane	0.400	0.481	-20.2	193	0.00
3	n-Nitrosodimethylamine	0.400	0.473	-18.2	178	0.00
4 S	2-Fluorophenol	0.400	0.548	-37.0#	201	0.00
5 S	Phenol-d6	0.400	0.569	-42.2#	215	0.00
6	bis(2-Chloroethyl)ether	0.400	0.542	-35.5#	210	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	180	-0.01
8 S	Nitrobenzene-d5	0.400	0.452	-13.0	194	0.00
9	Naphthalene	0.400	0.426	-6.5	184	0.00
10	Hexachlorobutadiene	0.400	0.304	24.0	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	163	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	171	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	164	0.00
14 S	2,4,6-Tribromophenol	0.400	0.396	1.0	158	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	146	0.00
16	Acenaphthylene	0.400	0.398	0.5	162	0.00
17	Acenaphthene	0.400	0.409	-2.2	165	-0.01
18	Fluorene	0.400	0.396	1.0	157	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	155	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.325	18.8	129	0.00
21	4-Bromophenyl-phenylether	0.400	0.367	8.3	140	0.00
22	Hexachlorobenzene	0.400	0.355	11.3	134	0.00
23	Atrazine	0.400	0.371	7.3	145	0.00
24	Pentachlorophenol	0.400	0.407	-1.7	159	0.00
25	Phenanthrene	0.400	0.405	-1.3	153	0.00
26	Anthracene	0.400	0.403	-0.8	156	-0.01
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	146	0.00
28	Fluoranthene	0.400	0.377	5.8	146	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	130	0.00
30	Pyrene	0.400	0.461	-15.3	146	0.00
31 S	Terphenyl-d14	0.400	0.439	-9.7	136	0.00
32	Benzo(a)anthracene	0.400	0.391	2.3	124	0.00
33	Chrysene	0.400	0.421	-5.2	135	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.502	-25.5#	169	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	120	0.00
36 I	Perylene-d12	0.400	0.400	0.0	134	0.00
37	Benzo(b)fluoranthene	0.400	0.384	4.0	126	0.00
38	Benzo(k)fluoranthene	0.400	0.395	1.3	125	0.00
39 C	Benzo(a)pyrene	0.400	0.393	1.8	129	0.00
40	Dibenzo(a,h)anthracene	0.400	0.370	7.5	119	0.00
41	Benzo(g,h,i)perylene	0.400	0.379	5.3	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4