

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092225\
 Data File : BN037850.D
 Acq On : 22 Sep 2025 10:55
 Operator : RC/JU
 Sample : PB169743BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB169743BL

Quant Time: Sep 22 11:55:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	6064	0.400	ng	-0.01
7) Naphthalene-d8	10.626	136	15212	0.400	ng	-0.02
13) Acenaphthene-d10	14.484	164	6032	0.400	ng	-0.01
19) Phenanthrene-d10	17.223	188	9991	0.400	ng	-0.01
29) Chrysene-d12	21.411	240	9227	0.400	ng	0.00
35) Perylene-d12	23.735	264	8954	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.391	112	4745	0.341	ng	0.00
5) Phenol-d6	6.980	99	5509	0.305	ng	0.00
8) Nitrobenzene-d5	8.971	82	4099	0.368	ng	-0.02
11) 2-Methylnaphthalene-d10	12.222	152	6310	0.313	ng	-0.02
14) 2,4,6-Tribromophenol	15.969	330	437	0.278	ng	-0.01
15) 2-Fluorobiphenyl	13.105	172	9105	0.361	ng	-0.01
27) Fluoranthene-d10	19.257	212	8329	0.332	ng	-0.01
31) Terphenyl-d14	19.856	244	5979	0.319	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092225\
Data File : BN037850.D
Acq On : 22 Sep 2025 10:55
Operator : RC/JU
Sample : PB169743BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB169743BL

Quant Time: Sep 22 11:55:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 14:20:01 2025
Response via : Initial Calibration

