

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
 Data File : BN011890.D
 Acq On : 23 Sep 2020 18:45
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092320

Quant Time: Sep 23 19:32:16 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.696	0.636	8.6	108	0.00
3	n-Nitrosodimethylamine	1.001	0.921	8.0	114	0.00
4 S	2-Fluorophenol	1.401	1.306	6.8	110	0.00
5 S	Phenol-d6	1.767	1.638	7.3	111	0.00
6	bis(2-Chloroethyl)ether	1.511	1.363	9.8	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.413	0.373	9.7	111	0.00
9	Naphthalene	1.146	1.039	9.3	109	0.00
10	Hexachlorobutadiene	0.187	0.172	8.0	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.575	9.2	109	0.00
12	2-Methylnaphthalene	0.729	0.660	9.5	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.194	0.168	13.4	109	0.00
15 S	2-Fluorobiphenyl	1.513	1.384	8.5	109	0.00
16	Acenaphthylene	1.928	1.689	12.4	109	0.00
17	Acenaphthene	1.320	1.216	7.9	114	0.00
18	Fluorene	1.612	1.444	10.4	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.066	13.2	112	0.00
21	4-Bromophenyl-phenylether	0.212	0.192	9.4	109	0.00
22	Hexachlorobenzene	0.247	0.232	6.1	112	0.00
23	Atrazine	0.203	0.179	11.8	111	0.00
24	Pentachlorophenol	0.121	0.102	15.7	105	0.00
25	Phenanthrene	1.227	1.096	10.7	109	0.00
26	Anthracene	1.075	0.944	12.2	109	0.00
27 SURR	Fluoranthene-d10	1.174	1.057	10.0	110	0.00
28	Fluoranthene	1.371	1.218	11.2	109	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
30	Pyrene	1.491	1.376	7.7	109	0.00
31 S	Terphenyl-d14	0.914	0.842	7.9	109	0.00
32	Benzo(a)anthracene	1.285	1.175	8.6	111	0.00
33	Chrysene	1.383	1.236	10.6	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.796	12.2	110	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.355	13.7	104	0.00
36 I	Perylene-d12	1.000	1.000	0.0	107	0.00
37	Benzo(b)fluoranthene	1.368	1.207	11.8	107	0.00
38	Benzo(k)fluoranthene	1.487	1.343	9.7	103	0.00
39 C	Benzo(a)pyrene	1.261	1.158	8.2	107	0.00
40	Dibenzo(a,h)anthracene	1.305	1.130	13.4	103	0.00
41	Benzo(g,h,i)perylene	1.410	1.248	11.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN092320