

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
 Data File : BN011890.D
 Acq On : 23 Sep 2020 18:45
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092320

Quant Time: Sep 23 19:32:16 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.366	8.5	108	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	114	0.00
4 S	2-Fluorophenol	0.400	0.373	6.8	110	0.00
5 S	Phenol-d6	0.400	0.371	7.3	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.361	9.8	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	111	0.00
9	Naphthalene	0.400	0.363	9.3	109	0.00
10	Hexachlorobutadiene	0.400	0.368	8.0	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	109	0.00
12	2-Methylnaphthalene	0.400	0.362	9.5	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.347	13.3	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	109	0.00
16	Acenaphthylene	0.400	0.350	12.5	109	0.00
17	Acenaphthene	0.400	0.368	8.0	114	0.00
18	Fluorene	0.400	0.358	10.5	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.349	12.8	112	0.00
21	4-Bromophenyl-phenylether	0.400	0.362	9.5	109	0.00
22	Hexachlorobenzene	0.400	0.375	6.3	112	0.00
23	Atrazine	0.400	0.352	12.0	111	0.00
24	Pentachlorophenol	0.400	0.337	15.8	105	0.00
25	Phenanthrene	0.400	0.357	10.8	109	0.00
26	Anthracene	0.400	0.351	12.3	109	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	110	0.00
28	Fluoranthene	0.400	0.355	11.3	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	0.00
30	Pyrene	0.400	0.369	7.8	109	0.00
31 S	Terphenyl-d14	0.400	0.369	7.8	109	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	111	0.00
33	Chrysene	0.400	0.357	10.8	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	110	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	104	0.00
36 I	Perylene-d12	0.400	0.400	0.0	107	0.00
37	Benzo(b)fluoranthene	0.400	0.353	11.8	107	0.00
38	Benzo(k)fluoranthene	0.400	0.361	9.8	103	0.00
39 C	Benzo(a)pyrene	0.400	0.367	8.3	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.354	11.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN092320