

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
 Data File : BN011893.D
 Acq On : 24 Sep 2020 00:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 01:25:17 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.353	11.8	103	0.00
3	n-Nitrosodimethylamine	0.400	0.395	1.3	121	0.00
4 S	2-Fluorophenol	0.400	0.369	7.8	108	0.00
5 S	Phenol-d6	0.400	0.368	8.0	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	-0.01
8 S	Nitrobenzene-d5	0.400	0.366	8.5	110	0.00
9	Naphthalene	0.400	0.366	8.5	108	-0.01
10	Hexachlorobutadiene	0.400	0.370	7.5	106	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.367	8.3	109	0.00
12	2-Methylnaphthalene	0.400	0.366	8.5	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.400	0.358	10.5	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	109	0.00
16	Acenaphthylene	0.400	0.352	12.0	111	-0.01
17	Acenaphthene	0.400	0.353	11.8	110	0.00
18	Fluorene	0.400	0.357	10.8	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.340	15.0	110	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	111	0.00
22	Hexachlorobenzene	0.400	0.371	7.3	112	0.00
23	Atrazine	0.400	0.356	11.0	113	0.00
24	Pentachlorophenol	0.400	0.352	12.0	110	0.00
25	Phenanthrene	0.400	0.359	10.3	111	0.00
26	Anthracene	0.400	0.353	11.8	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	111	0.00
28	Fluoranthene	0.400	0.361	9.8	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
30	Pyrene	0.400	0.367	8.3	110	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	111	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	110	-0.01
33	Chrysene	0.400	0.370	7.5	114	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.348	13.0	110	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	113	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	113	-0.01
37	Benzo(b)fluoranthene	0.400	0.362	9.5	116	-0.01
38	Benzo(k)fluoranthene	0.400	0.350	12.5	105	-0.01
39 C	Benzo(a)pyrene	0.400	0.367	8.3	113	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.357	10.8	113	-0.02
41	Benzo(g,h,i)perylene	0.400	0.358	10.5	111	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4