

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
 Data File : BN011907.D
 Acq On : 24 Sep 2020 10:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 11:13:03 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	145	0.00
2	1,4-Dioxane	0.400	0.342	14.5	137	0.00
3	n-Nitrosodimethylamine	0.400	0.343	14.2	145	0.00
4 S	2-Fluorophenol	0.400	0.381	4.8	153	0.00
5 S	Phenol-d6	0.400	0.381	4.8	155	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	152	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	147	-0.01
8 S	Nitrobenzene-d5	0.400	0.357	10.8	148	0.00
9	Naphthalene	0.400	0.361	9.8	147	-0.01
10	Hexachlorobutadiene	0.400	0.365	8.8	144	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.365	8.8	149	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	147	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	152	0.00
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	153	-0.01
15 S	2-Fluorobiphenyl	0.400	0.362	9.5	149	0.00
16	Acenaphthylene	0.400	0.350	12.5	151	-0.01
17	Acenaphthene	0.400	0.350	12.5	150	0.00
18	Fluorene	0.400	0.350	12.5	148	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	149	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.337	15.8	148	0.00
21	4-Bromophenyl-phenylether	0.400	0.353	11.8	146	0.00
22	Hexachlorobenzene	0.400	0.363	9.3	149	0.00
23	Atrazine	0.400	0.353	11.8	153	0.00
24	Pentachlorophenol	0.400	0.327	18.3	140	0.00
25	Phenanthrene	0.400	0.353	11.8	148	0.00
26	Anthracene	0.400	0.352	12.0	150	-0.01
27 SURR	Fluoranthene-d10	0.400	0.355	11.3	149	0.00
28	Fluoranthene	0.400	0.348	13.0	147	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	146	-0.01
30	Pyrene	0.400	0.370	7.5	146	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	148	0.00
32	Benzo(a)anthracene	0.400	0.353	11.8	143	-0.01
33	Chrysene	0.400	0.371	7.3	151	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.343	14.2	143	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	142	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	145	-0.01
37	Benzo(b)fluoranthene	0.400	0.359	10.3	147	-0.01
38	Benzo(k)fluoranthene	0.400	0.356	11.0	138	-0.01
39 C	Benzo(a)pyrene	0.400	0.364	9.0	144	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.355	11.3	143	-0.02
41	Benzo(g,h,i)perylene	0.400	0.351	12.3	139	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4