

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
 Data File : BN011893.D
 Acq On : 24 Sep 2020 00:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 01:25:17 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.683	152	3363	0.40	ng	0.00
7) Naphthalene-d8	10.453	136	13900	0.40	ng	#-0.01
13) Acenaphthene-d10	14.319	164	7271	0.40	ng	0.00
19) Phenanthrene-d10	17.065	188	14586	0.40	ng	0.00
29) Chrysene-d12	21.258	240	13335	0.40	ng	#-0.01
36) Perylene-d12	23.473	264	13342	0.40	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	4350	0.37	ng	0.00
5) Phenol-d6	6.853	99	5460	0.37	ng	0.00
8) Nitrobenzene-d5	8.831	82	5240	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.051	152	8082	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.816	330	1264	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.936	172	9934	0.36	ng	0.00
27) Fluoranthene-d10	19.097	212	15397	0.36	ng	0.00
31) Terphenyl-d14	19.702	244	11291	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.253	88	2067	0.35	ng	92
3) n-Nitrosodimethylamine	3.559	42	3327	0.40	ng	94
6) bis(2-Chloroethyl)ether	7.119	93	4853	0.38	ng	96
9) Naphthalene	10.503	128	14574	0.37	ng	99
10) Hexachlorobutadiene	10.795	225	2411	0.37	ng	# 98
12) 2-Methylnaphthalene	12.126	142	9270	0.37	ng	95
16) Acenaphthylene	14.027	152	12339	0.35	ng	100
17) Acenaphthene	14.382	154	8478	0.35	ng	91
18) Fluorene	15.372	166	10465	0.36	ng	99
20) 4,6-Dinitro-2-methylph...	15.448	198	945	0.34	ng	# 29
21) 4-Bromophenyl-phenylether	16.262	248	2831	0.37	ng	# 77
22) Hexachlorobenzene	16.372	284	3348	0.37	ng	# 84
23) Atrazine	16.530	200	2635	0.36	ng	# 94
24) Pentachlorophenol	16.712	266	1547	0.35	ng	97
25) Phenanthrene	17.102	178	16063	0.36	ng	99
26) Anthracene	17.199	178	13842	0.35	ng	99
28) Fluoranthene	19.127	202	18030	0.36	ng	# 95
30) Pyrene	19.494	202	18271	0.37	ng	100
32) Benzo(a)anthracene	21.237	228	15345	0.36	ng	98
33) Chrysene	21.290	228	17039	0.37	ng	97
34) Bis(2-ethylhexyl)phtha...	21.184	149	10504	0.35	ng	# 91
35) Indeno(1,2,3-cd)pyrene	25.695	276	19469	0.37	ng	# 90
37) Benzo(b)fluoranthene	22.809	252	16528	0.36	ng	# 86
38) Benzo(k)fluoranthene	22.854	252	17343	0.35	ng	# 89
39) Benzo(a)pyrene	23.378	252	15450	0.37	ng	# 78
40) Dibenzo(a,h)anthracene	25.709	278	15551	0.36	ng	# 90
41) Benzo(g,h,i)perylene	26.373	276	16841	0.36	ng	# 90

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092320\
Data File : BN011893.D
Acq On : 24 Sep 2020 00:51
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Sep 24 01:25:17 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 23 18:41:41 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.683 min Scan# 584
Delta R.T. -0.008 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 3363
Ion Ratio Lower Upper
152 100
150 157.8 118.3 177.5
115 74.2 51.3 76.9



#2
1,4-Dioxane
Concen: 0.35 ng
RT: 3.253 min Scan# 34
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion: 88 Resp: 2067
Ion Ratio Lower Upper
88 100
58 87.0 63.3 94.9
43 45.9 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.40 ng
RT: 3.559 min Scan# 72
Delta R.T. -0.008 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 3327
Ion Ratio Lower Upper
42 100
74 86.7 64.4 96.6
44 8.3 6.7 10.1



#4
2-Fluorophenol
Concen: 0.37 ng
RT: 5.291 min Scan# 287
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion: 112 Resp: 4350
Ion Ratio Lower Upper
112 100
64 70.1 49.0 73.6
63 44.0 31.8 47.6









#5
Phenol-d6
Concen: 0.37 ng
RT: 6.853 min Scan# 481
Delta R.T. -0.008 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 5460
Ion Ratio Lower Upper
99 100
42 30.8 23.4 35.0
71 45.8 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.119 min Scan# 514
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion: 93 Resp: 4853
Ion Ratio Lower Upper
93 100
63 80.9 61.1 91.7
95 30.7 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.453 min Scan# 831
Delta R.T. -0.013 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	13900
Ion Ratio	Lower	Upper	
136	100		
137	11.6	9.5	14.3
54	12.3	5.6	8.4#
68	7.3	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.831 min Scan# 703
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:	82	Resp:	5240
Ion Ratio	Lower	Upper	
82	100		
128	36.7	34.2	51.2
54	56.7	43.1	64.7









#9
Naphthalene
Concen: 0.37 ng
RT: 10.503 min Scan# 835
Delta R.T. -0.013 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	14574
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.7	14.5
127	13.7	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.795 min Scan# 858
Delta R.T. -0.013 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:	225	Resp:	2411
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.0	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.051 min Scan# 1035
Delta R.T. -0.004 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 8082
Ion Ratio Lower Upper
152 100
151 19.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.126 min Scan# 1052
Delta R.T. -0.004 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:142 Resp: 9270
Ion Ratio Lower Upper
142 100
141 91.0 70.0 105.0
115 48.7 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.319 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 7271
Ion Ratio Lower Upper
164 100
162 99.3 83.6 125.4
160 46.5 41.2 61.8



#14
2,4,6-Tribromophenol
Concen: 0.36 ng
RT: 15.816 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:330 Resp: 1264
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 54.1 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.36 ng
RT: 12.936 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	9934
Ion	Ratio	Lower	Upper
172	100		
171	35.6	27.4	41.2
170	25.1	17.8	26.6



#16
Acenaphthylene
Concen: 0.35 ng
RT: 14.027 min Scan# 1257
Delta R.T. -0.013 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:	152	Resp:	12339
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.9	23.9
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.35 ng
RT: 14.382 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 8478
Ion Ratio Lower Upper
154 100
153 115.5 83.4 125.2
152 58.0 42.8 64.2



#18
Fluorene
Concen: 0.36 ng
RT: 15.372 min Scan# 1363
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:166 Resp: 10465
Ion Ratio Lower Upper
166 100
165 99.7 78.8 118.2
167 13.5 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.065 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 14586
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.3 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.448 min Scan# 1369
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:198 Resp: 945
Ion Ratio Lower Upper
198 100
51 124.8 45.0 67.4#
105 60.5 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.262 min Scan# 1434
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 2831
Ion Ratio Lower Upper
248 100
250 100.2 82.7 124.1
141 87.1 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.37 ng
RT: 16.372 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:284 Resp: 3348
Ion Ratio Lower Upper
284 100
142 57.0 32.4 48.6#
249 31.8 26.8 40.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.530 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 2635
Ion Ratio Lower Upper
200 100
173 29.0 18.7 28.1#
215 47.5 40.2 60.4



#24
Pentachlorophenol
Concen: 0.35 ng
RT: 16.712 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:266 Resp: 1547
Ion Ratio Lower Upper
266 100
264 63.0 51.2 76.8
268 61.9 52.4 78.6









#25
Phenanthrene
Concen: 0.36 ng
RT: 17.102 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178	Resp:	16063
Ion Ratio	Lower	Upper
178	100	
176	19.5	15.1 22.7
179	15.4	12.3 18.5



#26
Anthracene
Concen: 0.35 ng
RT: 17.199 min Scan# 1511
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:178	Resp:	13842
Ion Ratio	Lower	Upper
178	100	
176	18.9	14.6 21.8
179	15.7	12.4 18.6









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.097 min Scan# 1796
Delta R.T. -0.005 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 15397
Ion Ratio Lower Upper
212 100
106 14.0 6.4 9.6#
104 8.3 3.9 5.9#



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.127 min Scan# 1802
Delta R.T. -0.005 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:202 Resp: 18030
Ion Ratio Lower Upper
202 100
101 12.2 5.4 8.2#
203 16.7 13.8 20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.258 min Scan# 2097

Delta R.T. -0.011 min

Lab File: BN011893.D

Acq: 24 Sep 2020 00:51

Instrument :





BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion:240 Resp: 13335

Ion	Ratio	Lower	Upper
240	100		
120	11.7	3.5	5.3#
236	26.6	20.5	30.7



#30

Pyrene

Concen: 0.37 ng

RT: 19.494 min Scan# 1876




Delta R.T. 0.000 min


Lab File: BN011893.D

Acq: 24 Sep 2020 00:51

Tgt Ion:202 Resp: 18271

Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.5	24.7
203	17.4	13.9	20.9









#31
Terphenyl-d14
Concen: 0.37 ng
RT: 19.702 min Scan# 1918
Delta R.T. -0.005 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 11291
Ion Ratio Lower Upper
244 100
212 9.7 7.3 10.9
122 15.1 7.8 11.6#



#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.237 min Scan# 2095
Delta R.T. -0.011 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:228 Resp: 15345
Ion Ratio Lower Upper
228 100
226 27.9 21.0 31.4
229 19.8 16.0 24.0









#33
Chrysene
Concen: 0.37 ng
RT: 21.290 min Scan# 2100
Delta R.T. -0.011 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 17039
Ion Ratio Lower Upper
228 100
226 31.2 23.3 34.9
229 19.5 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.184 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:149 Resp: 10504
Ion Ratio Lower Upper
149 100
167 24.1 23.6 35.4
279 3.8 4.0 6.0#









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 25.695 min Scan# 3166
Delta R.T. -0.010 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 19469
Ion Ratio Lower Upper
276 100
138 25.8 13.0 19.4#
277 24.3 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.473 min Scan# 2517
Delta R.T. -0.014 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Tgt Ion:264 Resp: 13342
Ion Ratio Lower Upper
264 100
260 24.5 19.6 29.4
265 60.5 38.6 58.0#









#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.809 min Scan# 2323
Delta R.T. -0.014 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 16528
Ion Ratio Lower Upper
252 100
253 27.8 19.1 28.7
125 19.8 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.854 min Scan# 2336
Delta R.T. -0.014 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51


Tgt Ion:252 Resp: 17343
Ion Ratio Lower Upper
252 100
253 27.0 19.3 28.9
125 16.8 6.0 9.0#








#39
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.378 min Scan# 2489
Delta R.T. -0.014 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 15450
Ion Ratio Lower Upper
252 100
253 30.3 19.8 29.6#
125 28.1 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.709 min Scan# 3170
Delta R.T. -0.017 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51


Tgt Ion:278 Resp: 15551
Ion Ratio Lower Upper
278 100
139 18.7 8.6 13.0#
279 27.6 19.8 29.8





#41
Benzo(g,h,i)perylene
Concen: 0.36 ng
RT: 26.373 min Scan# 3364
Delta R.T. -0.014 min
Lab File: BN011893.D
Acq: 24 Sep 2020 00:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.6	19.4	29.0
138	24.3	11.3	16.9#

