

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037867.D
 Acq On : 23 Sep 2025 15:51
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Sep 23 16:23:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	6080	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	17301	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	9570	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	17948	0.400	ng	0.00
29) Chrysene-d12	21.411	240	18531	0.400	ng	# 0.00
35) Perylene-d12	23.736	264	15649	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	75850	5.467	ng	0.00
5) Phenol-d6	6.980	99	102802	5.639	ng	0.00
8) Nitrobenzene-d5	8.971	82	73146	5.543	ng	0.00
11) 2-Methylnaphthalene-d10	12.222	152	154332	6.422	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	26327	6.816	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	210748	5.379	ng	0.00
27) Fluoranthene-d10	19.257	212	307332	6.807	ng	0.00
31) Terphenyl-d14	19.856	244	206917	5.606	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.290	88	28576	4.630	ng	98
3) n-Nitrosodimethylamine	3.593	42	41574	5.063	ng	# 92
6) bis(2-Chloroethyl)ether	7.240	93	86881	5.134	ng	99
9) Naphthalene	10.680	128	250883	5.264	ng	98
10) Hexachlorobutadiene	10.979	225	40670	5.003	ng	# 100
12) 2-Methylnaphthalene	12.293	142	173335	5.567	ng	99
16) Acenaphthylene	14.195	152	250496	5.765	ng	100
17) Acenaphthene	14.537	154	171955	5.554	ng	94
18) Fluorene	15.522	166	209070	5.583	ng	97
20) 4,6-Dinitro-2-methylph...	15.597	198	18771	4.997	ng	# 59
21) 4-Bromophenyl-phenylether	16.416	248	65230	5.577	ng	92
22) Hexachlorobenzene	16.540	284	74475	5.252	ng	99
23) Atrazine	16.689	200	55972	6.284	ng	96
24) Pentachlorophenol	16.875	266	36456	7.492	ng	96
25) Phenanthrene	17.260	178	325691	5.514	ng	99
26) Anthracene	17.359	178	305295	5.958	ng	100
28) Fluoranthene	19.285	202	384292	5.908	ng	99
30) Pyrene	19.647	202	388849	5.315	ng	100
32) Benzo(a)anthracene	21.393	228	362927	5.604	ng	98
33) Chrysene	21.447	228	363421	5.189	ng	99
34) Bis(2-ethylhexyl)phtha...	21.331	149	156492	6.109	ng	98
36) Indeno(1,2,3-cd)pyrene	26.115	276	422337	5.906	ng	99
37) Benzo(b)fluoranthene	23.034	252	387512	5.742	ng	# 93
38) Benzo(k)fluoranthene	23.081	252	376732	5.548	ng	# 92
39) Benzo(a)pyrene	23.636	252	314913	5.900	ng	# 88
40) Dibenzo(a,h)anthracene	26.133	278	333427	5.971	ng	96
41) Benzo(g,h,i)perylene	26.840	276	346868	5.913	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037867.D
Acq On : 23 Sep 2025 15:51
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :


Quant Time: Sep 23 16:23:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.825 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51


Instrument :
 BNA_N
 ClientSampleId :





Tgt Ion:152 Resp: 6080
 Ion Ratio Lower Upper
 152 100
 150 155.7 121.0 181.4
 115 58.9 47.4 71.0




#2
 1,4-Dioxane
 Concen: 4.630 ng
 RT: 3.290 min Scan# 29
 Delta R.T. -0.007 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51



Tgt Ion: 88 Resp: 28576
 Ion Ratio Lower Upper
 88 100
 43 29.8 26.6 39.8
 58 73.3 58.9 88.3





#3

n-Nitrosodimethylamine

Concen: 5.063 ng

RT: 3.593 min Scan# 71

Delta R.T. -0.007 min


Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :






Tgt Ion: 42 Resp: 41574

Ion Ratio Lower Upper

42 100

74 117.0 93.4 140.0

44 6.5 22.2 33.4#



#4

2-Fluorophenol

Concen: 5.467 ng

RT: 5.392 min Scan# 320

Delta R.T. 0.007 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion: 112 Resp: 75850

Ion Ratio Lower Upper

112 100

64 55.3 44.6 66.8

63 29.7 24.9 37.3








#5
Phenol-d6
Concen: 5.639 ng
RT: 6.980 min Scan# 54
Delta R.T. 0.007 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 99 Resp: 102802
Ion Ratio Lower Upper
99 100
42 20.6 17.2 25.8
71 33.3 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 5.134 ng
RT: 7.240 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion: 93 Resp: 86881
Ion Ratio Lower Upper
93 100
63 74.6 60.1 90.1
95 32.2 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :

Tgt Ion:	136	Resp:	17301
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	7.8	5.8	8.8
68	5.8	4.3	6.5



#8
Nitrobenzene-d5
Concen: 5.543 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:	82	Resp:	73146
Ion	Ratio	Lower	Upper
82	100		
128	41.2	34.8	52.2
54	52.4	42.2	63.2








#9
Naphthalene
Concen: 5.264 ng
RT: 10.680 min Scan# 946
Delta R.T. 0.011 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:128 Resp: 250883
Ion Ratio Lower Upper
128 100
129 10.8 9.2 13.8
127 12.8 11.0 16.6



#10
Hexachlorobutadiene
Concen: 5.003 ng
RT: 10.979 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:225 Resp: 40670
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 6.422 ng
RT: 12.222 min Scan# 1164
Delta R.T. 0.005 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :

Tgt Ion:152 Resp: 154332
Ion Ratio Lower Upper
152 100
151 21.2 17.4 26.2



#12
2-Methylnaphthalene
Concen: 5.567 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:142 Resp: 173335
Ion Ratio Lower Upper
142 100
141 88.7 71.2 106.8
115 35.2 29.7 44.5








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:164 Resp: 9570
Ion Ratio Lower Upper
164 100
162 105.9 84.8 127.2
160 55.3 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 6.816 ng
RT: 15.969 min Scan# 1590
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:330 Resp: 26327
Ion Ratio Lower Upper
330 100
332 97.4 77.2 115.8
141 41.4 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 5.379 ng
RT: 13.094 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:172 Resp: 210748
Ion Ratio Lower Upper
172 100
171 35.8 28.6 43.0
170 24.4 19.7 29.5



#16
Acenaphthylene
Concen: 5.765 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:152 Resp: 250496
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.1 10.5 15.7








#17
Acenaphthene
Concen: 5.554 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:154 Resp: 171955
Ion Ratio Lower Upper
154 100
153 108.8 90.5 135.7
152 56.5 51.8 77.8



#18
Fluorene
Concen: 5.583 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:166 Resp: 209070
Ion Ratio Lower Upper
166 100
165 95.7 79.0 118.4
167 12.9 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1

Delta R.T. -0.000 min


Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:	188	Resp:	17948
Ion Ratio	100	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.8	7.6	11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 4.997 ng




RT: 15.597 min Scan# 1560


Delta R.T. -0.000 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Tgt Ion:	198	Resp:	18771
Ion Ratio	100	Lower	Upper
198	100		
51	30.7	59.1	88.7#
105	32.7	41.3	61.9#









#21
4-Bromophenyl-phenylether
Concen: 5.577 ng
RT: 16.416 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:248 Resp: 65230
Ion Ratio Lower Upper
248 100
250 94.2 77.6 116.4
141 72.6 48.8 73.2



#22
Hexachlorobenzene
Concen: 5.252 ng
RT: 16.540 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:284 Resp: 74475
Ion Ratio Lower Upper
284 100
142 38.4 30.9 46.3
249 29.6 23.9 35.9








#23
 Atrazine
 Concen: 6.284 ng
 RT: 16.689 min Scan# 1648
 Delta R.T. -0.000 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:200 Resp: 55972
 Ion Ratio Lower Upper
 200 100
 173 21.8 18.3 27.5
 215 45.5 38.6 58.0



#24
 Pentachlorophenol
 Concen: 7.492 ng
 RT: 16.875 min Scan# 1663
 Delta R.T. 0.000 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51

Tgt Ion:266 Resp: 36456
 Ion Ratio Lower Upper
 266 100
 264 62.2 50.9 76.3
 268 63.4 54.3 81.5









#25
 Phenanthrene
 Concen: 5.514 ng
 RT: 17.260 min Scan# 1694
 Delta R.T. -0.000 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51

Instrument :
 BNA_N
 ClientSampleId :

Tgt Ion:178 Resp: 325691




Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.7	23.5
179	15.1	12.3	18.5




#26
 Anthracene
 Concen: 5.958 ng
 RT: 17.359 min Scan# 1702
 Delta R.T. -0.000 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51

Tgt Ion:178 Resp: 305295

Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.4
179	15.2	12.3	18.5









#27
Fluoranthene-d10
Concen: 6.807 ng
RT: 19.257 min Scan# 1954
Delta R.T. 0.005 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:212 Resp: 307332
Ion Ratio Lower Upper
212 100
106 15.7 12.6 19.0
104 9.0 7.4 11.0



#28
Fluoranthene
Concen: 5.908 ng
RT: 19.285 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51


Tgt Ion:202 Resp: 384292
Ion Ratio Lower Upper
202 100
101 11.7 9.3 13.9
203 17.3 13.6 20.4








#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:240 Resp: 18531
Ion Ratio Lower Upper
240 100
120 10.0 11.4 17.2#
236 27.5 23.0 34.6



#30
Pyrene
Concen: 5.315 ng
RT: 19.647 min Scan# 2039
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51


Tgt Ion:202 Resp: 388849
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.7 14.2 21.4







#31
 Terphenyl-d14
 Concen: 5.606 ng
 RT: 19.856 min Scan# 2084
 Delta R.T. 0.005 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51

Instrument :
 BNA_N
 ClientSampleId :







Tgt Ion:244 Resp: 206917
 Ion Ratio Lower Upper
 244 100
 212 8.8 7.6 11.4
 122 15.3 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 5.604 ng
 RT: 21.393 min Scan# 2271
 Delta R.T. 0.009 min
 Lab File: BN037867.D
 Acq: 23 Sep 2025 15:51

Tgt Ion:228 Resp: 362927
 Ion Ratio Lower Upper
 228 100
 226 26.9 22.6 33.8
 229 19.7 15.6 23.4





#33

Chrysene

Concen: 5.189 ng

RT: 21.447 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:228 Resp: 363421

Ion Ratio Lower Upper

228 100

226 29.6 24.6 37.0

229 19.5 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 6.109 ng

RT: 21.331 min Scan# 2264

Delta R.T. 0.009 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:149 Resp: 156492


Ion Ratio Lower Upper

149 100

167 26.8 20.4 30.6

279 3.2 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.736 min Scan# 21

Delta R.T. 0.009 min


Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:264 Resp: 15649

Ion Ratio Lower Upper

264 100

260 26.8 21.5 32.3

265 63.5 53.8 80.6



#36

Indeno(1,2,3-cd)pyrene

Concen: 5.906 ng

RT: 26.115 min Scan# 3653

Delta R.T. 0.009 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:276 Resp: 422337

Ion Ratio Lower Upper

276 100

138 26.3 21.4 32.0


277 25.1 20.3 30.5







#37
Benzo(b)fluoranthene
Concen: 5.742 ng
RT: 23.034 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:252 Resp: 387512
Ion Ratio Lower Upper
252 100
253 22.0 19.4 29.0
125 11.1 13.0 19.4#



#38
Benzo(k)fluoranthene
Concen: 5.548 ng
RT: 23.081 min Scan# 2615
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51


Tgt Ion:252 Resp: 376732
Ion Ratio Lower Upper
252 100
253 21.8 19.5 29.3
125 11.2 13.0 19.4#








#39
Benzo(a)pyrene
Concen: 5.900 ng
RT: 23.636 min Scan# 21
Delta R.T. 0.012 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:252 Resp: 314913
Ion Ratio Lower Upper
252 100
253 22.0 20.6 30.8
125 12.3 16.7 25.1#



#40
Dibenzo(a,h)anthracene
Concen: 5.971 ng
RT: 26.133 min Scan# 3659
Delta R.T. 0.015 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:278 Resp: 333427
Ion Ratio Lower Upper
278 100
139 17.5 15.4 23.2
279 23.6 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 5.913 ng

RT: 26.840 min Scan# 3901

Delta R.T. 0.012 min


Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :



Tgt Ion:276 Resp: 346868

Ion Ratio Lower Upper

276 100

277 23.8 20.2 30.4

138 22.7 19.8 29.8

