

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011939.D
 Acq On : 25 Sep 2020 07:41
 Operator : CG/JU
 Sample : L4082-21DL 2X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D2-20200916DL

Quant Time: Sep 25 08:54:24 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.683	152	2744	0.40	ng	0.00
7) Naphthalene-d8	10.453	136	11640	0.40	ng	#-0.01
13) Acenaphthene-d10	14.306	164	6400	0.40	ng	-0.01
19) Phenanthrene-d10	17.053	188	12752	0.40	ng	#-0.01
29) Chrysene-d12	21.258	240	11439	0.40	ng	#-0.01
36) Perylene-d12	23.467	264	11528	0.40	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	628	0.07	ng	0.00
5) Phenol-d6	6.861	99	374	0.03	ng	0.00
8) Nitrobenzene-d5	8.830	82	1582	0.13	ng	0.00
11) 2-Methylnaphthalene-d10	12.051	152	2365	0.13	ng	0.00
14) 2,4,6-Tribromophenol	15.803	330	572	0.18	ng	-0.01
15) 2-Fluorobiphenyl	12.936	172	3780	0.16	ng	0.00
27) Fluoranthene-d10	19.092	212	5768	0.15	ng	0.00
31) Terphenyl-d14	19.697	244	5806	0.22	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.245	88	16301	3.41	ng	95
34) Bis(2-ethylhexyl)phtha...	21.173	149	1204	0.05	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
Data File : BN011939.D
Acq On : 25 Sep 2020 07:41
Operator : CG/JU
Sample : L4082-21DL 2X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D2-20200916DL

Quant Time: Sep 25 08:54:24 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 23 18:41:41 2020
Response via : Initial Calibration

