

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011909.D
 Acq On : 24 Sep 2020 12:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 12:53:25 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.696	0.627	9.9	98	0.00
3	n-Nitrosodimethylamine	1.001	0.941	6.0	108	0.00
4 S	2-Fluorophenol	1.401	1.319	5.9	103	0.00
5 S	Phenol-d6	1.767	1.638	7.3	102	0.00
6	bis(2-Chloroethyl)ether	1.511	1.400	7.3	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.413	0.372	9.9	103	0.01
9	Naphthalene	1.146	1.044	8.9	101	0.00
10	Hexachlorobutadiene	0.187	0.175	6.4	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.571	9.8	101	0.00
12	2-Methylnaphthalene	0.729	0.653	10.4	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.01
14 S	2,4,6-Tribromophenol	0.194	0.175	9.8	106	0.00
15 S	2-Fluorobiphenyl	1.513	1.382	8.7	102	0.00
16	Acenaphthylene	1.928	1.703	11.7	104	0.00
17	Acenaphthene	1.320	1.184	10.3	104	0.01
18	Fluorene	1.612	1.465	9.1	105	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.01
20	4,6-Dinitro-2-methylphenol	0.076	0.061	19.7	99	0.00
21	4-Bromophenyl-phenylether	0.212	0.183	13.7	101	0.01
22	Hexachlorobenzene	0.247	0.225	8.9	105	0.00
23	Atrazine	0.203	0.171	15.8	103	0.00
24	Pentachlorophenol	0.121	0.104	14.0	103	0.00
25	Phenanthrene	1.227	1.083	11.7	104	0.00
26	Anthracene	1.075	0.918	14.6	102	0.00
27 SURR	Fluoranthene-d10	1.174	1.030	12.3	104	0.00
28	Fluoranthene	1.371	1.186	13.5	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
30	Pyrene	1.491	1.385	7.1	102	0.00
31 S	Terphenyl-d14	0.914	0.846	7.4	102	0.00
32	Benzo(a)anthracene	1.285	1.112	13.5	98	0.01
33	Chrysene	1.383	1.305	5.6	107	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.741	18.3	95	0.01
35	Indeno(1,2,3-cd)pyrene	1.570	1.497	4.6	107	0.01
36 I	Perylene-d12	1.000	1.000	0.0	104	0.00
37	Benzo(b)fluoranthene	1.368	1.168	14.6	101	0.00
38	Benzo(k)fluoranthene	1.487	1.347	9.4	101	0.01
39 C	Benzo(a)pyrene	1.261	1.111	11.9	101	0.00
40	Dibenzo(a,h)anthracene	1.305	1.194	8.5	107	0.01
41	Benzo(g,h,i)perylene	1.410	1.301	7.7	106	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4