

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011926.D
 Acq On : 24 Sep 2020 22:35
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 01:41:29 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.696	0.734	-5.5	119	0.00
3	n-Nitrosodimethylamine	1.001	1.049	-4.8	124	0.00
4 S	2-Fluorophenol	1.401	1.614	-15.2	130	0.00
5 S	Phenol-d6	1.767	2.062	-16.7	133	0.00
6	bis(2-Chloroethyl)ether	1.511	1.535	-1.6	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.413	0.438	-6.1	127	0.00
9	Naphthalene	1.146	1.189	-3.8	121	0.00
10	Hexachlorobutadiene	0.187	0.200	-7.0	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.658	-3.9	122	0.00
12	2-Methylnaphthalene	0.729	0.765	-4.9	123	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.01
14 S	2,4,6-Tribromophenol	0.194	0.213	-9.8	138	0.00
15 S	2-Fluorobiphenyl	1.513	1.635	-8.1	129	0.00
16	Acenaphthylene	1.928	1.958	-1.6	127	0.00
17	Acenaphthene	1.320	1.376	-4.2	129	0.00
18	Fluorene	1.612	1.625	-0.8	123	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.071	6.6	120	0.00
21	4-Bromophenyl-phenylether	0.212	0.213	-0.5	122	0.01
22	Hexachlorobenzene	0.247	0.261	-5.7	126	0.00
23	Atrazine	0.203	0.206	-1.5	128	0.00
24	Pentachlorophenol	0.121	0.114	5.8	118	0.00
25	Phenanthrene	1.227	1.219	0.7	122	0.00
26	Anthracene	1.075	1.078	-0.3	125	0.00
27 SURR	Fluoranthene-d10	1.174	1.184	-0.9	124	0.00
28	Fluoranthene	1.371	1.379	-0.6	124	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
30	Pyrene	1.491	1.614	-8.2	124	0.00
31 S	Terphenyl-d14	0.914	1.019	-11.5	128	0.00
32	Benzo(a)anthracene	1.285	1.331	-3.6	121	0.00
33	Chrysene	1.383	1.438	-4.0	122	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.904	0.3	120	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.735	-10.5	129	0.00
36 I	Perylene-d12	1.000	1.000	0.0	112	0.00
37	Benzo(b)fluoranthene	1.368	1.340	2.0	124	0.00
38	Benzo(k)fluoranthene	1.487	1.574	-5.9	126	0.00
39 C	Benzo(a)pyrene	1.261	1.323	-4.9	128	0.00
40	Dibenzo(a,h)anthracene	1.305	1.365	-4.6	131	0.00
41	Benzo(g,h,i)perylene	1.410	1.454	-3.1	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4