

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011926.D
 Acq On : 24 Sep 2020 22:35
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 01:41:29 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.422	-5.5	119	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	124	0.00
4 S	2-Fluorophenol	0.400	0.461	-15.3	130	0.00
5 S	Phenol-d6	0.400	0.467	-16.8	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	-0.01
8 S	Nitrobenzene-d5	0.400	0.424	-6.0	127	-0.01
9	Naphthalene	0.400	0.415	-3.7	121	-0.01
10	Hexachlorobutadiene	0.400	0.427	-6.7	121	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.415	-3.7	122	0.00
12	2-Methylnaphthalene	0.400	0.420	-5.0	123	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.439	-9.7	138	-0.01
15 S	2-Fluorobiphenyl	0.400	0.432	-8.0	129	0.00
16	Acenaphthylene	0.400	0.406	-1.5	127	-0.01
17	Acenaphthene	0.400	0.417	-4.2	129	-0.01
18	Fluorene	0.400	0.403	-0.8	123	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.374	6.5	120	0.00
21	4-Bromophenyl-phenylether	0.400	0.402	-0.5	122	0.00
22	Hexachlorobenzene	0.400	0.422	-5.5	126	0.00
23	Atrazine	0.400	0.405	-1.3	128	0.00
24	Pentachlorophenol	0.400	0.377	5.8	118	0.00
25	Phenanthrene	0.400	0.397	0.8	122	0.00
26	Anthracene	0.400	0.401	-0.3	125	-0.01
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	124	0.00
28	Fluoranthene	0.400	0.402	-0.5	124	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	-0.01
30	Pyrene	0.400	0.433	-8.2	124	0.00
31 S	Terphenyl-d14	0.400	0.446	-11.5	128	0.00
32	Benzo(a)anthracene	0.400	0.414	-3.5	121	-0.01
33	Chrysene	0.400	0.416	-4.0	122	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.399	0.3	120	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	129	-0.02
36 I	Perylene-d12	0.400	0.400	0.0	112	-0.02
37	Benzo(b)fluoranthene	0.400	0.392	2.0	124	-0.02
38	Benzo(k)fluoranthene	0.400	0.423	-5.7	126	-0.02
39 C	Benzo(a)pyrene	0.400	0.420	-5.0	128	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.418	-4.5	131	-0.03
41	Benzo(g,h,i)perylene	0.400	0.413	-3.2	127	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4