

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011928.D
 Acq On : 25 Sep 2020 01:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 01:41:43 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.696	0.627	9.9	108	0.00
3	n-Nitrosodimethylamine	1.001	0.941	6.0	118	0.00
4 S	2-Fluorophenol	1.401	1.343	4.1	115	0.00
5 S	Phenol-d6	1.767	1.701	3.7	117	0.00
6	bis(2-Chloroethyl)ether	1.511	1.392	7.9	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.413	0.373	9.7	114	0.00
9	Naphthalene	1.146	1.044	8.9	113	0.00
10	Hexachlorobutadiene	0.187	0.175	6.4	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.577	8.8	113	0.00
12	2-Methylnaphthalene	0.729	0.668	8.4	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.194	0.185	4.6	130	0.00
15 S	2-Fluorobiphenyl	1.513	1.354	10.5	116	0.00
16	Acenaphthylene	1.928	1.707	11.5	120	0.00
17	Acenaphthene	1.320	1.176	10.9	120	0.00
18	Fluorene	1.612	1.437	10.9	119	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.062	18.4	113	0.00
21	4-Bromophenyl-phenylether	0.212	0.193	9.0	120	0.01
22	Hexachlorobenzene	0.247	0.233	5.7	122	0.00
23	Atrazine	0.203	0.180	11.3	121	0.00
24	Pentachlorophenol	0.121	0.106	12.4	119	0.00
25	Phenanthrene	1.227	1.092	11.0	118	0.00
26	Anthracene	1.075	0.949	11.7	119	0.00
27 SURR	Fluoranthene-d10	1.174	1.028	12.4	116	0.00
28	Fluoranthene	1.371	1.233	10.1	120	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	117	-0.01
30	Pyrene	1.491	1.404	5.8	119	0.00
31 S	Terphenyl-d14	0.914	0.882	3.5	122	0.00
32	Benzo(a)anthracene	1.285	1.109	13.7	112	0.00
33	Chrysene	1.383	1.345	2.7	127	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.761	16.1	112	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.522	3.1	125	0.00
36 I	Perylene-d12	1.000	1.000	0.0	122	0.00
37	Benzo(b)fluoranthene	1.368	1.191	12.9	120	0.00
38	Benzo(k)fluoranthene	1.487	1.385	6.9	121	0.00
39 C	Benzo(a)pyrene	1.261	1.173	7.0	124	0.00
40	Dibenzo(a,h)anthracene	1.305	1.206	7.6	126	0.00
41	Benzo(g,h,i)perylene	1.410	1.328	5.8	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4