

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092420\
 Data File : BN011928.D
 Acq On : 25 Sep 2020 01:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 01:41:43 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 18:41:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.360	10.0	108	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	118	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	115	0.00
5 S	Phenol-d6	0.400	0.385	3.8	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.369	7.8	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.361	9.8	114	0.00
9	Naphthalene	0.400	0.364	9.0	113	0.00
10	Hexachlorobutadiene	0.400	0.373	6.8	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	113	0.00
12	2-Methylnaphthalene	0.400	0.366	8.5	114	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.400	0.380	5.0	130	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	116	0.00
16	Acenaphthylene	0.400	0.354	11.5	120	0.00
17	Acenaphthene	0.400	0.356	11.0	120	0.00
18	Fluorene	0.400	0.357	10.8	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	118	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.325	18.8	113	0.00
21	4-Bromophenyl-phenylether	0.400	0.365	8.8	120	0.01
22	Hexachlorobenzene	0.400	0.376	6.0	122	0.00
23	Atrazine	0.400	0.354	11.5	121	0.00
24	Pentachlorophenol	0.400	0.352	12.0	119	0.00
25	Phenanthrene	0.400	0.356	11.0	118	0.00
26	Anthracene	0.400	0.353	11.8	119	0.00
27 SURR	Fluoranthene-d10	0.400	0.350	12.5	116	0.00
28	Fluoranthene	0.400	0.360	10.0	120	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	117	-0.01
30	Pyrene	0.400	0.377	5.8	119	0.00
31 S	Terphenyl-d14	0.400	0.386	3.5	122	0.00
32	Benzo(a)anthracene	0.400	0.345	13.8	112	0.00
33	Chrysene	0.400	0.389	2.8	127	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.336	16.0	112	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.388	3.0	125	0.00
36 I	Perylene-d12	0.400	0.400	0.0	122	0.00
37	Benzo(b)fluoranthene	0.400	0.348	13.0	120	0.00
38	Benzo(k)fluoranthene	0.400	0.373	6.8	121	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	124	0.00
40	Dibenzo(a,h)anthracene	0.400	0.369	7.8	126	0.00
41	Benzo(g,h,i)perylene	0.400	0.377	5.8	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4