

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011953.D
 Acq On : 25 Sep 2020 17:01
 Operator : CG/JU
 Sample : L4119-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE103D2-20200921

Quant Time: Sep 25 17:33:12 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.683	152	2471	0.40	ng	0.00
7) Naphthalene-d8	10.453	136	10119	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	5578	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	10927	0.40	ng	# 0.00
29) Chrysene-d12	21.258	240	9479	0.40	ng	# 0.00
36) Perylene-d12	23.473	264	9704	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	925	0.11	ng	0.00
5) Phenol-d6	6.853	99	577	0.05	ng	0.00
8) Nitrobenzene-d5	8.818	82	3273	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.046	152	4452	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.803	330	831	0.31	ng	0.00
15) 2-Fluorobiphenyl	12.936	172	7889	0.37	ng	0.00
27) Fluoranthene-d10	19.092	212	10768	0.34	ng	0.00
31) Terphenyl-d14	19.697	244	9974	0.46	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	2549	0.59	ng	Qvalue 92
34) Bis(2-ethylhexyl)phtha...	21.173	149	872	0.04	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
Data File : BN011953.D
Acq On : 25 Sep 2020 17:01
Operator : CG/JU
Sample : L4119-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE103D2-20200921

Quant Time: Sep 25 17:33:12 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 11:18:41 2020
Response via : Initial Calibration

