

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011957.D
 Acq On : 25 Sep 2020 19:25
 Operator : CG/JU
 Sample : L4083-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE104D3-20200917

Quant Time: Sep 26 04:20:20 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.683	152	2536	0.40	ng	0.00
7) Naphthalene-d8	10.453	136	10468	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	5784	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	11447	0.40	ng	# 0.00
29) Chrysene-d12	21.259	240	10059	0.40	ng	# 0.00
36) Perylene-d12	23.467	264	10436	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	857	0.10	ng	0.00
5) Phenol-d6	6.853	99	575	0.05	ng	0.00
8) Nitrobenzene-d5	8.818	82	2796	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.046	152	4449	0.27	ng	0.00
14) 2,4,6-Tribromophenol	15.803	330	730	0.26	ng	0.00
15) 2-Fluorobiphenyl	12.936	172	6087	0.28	ng	0.00
27) Fluoranthene-d10	19.092	212	10809	0.32	ng	0.00
31) Terphenyl-d14	19.698	244	9265	0.40	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.174	149	1185	0.05	ng	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
Data File : BN011957.D
Acq On : 25 Sep 2020 19:25
Operator : CG/JU
Sample : L4083-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE104D3-20200917

Quant Time: Sep 26 04:20:20 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 11:18:41 2020
Response via : Initial Calibration

