

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011967.D
 Acq On : 26 Sep 2020 03:14
 Operator : CG/JU
 Sample : L4119-19
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP06-20200922

Quant Time: Sep 26 05:09:53 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.683	152	2542	0.40	ng	0.00
7) Naphthalene-d8	10.453	136	10317	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	5501	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	10560	0.40	ng	0.00
29) Chrysene-d12	21.248	240	9805	0.40	ng	#-0.01
36) Perylene-d12	23.460	264	10287	0.40	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	901	0.10	ng	0.00
5) Phenol-d6	6.845	99	634	0.06	ng	0.00
8) Nitrobenzene-d5	8.818	82	3510	0.33	ng	0.00
11) 2-Methylnaphthalene-d10	12.042	152	4825	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.803	330	757	0.28	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	7045	0.34	ng	-0.01
27) Fluoranthene-d10	19.087	212	10488	0.34	ng	0.00
31) Terphenyl-d14	19.692	244	9128	0.41	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	1643	0.37	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.173	149	1867	0.08	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
Data File : BN011967.D
Acq On : 26 Sep 2020 03:14
Operator : CG/JU
Sample : L4119-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP06-20200922

Quant Time: Sep 26 05:09:53 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 11:18:41 2020
Response via : Initial Calibration

