

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011943.D
 Acq On : 25 Sep 2020 10:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 11:18:59 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.696	0.706	-1.4	86	0.00
3	n-Nitrosodimethylamine	1.001	1.208	-20.7	108	0.00
4 S	2-Fluorophenol	1.401	1.590	-13.5	97	0.00
5 S	Phenol-d6	1.767	2.013	-13.9	98	0.00
6	bis(2-Chloroethyl)ether	1.511	1.459	3.4	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
8 S	Nitrobenzene-d5	0.413	0.474	-14.8	100	0.00
9	Naphthalene	1.146	1.195	-4.3	89	0.00
10	Hexachlorobutadiene	0.187	0.208	-11.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.655	-3.5	88	0.00
12	2-Methylnaphthalene	0.729	0.761	-4.4	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.194	0.213	-9.8	98	0.00
15 S	2-Fluorobiphenyl	1.513	1.660	-9.7	92	0.00
16	Acenaphthylene	1.928	2.007	-4.1	92	0.00
17	Acenaphthene	1.320	1.353	-2.5	90	0.00
18	Fluorene	1.612	1.656	-2.7	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.078	-2.6	89	0.00
21	4-Bromophenyl-phenylether	0.212	0.222	-4.7	86	0.00
22	Hexachlorobenzene	0.247	0.273	-10.5	90	0.00
23	Atrazine	0.203	0.213	-4.9	90	0.00
24	Pentachlorophenol	0.121	0.118	2.5	83	0.00
25	Phenanthrene	1.227	1.253	-2.1	85	0.00
26	Anthracene	1.075	1.099	-2.2	86	0.00
27 SURR	Fluoranthene-d10	1.174	1.203	-2.5	85	0.00
28	Fluoranthene	1.371	1.348	1.7	82	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
30	Pyrene	1.491	1.575	-5.6	85	0.00
31 S	Terphenyl-d14	0.914	0.990	-8.3	87	0.00
32	Benzo(a)anthracene	1.285	1.308	-1.8	84	0.00
33	Chrysene	1.383	1.423	-2.9	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.919	-1.3	86	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.776	-13.1	92	0.00
36 I	Perylene-d12	1.000	1.000	0.0	78	0.00
37	Benzo(b)fluoranthene	1.368	1.395	-2.0	90	0.00
38	Benzo(k)fluoranthene	1.487	1.547	-4.0	87	0.00
39 C	Benzo(a)pyrene	1.261	1.380	-9.4	94	0.00
40	Dibenzo(a,h)anthracene	1.305	1.367	-4.8	92	0.00
41	Benzo(g,h,i)perylene	1.410	1.474	-4.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4