

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011960.D
 Acq On : 25 Sep 2020 22:26
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 04:20:57 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.696	0.632	9.2	93	0.00
3	n-Nitrosodimethylamine	1.001	1.003	-0.2	108	0.00
4 S	2-Fluorophenol	1.401	1.367	2.4	100	0.00
5 S	Phenol-d6	1.767	1.700	3.8	100	0.00
6	bis(2-Chloroethyl)ether	1.511	1.402	7.2	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.413	0.381	7.7	100	0.00
9	Naphthalene	1.146	1.051	8.3	97	0.00
10	Hexachlorobutadiene	0.187	0.179	4.3	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.579	8.5	97	0.00
12	2-Methylnaphthalene	0.729	0.662	9.2	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.194	0.186	4.1	111	0.00
15 S	2-Fluorobiphenyl	1.513	1.387	8.3	100	-0.01
16	Acenaphthylene	1.928	1.741	9.7	104	0.00
17	Acenaphthene	1.320	1.196	9.4	103	0.00
18	Fluorene	1.612	1.454	9.8	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.064	15.8	100	-0.01
21	4-Bromophenyl-phenylether	0.212	0.188	11.3	99	0.00
22	Hexachlorobenzene	0.247	0.234	5.3	104	-0.01
23	Atrazine	0.203	0.181	10.8	103	-0.01
24	Pentachlorophenol	0.121	0.100	17.4	96	0.00
25	Phenanthrene	1.227	1.121	8.6	103	-0.01
26	Anthracene	1.075	0.938	12.7	100	0.00
27 SURR	Fluoranthene-d10	1.174	1.027	12.5	99	0.00
28	Fluoranthene	1.371	1.239	9.6	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	97	-0.01
30	Pyrene	1.491	1.424	4.5	101	0.00
31 S	Terphenyl-d14	0.914	0.868	5.0	100	0.00
32	Benzo(a)anthracene	1.285	1.105	14.0	93	0.00
33	Chrysene	1.383	1.336	3.4	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.775	14.6	95	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.611	-2.6	110	0.00
36 I	Perylene-d12	1.000	1.000	0.0	104	0.00
37	Benzo(b)fluoranthene	1.368	1.195	12.6	103	0.00
38	Benzo(k)fluoranthene	1.487	1.375	7.5	103	0.00
39 C	Benzo(a)pyrene	1.261	1.180	6.4	106	0.00
40	Dibenzo(a,h)anthracene	1.305	1.219	6.6	108	0.00
41	Benzo(g,h,i)perylene	1.410	1.346	4.5	109	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4