

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092520\
 Data File : BN011977.D
 Acq On : 26 Sep 2020 09:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 28 00:47:42 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	62	0.00
2	1,4-Dioxane	0.696	0.583	16.2	58	0.00
3	n-Nitrosodimethylamine	1.001	1.179	-17.8	86	0.00
4 S	2-Fluorophenol	1.401	1.365	2.6	68	0.00
5 S	Phenol-d6	1.767	1.728	2.2	69	0.00
6	bis(2-Chloroethyl)ether	1.511	1.391	7.9	65	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
8 S	Nitrobenzene-d5	0.413	0.401	2.9	72	0.00
9	Naphthalene	1.146	1.057	7.8	67	0.00
10	Hexachlorobutadiene	0.187	0.178	4.8	67	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.567	10.4	65	0.00
12	2-Methylnaphthalene	0.729	0.670	8.1	67	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.194	0.199	-2.6	80	0.00
15 S	2-Fluorobiphenyl	1.513	1.402	7.3	69	-0.01
16	Acenaphthylene	1.928	1.774	8.0	71	0.00
17	Acenaphthene	1.320	1.201	9.0	70	0.00
18	Fluorene	1.612	1.475	8.5	70	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.065	14.5	67	0.00
21	4-Bromophenyl-phenylether	0.212	0.194	8.5	68	0.00
22	Hexachlorobenzene	0.247	0.234	5.3	69	-0.01
23	Atrazine	0.203	0.192	5.4	73	-0.01
24	Pentachlorophenol	0.121	0.105	13.2	66	-0.01
25	Phenanthrene	1.227	1.123	8.5	69	-0.01
26	Anthracene	1.075	0.961	10.6	68	0.00
27 SURR	Fluoranthene-d10	1.174	1.040	11.4	66	0.00
28	Fluoranthene	1.371	1.261	8.0	69	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	66	-0.01
30	Pyrene	1.491	1.445	3.1	69	0.00
31 S	Terphenyl-d14	0.914	0.877	4.0	68	0.00
32	Benzo(a)anthracene	1.285	1.130	12.1	64	0.00
33	Chrysene	1.383	1.341	3.0	71	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.843	7.1	70	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.512	3.7	70	0.00
36 I	Perylene-d12	1.000	1.000	0.0	69	-0.01
37	Benzo(b)fluoranthene	1.368	1.257	8.1	72	0.00
38	Benzo(k)fluoranthene	1.487	1.331	10.5	66	0.00
39 C	Benzo(a)pyrene	1.261	1.158	8.2	69	0.00
40	Dibenzo(a,h)anthracene	1.305	1.198	8.2	71	-0.01
41	Benzo(g,h,i)perylene	1.410	1.279	9.3	69	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4