

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
 Data File : BN016486.D
 Acq On : 25 Sep 2021 22:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 04:16:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	21819	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	87806	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	42159	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	76951	0.400	ng	0.00
29) Chrysene-d12	21.249	240	74013	0.400	ng	# 0.00
35) Perylene-d12	23.515	264	64288	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	19697	0.402	ng	0.00
5) Phenol-d6	6.849	99	22614	0.391	ng	0.00
8) Nitrobenzene-d5	8.810	82	26531	0.427	ng	0.00
11) 2-Methylnaphthalene-d10	12.030	152	51050	0.392	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	3794	0.366	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	71974	0.407	ng	-0.01
27) Fluoranthene-d10	19.088	212	77328	0.366	ng	0.00
31) Terphenyl-d14	19.698	244	68036	0.395	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	2885	0.309	ng	# 89
3) n-Nitrosodimethylamine	3.604	42	7001	0.312	ng	# 13
6) bis(2-Chloroethyl)ether	7.108	93	24611	0.421	ng	90
9) Naphthalene	10.483	128	92888	0.397	ng	99
10) Hexachlorobutadiene	10.774	225	20041	0.430	ng	# 99
12) 2-Methylnaphthalene	12.101	142	58819	0.396	ng	97
16) Acenaphthylene	14.002	152	73996	0.432	ng	100
17) Acenaphthene	14.358	154	47965	0.394	ng	94
18) Fluorene	15.347	166	57860	0.396	ng	98
20) 4,6-Dinitro-2-methylph...	15.436	198	875	0.245	ng	# 87
21) 4-Bromophenyl-phenylether	16.251	248	18553	0.413	ng	# 90
22) Hexachlorobenzene	16.348	284	20574	0.462	ng	# 89
23) Atrazine	16.519	200	7936	0.321	ng	93
24) Pentachlorophenol	16.701	266	4463	0.439	ng	98
25) Phenanthrene	17.091	178	93454	0.408	ng	99
26) Anthracene	17.176	178	73757	0.387	ng	99
28) Fluoranthene	19.118	202	102387	0.408	ng	98
30) Pyrene	19.480	202	98978	0.427	ng	99
32) Benzo(a)anthracene	21.238	228	80688	0.376	ng	100
33) Chrysene	21.291	228	100041	0.434	ng	99
34) Bis(2-ethylhexyl)phtha...	21.185	149	18334	0.313	ng	95
36) Indeno(1,2,3-cd)pyrene	25.816	276	72816	0.350	ng	98
37) Benzo(b)fluoranthene	22.834	252	75598	0.352	ng	97
38) Benzo(k)fluoranthene	22.879	252	91711	0.442	ng	99
39) Benzo(a)pyrene	23.416	252	63504	0.363	ng	# 96
40) Dibenzo(a,h)anthracene	25.833	278	59190	0.334	ng	96
41) Benzo(g,h,i)perylene	26.510	276	57868	0.318	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
 Data File : BN016486.D
 Acq On : 25 Sep 2021 22:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 04:16:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.670 min Scan# 498
 Delta R.T. 0.000 min
 Lab File: BN016486.D
 Acq: 25 Sep 2021 22:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 21819
 Ion Ratio Lower Upper
 152 100
 150 153.5 125.6 188.4
 115 57.4 50.2 75.2



#2
 1,4-Dioxane
 Concen: 0.309 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN016486.D
 Acq: 25 Sep 2021 22:59

Tgt Ion: 88 Resp: 2885
 Ion Ratio Lower Upper
 88 100
 43 39.7 20.3 30.5#
 58 81.6 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 0.312 ng
RT: 3.604 min Scan# 57
Delta R.T. -0.027 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 7001
Ion Ratio Lower Upper
42 100
74 35.5 113.6 170.4#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.402 ng
RT: 5.273 min Scan# 238
Delta R.T. -0.009 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion: 112 Resp: 19697
Ion Ratio Lower Upper
112 100
64 64.4 47.3 70.9
63 34.2 23.4 35.0









#5
Phenol-d6
Concen: 0.391 ng
RT: 6.849 min Scan# 409
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 22614
Ion Ratio Lower Upper
99 100
42 28.3 14.0 21.0#
71 32.5 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.421 ng
RT: 7.108 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion: 93 Resp: 24611
Ion Ratio Lower Upper
93 100
63 91.4 63.4 95.0
95 32.8 26.8 40.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.012 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	87806
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	10.0	5.3	7.9#
68	5.6	3.9	5.9



#8
Nitrobenzene-d5
Concen: 0.427 ng
RT: 8.810 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:	82	Resp:	26531
Ion Ratio	Lower	Upper	
82	100		
128	40.6	31.7	47.5
54	66.4	46.4	69.6









#9
Naphthalene
Concen: 0.397 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 92888
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.9 10.1 15.1



#10
Hexachlorobutadiene
Concen: 0.430 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:225 Resp: 20041
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.392 ng
RT: 12.030 min Scan# 937
Delta R.T. -0.004 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 51050
Ion Ratio Lower Upper
152 100
151 20.3 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.396 ng
RT: 12.101 min Scan# 953
Delta R.T. -0.004 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:142 Resp: 58819
Ion Ratio Lower Upper
142 100
141 88.7 69.8 104.8
115 35.2 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 42159
Ion Ratio Lower Upper
164 100
162 99.4 79.0 118.4
160 49.1 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.366 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:330 Resp: 3794
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 44.9 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.407 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.012 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	71974
Ion	Ratio	Lower	Upper
172	100		
171	35.2	27.8	41.6
170	24.5	18.2	27.2



#16
Acenaphthylene
Concen: 0.432 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.012 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:	152	Resp:	73996
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.2	22.8
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.394 ng
RT: 14.358 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	154	Resp:	47965
Ion	Ratio	Lower	Upper
154	100		
153	113.5	89.3	133.9
152	62.5	42.2	63.4



#18
Fluorene
Concen: 0.396 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:	166	Resp:	57860
Ion	Ratio	Lower	Upper
166	100		
165	97.9	76.7	115.1
167	13.5	10.9	16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188	Resp:	76951
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	9.1	10.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.245 ng
RT: 15.436 min Scan# 1275
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:198	Resp:	875
Ion Ratio	Lower	Upper
198	100	
182	26.0	26.6
77	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.413 ng
RT: 16.251 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	106.1	80.5	120.7
141	30.3	34.8	52.2#



#22
Hexachlorobenzene
Concen: 0.462 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.8	22.6	34.0#
249	32.3	24.4	36.6









#23
Atrazine
Concen: 0.321 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.012 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 7936
Ion Ratio Lower Upper
200 100
173 28.0 18.7 28.1
215 48.0 41.6 62.4



#24
Pentachlorophenol
Concen: 0.439 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:266 Resp: 4463
Ion Ratio Lower Upper
266 100
264 62.3 50.6 75.8
268 63.3 52.6 79.0









#25
Phenanthrene
Concen: 0.408 ng
RT: 17.091 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 93454
Ion Ratio Lower Upper
178 100
176 18.9 14.4 21.6
179 15.6 12.2 18.4



#26
Anthracene
Concen: 0.387 ng
RT: 17.176 min Scan# 1416
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:178 Resp: 73757
Ion Ratio Lower Upper
178 100
176 18.3 13.9 20.9
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.366 ng
RT: 19.088 min Scan# 1701
Delta R.T. -0.005 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 77328
Ion Ratio Lower Upper
212 100
106 10.7 9.8 14.6
104 6.0 5.4 8.2



#28
Fluoranthene
Concen: 0.408 ng
RT: 19.118 min Scan# 1707
Delta R.T. -0.005 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:202 Resp: 102387
Ion Ratio Lower Upper
202 100
101 9.1 8.5 12.7
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 74013
Ion Ratio Lower Upper
240 100
120 9.1 5.3 7.9#
236 26.5 19.8 29.8



#30
Pyrene
Concen: 0.427 ng
RT: 19.480 min Scan# 1780
Delta R.T. -0.005 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:202 Resp: 98978
Ion Ratio Lower Upper
202 100
200 20.5 15.8 23.6
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.395 ng
 RT: 19.698 min Scan# 1824
 Delta R.T. -0.005 min
 Lab File: BN016486.D
 Acq: 25 Sep 2021 22:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244	Resp:	68036
Ion Ratio	Lower	Upper
244	100	
212	7.6	5.5 8.3
122	9.5	9.0 13.4



#32
 Benzo(a)anthracene
 Concen: 0.376 ng
 RT: 21.238 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: BN016486.D
 Acq: 25 Sep 2021 22:59

Tgt Ion:228	Resp:	80688
Ion Ratio	Lower	Upper
228	100	
226	26.1	20.8 31.2
229	19.6	15.8 23.6









#33
Chrysene
Concen: 0.434 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	100041
Ion Ratio	Lower	Upper	
228	100		
226	28.5	23.0	34.4
229	19.9	15.6	23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.313 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.010 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:	149	Resp:	18334
Ion Ratio	Lower	Upper	
149	100		
167	26.6	23.4	35.0
279	4.0	3.9	5.9









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.515 min Scan# 2436
Delta R.T. -0.003 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	264	Resp:	64288
Ion Ratio	Lower	Upper	
264	100		
260	24.7	18.9	28.3
265	23.1	23.7	35.5#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.350 ng
RT: 25.816 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:	276	Resp:	72816
Ion Ratio	Lower	Upper	
276	100		
138	22.4	19.3	28.9
277	24.9	20.2	30.2









#37
Benzo(b)fluoranthene
Concen: 0.352 ng
RT: 22.834 min Scan# 2237
Delta R.T. -0.003 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 75598
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 8.5 8.4 12.6



#38
Benzo(k)fluoranthene
Concen: 0.442 ng
RT: 22.879 min Scan# 2250
Delta R.T. -0.003 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Tgt Ion:252 Resp: 91711
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.2
125 10.2 8.2 12.4





#39

Benzo(a)pyrene

Concen: 0.363 ng

RT: 23.416 min Scan# 2407

Delta R.T. -0.003 min

Lab File: BN016486.D

Acq: 25 Sep 2021 22:59





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion:252	Resp:	63504
Ion Ratio	Lower	Upper
252	100	
253	22.2	18.7 28.1
125	9.6	9.9 14.9#



#40

Dibenzo(a,h)anthracene

Concen: 0.334 ng




RT: 25.833 min Scan# 3113


Delta R.T. -0.003 min

Lab File: BN016486.D

Acq: 25 Sep 2021 22:59

Tgt Ion:278	Resp:	59190
Ion Ratio	Lower	Upper
278	100	
139	16.2	14.2 21.2
279	24.1	21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.318 ng
RT: 26.510 min Scan# 3311
Delta R.T. -0.003 min
Lab File: BN016486.D
Acq: 25 Sep 2021 22:59

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	57868
Ion Ratio	Lower	Upper	
276	100		
277	23.8	20.5	30.7
138	18.7	18.2	27.4

