

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
 Data File : BN016522.D
 Acq On : 26 Sep 2021 22:26
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Sep 27 06:42:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.661	152	10183	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	23530	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	6988	0.400	ng	# 0.00
19) Phenanthrene-d10	17.042	188	14592	0.400	ng	0.00
29) Chrysene-d12	21.249	240	15696	0.400	ng	0.00
35) Perylene-d12	23.512	264	16158	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	1858	0.081	ng	0.00
5) Phenol-d6	6.840	99	5198	0.193	ng	0.00
8) Nitrobenzene-d5	8.797	82	13263	0.796	ng	-0.01
11) 2-Methylnaphthalene-d10	12.025	152	8902	0.255	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	1063	0.534	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	14615	0.498	ng	-0.01
27) Fluoranthene-d10	19.083	212	16986	0.424	ng	0.00
31) Terphenyl-d14	19.698	244	12974	0.356	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	204	0.047	ng	# 75
6) bis(2-Chloroethyl)ether	7.108	93	8555	0.314	ng	# 81
9) Naphthalene	10.483	128	24276	0.387	ng	98
10) Hexachlorobutadiene	10.774	225	13313	1.067	ng	# 99
12) 2-Methylnaphthalene	12.101	142	10544	0.265	ng	# 93
16) Acenaphthylene	14.002	152	13626	0.480	ng	99
17) Acenaphthene	14.358	154	8628	0.427	ng	# 57
18) Fluorene	15.347	166	10779	0.445	ng	88
20) 4,6-Dinitro-2-methylph...	15.423	198	327	0.338	ng	96
21) 4-Bromophenyl-phenylether	16.239	248	3077	0.361	ng	# 82
22) Hexachlorobenzene	16.348	284	6202	0.735	ng	# 87
23) Atrazine	16.519	200	2087	0.394	ng	96
24) Pentachlorophenol	16.701	266	1395	0.566	ng	99
25) Phenanthrene	17.078	178	16780	0.386	ng	97
26) Anthracene	17.176	178	12744	0.352	ng	98
28) Fluoranthene	19.113	202	20744	0.436	ng	99
30) Pyrene	19.475	202	19106	0.389	ng	99
32) Benzo(a)anthracene	21.227	228	17204	0.378	ng	98
33) Chrysene	21.281	228	21725	0.444	ng	97
34) Bis(2-ethylhexyl)phtha...	21.185	149	8269	0.666	ng	# 94
36) Indeno(1,2,3-cd)pyrene	25.805	276	24337	0.465	ng	99
37) Benzo(b)fluoranthene	22.831	252	18289	0.339	ng	99
38) Benzo(k)fluoranthene	22.875	252	29734	0.570	ng	98
39) Benzo(a)pyrene	23.409	252	14924	0.339	ng	97
40) Dibenzo(a,h)anthracene	25.826	278	21279	0.477	ng	97
41) Benzo(g,h,i)perylene	26.500	276	13402	0.293	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
Data File : BN016522.D
Acq On : 26 Sep 2021 22:26
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Sep 27 06:42:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.661 min Scan# 497
 Delta R.T. -0.009 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:152 Resp: 10183
 Ion Ratio Lower Upper
 152 100
 150 157.4 125.6 188.4
 115 58.2 50.2 75.2



#2
 1,4-Dioxane
 Concen: 0.047 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26

Tgt Ion: 88 Resp: 204
 Ion Ratio Lower Upper
 88 100
 43 0.0 20.3 30.5#
 58 71.6 69.9 104.9









#4
2-Fluorophenol
Concen: 0.081 ng
RT: 5.273 min Scan# 238
Delta R.T. -0.009 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 112 Resp: 1858
Ion Ratio Lower Upper
112 100
64 60.1 47.3 70.9
63 34.1 23.4 35.0



#5
Phenol-d6
Concen: 0.193 ng
RT: 6.840 min Scan# 408
Delta R.T. -0.009 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion: 99 Resp: 5198
Ion Ratio Lower Upper
99 100
42 37.1 14.0 21.0#
71 31.9 23.8 35.6









#6
bis(2-Chloroethyl)ether
Concen: 0.314 ng
RT: 7.108 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 93 Resp: 8555
Ion Ratio Lower Upper
93 100
63 101.7 63.4 95.0#
95 32.6 26.8 40.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion: 136 Resp: 23530
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 9.0 5.3 7.9#
68 5.1 3.9 5.9









#8
Nitrobenzene-d5
Concen: 0.796 ng
RT: 8.797 min Scan# 606
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 82 Resp: 13263
Ion Ratio Lower Upper
82 100
128 30.1 31.7 47.5#
54 80.5 46.4 69.6#



#9
Naphthalene
Concen: 0.387 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion: 128 Resp: 24276
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 14.0 10.1 15.1









#10
Hexachlorobutadiene
Concen: 1.067 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:225 Resp: 13313
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.5 77.3



#11
2-Methylnaphthalene-d10
Concen: 0.255 ng
RT: 12.025 min Scan# 936
Delta R.T. -0.009 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:152 Resp: 8902
Ion Ratio Lower Upper
152 100
151 21.5 16.1 24.1









#12
2-Methylnaphthalene
Concen: 0.265 ng
RT: 12.101 min Scan# 953
Delta R.T. -0.004 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:142 Resp: 10544
Ion Ratio Lower Upper
142 100
141 92.0 69.8 104.8
115 38.3 24.4 36.6#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:164 Resp: 6988
Ion Ratio Lower Upper
164 100
162 103.4 79.0 118.4
160 110.1 34.1 51.1#









#14
2,4,6-Tribromophenol
Concen: 0.534 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:330 Resp: 1063
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 48.0 25.0 37.4#



#15
2-Fluorobiphenyl
Concen: 0.498 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:172 Resp: 14615
Ion Ratio Lower Upper
172 100
171 35.3 27.8 41.6
170 25.5 18.2 27.2









#16
Acenaphthylene
Concen: 0.480 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.2	22.8
153	12.9	10.5	15.7



#17
Acenaphthene
Concen: 0.427 ng
RT: 14.358 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion	Ratio	Lower	Upper
154	100		
153	125.4	89.3	133.9
152	127.9	42.2	63.4









#18
Fluorene
Concen: 0.445 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:166 Resp: 10779
Ion Ratio Lower Upper
166 100
165 109.2 76.7 115.1
167 13.4 10.9 16.3



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:188 Resp: 14592
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.4 7.0 10.6









#20
4,6-Dinitro-2-methylphenol
Concen: 0.338 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.013 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:198 Resp: 327
Ion Ratio Lower Upper
198 100
182 30.8 26.6 39.8
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.361 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:248 Resp: 3077
Ion Ratio Lower Upper
248 100
250 94.2 80.5 120.7
141 72.2 34.8 52.2#









#22
Hexachlorobenzene
Concen: 0.735 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:284 Resp: 6202
Ion Ratio Lower Upper
284 100
142 40.5 22.6 34.0#
249 32.7 24.4 36.6



#23
Atrazine
Concen: 0.394 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:200 Resp: 2087
Ion Ratio Lower Upper
200 100
173 26.4 18.7 28.1
215 50.0 41.6 62.4









#24
Pentachlorophenol
Concen: 0.566 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	266	Resp:	1395
Ion	Ratio	Lower	Upper
266	100		
264	62.4	50.6	75.8
268	65.0	52.6	79.0



#25
Phenanthrene
Concen: 0.386 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.012 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26


Tgt Ion:	178	Resp:	16780
Ion	Ratio	Lower	Upper
178	100		
176	19.7	14.4	21.6
179	16.2	12.2	18.4








#26
 Anthracene
 Concen: 0.352 ng
 RT: 17.176 min Scan# 1416
 Delta R.T. 0.000 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26

Instrument :
 BNA_N
 ClientSampled :







Tgt Ion:	178	Resp:	12744
Ion Ratio	Lower	Upper	
178	100		
176	19.1	13.9	20.9
179	15.3	12.2	18.4



#27
 Fluoranthene-d10
 Concen: 0.424 ng
 RT: 19.083 min Scan# 1700
 Delta R.T. -0.010 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26

Tgt Ion:	212	Resp:	16986
Ion Ratio	Lower	Upper	
212	100		
106	11.2	9.8	14.6
104	6.6	5.4	8.2









#28
Fluoranthene
Concen: 0.436 ng
RT: 19.113 min Scan# 1706
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:202 Resp: 20744
Ion Ratio Lower Upper
202 100
101 9.6 8.5 12.7
203 17.3 13.8 20.8



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:240 Resp: 15696
Ion Ratio Lower Upper
240 100
120 7.5 5.3 7.9
236 26.7 19.8 29.8









#30
Pyrene
Concen: 0.389 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:202	Resp:	19106
Ion Ratio	Lower	Upper
202	100	
200	20.8	15.8 23.6
203	17.3	13.9 20.9



#31
Terphenyl-d14
Concen: 0.356 ng
RT: 19.698 min Scan# 1824
Delta R.T. -0.005 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:244	Resp:	12974
Ion Ratio	Lower	Upper
244	100	
212	7.9	5.5 8.3
122	9.9	9.0 13.4









#32
Benzo(a)anthracene
Concen: 0.378 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:228	Resp:	17204
Ion Ratio	Lower	Upper
228	100	
226	28.1	20.8 31.2
229	19.6	15.8 23.6



#33
Chrysene
Concen: 0.444 ng
RT: 21.281 min Scan# 2006
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:228	Resp:	21725
Ion Ratio	Lower	Upper
228	100	
226	31.0	23.0 34.4
229	19.5	15.6 23.4









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.666 ng
 RT: 21.185 min Scan# 1997
 Delta R.T. -0.010 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:149 Resp: 8269
 Ion Ratio Lower Upper
 149 100
 167 25.6 23.4 35.0
 279 3.8 3.9 5.9#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.512 min Scan# 2435
 Delta R.T. -0.007 min
 Lab File: BN016522.D
 Acq: 26 Sep 2021 22:26

Tgt Ion:264 Resp: 16158
 Ion Ratio Lower Upper
 264 100
 260 25.1 18.9 28.3
 265 26.2 23.7 35.5









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.465 ng
RT: 25.805 min Scan# 3105
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:276 Resp: 24337
Ion Ratio Lower Upper
276 100
138 24.9 19.3 28.9
277 24.8 20.2 30.2



#37
Benzo(b)fluoranthene
Concen: 0.339 ng
RT: 22.831 min Scan# 2236
Delta R.T. -0.007 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:252 Resp: 18289
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.2
125 9.5 8.4 12.6









#38
Benzo(k)fluoranthene
Concen: 0.570 ng
RT: 22.875 min Scan# 2249
Delta R.T. -0.007 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:252 Resp: 29734
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.2
125 11.3 8.2 12.4



#39
Benzo(a)pyrene
Concen: 0.339 ng
RT: 23.409 min Scan# 2405
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:252 Resp: 14924
Ion Ratio Lower Upper
252 100
253 22.5 18.7 28.1
125 10.6 9.9 14.9









#40
Dibenzo(a,h)anthracene
Concen: 0.477 ng
RT: 25.826 min Scan# 3111
Delta R.T. -0.010 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Instrument :
BNA_N
ClientSampled :

Tgt Ion:	278	Resp:	21279
Ion Ratio	Lower	Upper	
278	100		
139	17.0	14.2	21.2
279	24.4	21.3	31.9



#41
Benzo(g,h,i)perylene
Concen: 0.293 ng
RT: 26.500 min Scan# 3308
Delta R.T. -0.013 min
Lab File: BN016522.D
Acq: 26 Sep 2021 22:26

Tgt Ion:	276	Resp:	13402
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.5	30.7
138	21.0	18.2	27.4

