

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
 Data File : BN016526.D
 Acq On : 27 Sep 2021 01:26
 Operator : CG/JU
 Sample : PB139321BS
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Sep 27 05:33:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.661	152	12265	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	29458	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	9324	0.400	ng	# 0.00
19) Phenanthrene-d10	17.042	188	19685	0.400	ng	0.00
29) Chrysene-d12	21.249	240	20031	0.400	ng	0.00
35) Perylene-d12	23.508	264	19716	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	1779	0.065	ng	0.00
5) Phenol-d6	6.840	99	5706	0.176	ng	0.00
8) Nitrobenzene-d5	8.810	82	16265	0.780	ng	0.00
11) 2-Methylnaphthalene-d10	12.025	152	11601	0.266	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	1395	0.528	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	19370	0.495	ng	-0.01
27) Fluoranthene-d10	19.083	212	20560	0.381	ng	0.00
31) Terphenyl-d14	19.693	244	15839	0.340	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	176	0.033	ng	# 83
6) bis(2-Chloroethyl)ether	7.108	93	9939	0.302	ng	# 82
9) Naphthalene	10.483	128	31396	0.400	ng	98
10) Hexachlorobutadiene	10.774	225	16165	1.035	ng	# 99
12) 2-Methylnaphthalene	12.101	142	14161	0.284	ng	# 94
16) Acenaphthylene	14.002	152	18252	0.481	ng	100
17) Acenaphthene	14.358	154	12041	0.447	ng	# 57
18) Fluorene	15.347	166	15063	0.466	ng	91
20) 4,6-Dinitro-2-methylph...	15.423	198	488	0.358	ng	# 87
21) 4-Bromophenyl-phenylether	16.239	248	4113	0.358	ng	# 82
22) Hexachlorobenzene	16.348	284	8569	0.753	ng	# 87
23) Atrazine	16.519	200	2622	0.376	ng	95
24) Pentachlorophenol	16.701	266	1927	0.573	ng	99
25) Phenanthrene	17.078	178	23481	0.401	ng	98
26) Anthracene	17.176	178	17758	0.364	ng	98
28) Fluoranthene	19.113	202	26127	0.407	ng	99
30) Pyrene	19.480	202	22561	0.360	ng	99
32) Benzo(a)anthracene	21.227	228	21634	0.372	ng	97
33) Chrysene	21.281	228	28395	0.455	ng	97
34) Bis(2-ethylhexyl)phtha...	21.185	149	9602	0.606	ng	# 94
36) Indeno(1,2,3-cd)pyrene	25.805	276	31775	0.497	ng	98
37) Benzo(b)fluoranthene	22.827	252	22261	0.338	ng	98
38) Benzo(k)fluoranthene	22.872	252	40692	0.639	ng	97
39) Benzo(a)pyrene	23.409	252	17859	0.333	ng	96
40) Dibenzo(a,h)anthracene	25.822	278	27182	0.500	ng	97
41) Benzo(g,h,i)perylene	26.500	276	15980	0.286	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
Data File : BN016526.D
Acq On : 27 Sep 2021 01:26
Operator : CG/JU
Sample : PB139321BS
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Sep 27 05:33:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.661 min Scan# 497
 Delta R.T. -0.009 min
 Lab File: BN016526.D
 Acq: 27 Sep 2021 01:26




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:152 Resp: 12265
 Ion Ratio Lower Upper
 152 100
 150 155.4 125.6 188.4
 115 57.2 50.2 75.2



#2
 1,4-Dioxane
 Concen: 0.033 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN016526.D
 Acq: 27 Sep 2021 01:26

Tgt Ion: 88 Resp: 176
 Ion Ratio Lower Upper
 88 100
 43 52.8 20.3 30.5#
 58 81.3 69.9 104.9









#4
2-Fluorophenol
Concen: 0.065 ng
RT: 5.273 min Scan# 238
Delta R.T. -0.009 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:	112	Resp:	1779
Ion	Ratio	Lower	Upper
112	100		
64	56.0	47.3	70.9
63	32.5	23.4	35.0



#5
Phenol-d6
Concen: 0.176 ng
RT: 6.840 min Scan# 408
Delta R.T. -0.009 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:	99	Resp:	5706
Ion	Ratio	Lower	Upper
99	100		
42	36.9	14.0	21.0#
71	33.1	23.8	35.6









#6
bis(2-Chloroethyl)ether
Concen: 0.302 ng
RT: 7.108 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion: 93 Resp: 9939
Ion Ratio Lower Upper
93 100
63 100.9 63.4 95.0#
95 33.1 26.8 40.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:136 Resp: 29458
Ion Ratio Lower Upper
136 100
137 11.2 8.8 13.2
54 9.2 5.3 7.9#
68 5.2 3.9 5.9









#8
Nitrobenzene-d5
Concen: 0.780 ng
RT: 8.810 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
82	100		
128	42.8	31.7	47.5
54	64.2	46.4	69.6



#9
Naphthalene
Concen: 0.400 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.9	10.1	15.1









#10
Hexachlorobutadiene
Concen: 1.035 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:225 Resp: 16165
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.5 77.3



#11
2-Methylnaphthalene-d10
Concen: 0.266 ng
RT: 12.025 min Scan# 936
Delta R.T. -0.009 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:152 Resp: 11601
Ion Ratio Lower Upper
152 100
151 21.1 16.1 24.1









#12
2-Methylnaphthalene
Concen: 0.284 ng
RT: 12.101 min Scan# 953
Delta R.T. -0.004 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:142 Resp: 14161
Ion Ratio Lower Upper
142 100
141 90.3 69.8 104.8
115 38.0 24.4 36.6#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:164 Resp: 9324
Ion Ratio Lower Upper
164 100
162 103.5 79.0 118.4
160 115.7 34.1 51.1#









#14
2,4,6-Tribromophenol
Concen: 0.528 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 330 Resp: 1395
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 51.3 25.0 37.4#



#15
2-Fluorobiphenyl
Concen: 0.495 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion: 172 Resp: 19370
Ion Ratio Lower Upper
172 100
171 35.1 27.8 41.6
170 24.8 18.2 27.2









#16
Acenaphthylene
Concen: 0.481 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.2	22.8
153	12.8	10.5	15.7



#17
Acenaphthene
Concen: 0.447 ng
RT: 14.358 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion	Ratio	Lower	Upper
154	100		
153	124.0	89.3	133.9
152	129.9	42.2	63.4









#18
Fluorene
Concen: 0.466 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:166 Resp: 15063
Ion Ratio Lower Upper
166 100
165 106.3 76.7 115.1
167 13.4 10.9 16.3



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:188 Resp: 19685
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.4 7.0 10.6









#20
4,6-Dinitro-2-methylphenol
Concen: 0.358 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.013 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:198 Resp: 488
Ion Ratio Lower Upper
198 100
182 25.6 26.6 39.8#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.358 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:248 Resp: 4113
Ion Ratio Lower Upper
248 100
250 95.1 80.5 120.7
141 74.1 34.8 52.2#









#22
Hexachlorobenzene
Concen: 0.753 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
284	100		
142	40.7	22.6	34.0#
249	32.4	24.4	36.6



#23
Atrazine
Concen: 0.376 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.8	18.7	28.1
215	48.3	41.6	62.4









#24
Pentachlorophenol
Concen: 0.573 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	266	Resp:	1927
Ion Ratio	Lower	Upper	
266	100		
264	64.3	50.6	75.8
268	66.8	52.6	79.0



#25
Phenanthrene
Concen: 0.401 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.012 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26


Tgt Ion:	178	Resp:	23481
Ion Ratio	Lower	Upper	
178	100		
176	19.4	14.4	21.6
179	15.5	12.2	18.4








#26
 Anthracene
 Concen: 0.364 ng
 RT: 17.176 min Scan# 1416
 Delta R.T. 0.000 min
 Lab File: BN016526.D
 Acq: 27 Sep 2021 01:26

Instrument :
 BNA_N
 ClientSampled :







Tgt Ion:	178	Resp:	17758
Ion Ratio	Lower	Upper	
178	100		
176	18.8	13.9	20.9
179	15.4	12.2	18.4



#27
 Fluoranthene-d10
 Concen: 0.381 ng
 RT: 19.083 min Scan# 1700
 Delta R.T. -0.010 min
 Lab File: BN016526.D
 Acq: 27 Sep 2021 01:26

Tgt Ion:	212	Resp:	20560
Ion Ratio	Lower	Upper	
212	100		
106	11.2	9.8	14.6
104	6.3	5.4	8.2









#28
Fluoranthene
Concen: 0.407 ng
RT: 19.113 min Scan# 1706
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:202	Resp:	26127
Ion Ratio	Lower	Upper
202	100	
101	9.7	8.5 12.7
203	17.4	13.8 20.8



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:240	Resp:	20031
Ion Ratio	Lower	Upper
240	100	
120	7.2	5.3 7.9
236	26.8	19.8 29.8









#30
Pyrene
Concen: 0.360 ng
RT: 19.480 min Scan# 1780
Delta R.T. -0.005 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampled :


Tgt Ion:	202	Resp:	22561
Ion Ratio	Lower	Upper	
202	100		
200	20.7	15.8	23.6
203	17.6	13.9	20.9



#31
Terphenyl-d14
Concen: 0.340 ng
RT: 19.693 min Scan# 1823
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:	244	Resp:	15839
Ion Ratio	Lower	Upper	
244	100		
212	8.2	5.5	8.3
122	11.7	9.0	13.4









#32
Benzo(a)anthracene
Concen: 0.372 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:228	Resp:	21634
Ion Ratio	Lower	Upper
228	100	
226	28.2	20.8 31.2
229	19.1	15.8 23.6



#33
Chrysene
Concen: 0.455 ng
RT: 21.281 min Scan# 2006
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:228	Resp:	28395
Ion Ratio	Lower	Upper
228	100	
226	30.7	23.0 34.4
229	19.0	15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.606 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:149 Resp: 9602
Ion Ratio Lower Upper
149 100
167 25.7 23.4 35.0
279 3.7 3.9 5.9#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.508 min Scan# 2434
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:264 Resp: 19716
Ion Ratio Lower Upper
264 100
260 24.9 18.9 28.3
265 24.8 23.7 35.5









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.497 ng
RT: 25.805 min Scan# 3105
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:276 Resp: 31775
Ion Ratio Lower Upper
276 100
138 25.4 19.3 28.9
277 24.9 20.2 30.2



#37
Benzo(b)fluoranthene
Concen: 0.338 ng
RT: 22.827 min Scan# 2235
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:252 Resp: 22261
Ion Ratio Lower Upper
252 100
253 21.8 18.2 27.2
125 9.4 8.4 12.6









#38
Benzo(k)fluoranthene
Concen: 0.639 ng
RT: 22.872 min Scan# 2248
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:252 Resp: 40692
Ion Ratio Lower Upper
252 100
253 21.8 18.2 27.2
125 12.3 8.2 12.4



#39
Benzo(a)pyrene
Concen: 0.333 ng
RT: 23.409 min Scan# 2405
Delta R.T. -0.010 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:252 Resp: 17859
Ion Ratio Lower Upper
252 100
253 21.9 18.7 28.1
125 10.4 9.9 14.9









#40
Dibenzo(a,h)anthracene
Concen: 0.500 ng
RT: 25.822 min Scan# 3110
Delta R.T. -0.013 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Instrument :
BNA_N
ClientSampleId :

Tgt Ion:278 Resp: 27182
Ion Ratio Lower Upper
278 100
139 18.0 14.2 21.2
279 23.9 21.3 31.9



#41
Benzo(g,h,i)perylene
Concen: 0.286 ng
RT: 26.500 min Scan# 3308
Delta R.T. -0.013 min
Lab File: BN016526.D
Acq: 27 Sep 2021 01:26

Tgt Ion:276 Resp: 15980
Ion Ratio Lower Upper
276 100
277 24.6 20.5 30.7
138 20.9 18.2 27.4

