




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
Data File : BN016504.D
Acq On : 26 Sep 2021 10:26
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 37 Sample Multiplier: 1

Quant Time: Sep 27 07:00:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:04:53 2021
Response via : Initial Calibration



TIC: BN016504.D\data.ms

(70) Pentachlorophenol (C)




16.698min (-0.006) 16224.85 ng

response 88366

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.73
264.00	61.60	57.84
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
Data File : BN016504.D
Acq On : 26 Sep 2021 10:26
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 37 Sample Multiplier: 1

Quant Time: Sep 27 07:00:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 13:04:53 2021
Response via : Initial Calibration



TIC: BN016504.D\data.ms

(77) Benzidine

19.310min (-0.012) 0.00 ng

response 256479

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.96
183.00	13.20	11.92
0.00	0.00	0.00