

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
 Data File : BN016484.D
 Acq On : 25 Sep 2021 20:34
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 26 04:16:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	21163	0.40	ng	0.00
7) Naphthalene-d8	10.432	136	87428	0.40	ng	#-0.01
13) Acenaphthene-d10	14.294	164	43907	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	83764	0.40	ng	0.00
29) Chrysene-d12	21.248	240	82319	0.40	ng	# 0.00
35) Perylene-d12	23.515	264	74413	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	19154	0.40	ng	0.00
5) Phenol-d6	6.849	99	22380	0.40	ng	0.00
8) Nitrobenzene-d5	8.809	82	26717	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	12.029	152	51893	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	4252	0.38	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	73138	0.40	ng	-0.01
27) Fluoranthene-d10	19.087	212	92939	0.40	ng	0.00
31) Terphenyl-d14	19.698	244	76602	0.40	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	2699	0.30	ng	# 80
3) n-Nitrosodimethylamine	3.631	42	7954	0.36	ng	# 71
6) bis(2-Chloroethyl)ether	7.107	93	24115	0.43	ng	90
9) Naphthalene	10.482	128	92023	0.39	ng	99
10) Hexachlorobutadiene	10.774	225	19566	0.42	ng	# 99
12) 2-Methylnaphthalene	12.100	142	59672	0.40	ng	96
16) Acenaphthylene	14.002	152	77009	0.43	ng	100
17) Acenaphthene	14.357	154	50247	0.40	ng	96
18) Fluorene	15.347	166	61247	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.436	198	770	0.23	ng	# 87
21) 4-Bromophenyl-phenylether	16.250	248	19634	0.40	ng	# 94
22) Hexachlorobenzene	16.348	284	20934	0.43	ng	# 89
23) Atrazine	16.518	200	10087	0.35	ng	# 91
24) Pentachlorophenol	16.701	266	4657	0.43	ng	99
25) Phenanthrene	17.090	178	99901	0.40	ng	99
26) Anthracene	17.175	178	81036	0.39	ng	99
28) Fluoranthene	19.117	202	115578	0.42	ng	99
30) Pyrene	19.480	202	112330	0.44	ng	99
32) Benzo(a)anthracene	21.238	228	97362	0.41	ng	100
33) Chrysene	21.291	228	110796	0.43	ng	100
34) Bis(2-ethylhexyl)phtha...	21.185	149	26157	0.40	ng	96
36) Indeno(1,2,3-cd)pyrene	25.812	276	85966	0.36	ng	99
37) Benzo(b)fluoranthene	22.834	252	92692	0.37	ng	98
38) Benzo(k)fluoranthene	22.878	252	102761	0.43	ng	98
39) Benzo(a)pyrene	23.416	252	78738	0.39	ng	# 96
40) Dibenzo(a,h)anthracene	25.832	278	70513	0.34	ng	96
41) Benzo(g,h,i)perylene	26.510	276	69405	0.33	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092521\
Data File : BN016484.D
Acq On : 25 Sep 2021 20:34
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Sep 26 04:16:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.670 min Scan# 498
 Delta R.T. -0.000 min
 Lab File: BN016484.D
 Acq: 25 Sep 2021 20:34

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:152 Resp: 21163




Ion	Ratio	Lower	Upper
152	100		
150	153.0	125.6	188.4
115	57.0	50.2	75.2




#2
 1,4-Dioxane
 Concen: 0.30 ng
 RT: 3.281 min Scan# 22
 Delta R.T. -0.000 min
 Lab File: BN016484.D
 Acq: 25 Sep 2021 20:34

Tgt Ion: 88 Resp: 2699

Ion	Ratio	Lower	Upper
88	100		
43	42.7	20.3	30.5
58	102.5	69.9	104.9









#3
n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.631 min Scan# 60
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 7954
Ion Ratio Lower Upper
42 100
74 105.8 113.6 170.4#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.291 min Scan# 240
Delta R.T. 0.009 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion: 112 Resp: 19154
Ion Ratio Lower Upper
112 100
64 64.6 47.3 70.9
63 34.4 23.4 35.0









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.849 min Scan# 409
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 22380
Ion Ratio Lower Upper
99 100
42 27.3 14.0 21.0#
71 32.5 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.43 ng
RT: 7.107 min Scan# 437
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion: 93 Resp: 24115
Ion Ratio Lower Upper
93 100
63 90.7 63.4 95.0
95 32.8 26.8 40.2









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.013 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	87428
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	10.0	5.3	7.9#
68	5.7	3.9	5.9



#8
Nitrobenzene-d5
Concen: 0.43 ng
RT: 8.809 min Scan# 607
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:	82	Resp:	26717
Ion Ratio	Lower	Upper	
82	100		
128	41.2	31.7	47.5
54	66.1	46.4	69.6









#9
Naphthalene
Concen: 0.39 ng
RT: 10.482 min Scan# 739
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	92023
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.9	13.3
127	13.0	10.1	15.1



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:	225	Resp:	19566
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.029 min Scan# 937
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 51893
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.100 min Scan# 953
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:142 Resp: 59672
Ion Ratio Lower Upper
142 100
141 89.0 69.8 104.8
115 35.0 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 43907
Ion Ratio Lower Upper
164 100
162 99.4 79.0 118.4
160 48.0 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.38 ng
RT: 15.791 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:330 Resp: 4252
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 44.8 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.013 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172	Resp:	73138
Ion Ratio	Lower	Upper
172	100	
171	35.4	27.8 41.6
170	24.2	18.2 27.2



#16
Acenaphthylene
Concen: 0.43 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.013 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:152	Resp:	77009
Ion Ratio	Lower	Upper
152	100	
151	19.3	15.2 22.8
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.357 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 50247
Ion Ratio Lower Upper
154 100
153 112.4 89.3 133.9
152 59.8 42.2 63.4



#18
Fluorene
Concen: 0.40 ng
RT: 15.347 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:166 Resp: 61247
Ion Ratio Lower Upper
166 100
165 97.4 76.7 115.1
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 83764
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.3 7.0 10.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.23 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:198 Resp: 770
Ion Ratio Lower Upper
198 100
182 25.9 26.6 39.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.250 min Scan# 1340
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 19634
Ion Ratio Lower Upper
248 100
250 102.5 80.5 120.7
141 32.9 34.8 52.2#



#22
Hexachlorobenzene
Concen: 0.43 ng
RT: 16.348 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:284 Resp: 20934
Ion Ratio Lower Upper
284 100
142 38.8 22.6 34.0#
249 32.4 24.4 36.6









#23
Atrazine
Concen: 0.35 ng
RT: 16.518 min Scan# 1362
Delta R.T. -0.013 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 10087
Ion Ratio Lower Upper
200 100
173 29.1 18.7 28.1#
215 47.0 41.6 62.4



#24
Pentachlorophenol
Concen: 0.43 ng
RT: 16.701 min Scan# 1377
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:266 Resp: 4657
Ion Ratio Lower Upper
266 100
264 63.2 50.6 75.8
268 63.8 52.6 79.0









#25
Phenanthrene
Concen: 0.40 ng
RT: 17.090 min Scan# 1409
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 99901
Ion Ratio Lower Upper
178 100
176 18.9 14.4 21.6
179 15.5 12.2 18.4



#26
Anthracene
Concen: 0.39 ng
RT: 17.175 min Scan# 1416
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:178 Resp: 81036
Ion Ratio Lower Upper
178 100
176 18.2 13.9 20.9
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.087 min Scan# 1701
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 92939
Ion Ratio Lower Upper
212 100
106 10.6 9.8 14.6
104 6.0 5.4 8.2



#28
Fluoranthene
Concen: 0.42 ng
RT: 19.117 min Scan# 1707
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:202 Resp: 115578
Ion Ratio Lower Upper
202 100
101 9.3 8.5 12.7
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.248 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 82319
Ion Ratio Lower Upper
240 100
120 10.0 5.3 7.9#
236 26.7 19.8 29.8



#30
Pyrene
Concen: 0.44 ng
RT: 19.480 min Scan# 1780
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:202 Resp: 112330
Ion Ratio Lower Upper
202 100
200 20.4 15.8 23.6
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.40 ng
RT: 19.698 min Scan# 1824
Delta R.T. -0.005 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 76602
Ion Ratio Lower Upper
244 100
212 7.7 5.5 8.3
122 10.4 9.0 13.4



#32
Benzo(a)anthracene
Concen: 0.41 ng
RT: 21.238 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:228 Resp: 97362
Ion Ratio Lower Upper
228 100
226 26.2 20.8 31.2
229 19.8 15.8 23.6









#33
Chrysene
Concen: 0.43 ng
RT: 21.291 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:228 Resp: 110796




Ion	Ratio	Lower	Upper
228	100		
226	28.7	23.0	34.4
229	19.8	15.6	23.4




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.40 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.011 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:149 Resp: 26157

Ion	Ratio	Lower	Upper
149	100		
167	27.0	23.4	35.0
279	4.1	3.9	5.9









#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.515 min Scan# 2436
 Delta R.T. -0.004 min
 Lab File: BN016484.D
 Acq: 25 Sep 2021 20:34

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:264 Resp: 74413




Ion	Ratio	Lower	Upper
264	100		
260	24.6	18.9	28.3
265	23.4	23.7	35.5#




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.36 ng
 RT: 25.812 min Scan# 3107
 Delta R.T. -0.004 min
 Lab File: BN016484.D
 Acq: 25 Sep 2021 20:34

Tgt Ion:276 Resp: 85966

Ion	Ratio	Lower	Upper
276	100		
138	23.0	19.3	28.9
277	25.0	20.2	30.2









#37
Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 22.834 min Scan# 2237
Delta R.T. -0.004 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 92692
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.2
125 8.6 8.4 12.6



#38
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 22.878 min Scan# 2250
Delta R.T. -0.004 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:252 Resp: 102761
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.2
125 11.7 8.2 12.4









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.416 min Scan# 2407
Delta R.T. -0.004 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 78738
Ion Ratio Lower Upper
252 100
253 22.3 18.7 28.1
125 9.7 9.9 14.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.34 ng
RT: 25.832 min Scan# 3113
Delta R.T. -0.004 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Tgt Ion:278 Resp: 70513
Ion Ratio Lower Upper
278 100
139 17.6 14.2 21.2
279 23.7 21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.33 ng
RT: 26.510 min Scan# 3311
Delta R.T. -0.004 min
Lab File: BN016484.D
Acq: 25 Sep 2021 20:34

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:276 Resp: 69405
Ion Ratio Lower Upper
276 100
277 24.0 20.5 30.7
138 18.7 18.2 27.4

