

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092722\
 Data File : BN021925.D
 Acq On : 26 Sep 2022 19:36
 Operator : CG/JU
 Sample : N4662-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXZ24

Quant Time: Sep 27 00:26:02 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.042	152	8004	0.400	ng/u1	0.00
4) Naphthalene-d8	10.872	136	26165	0.400	ng/u1	#-0.01
9) Acenaphthene-d10	14.698	164	14628	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.451	188	29345	0.400	ng/u1	0.00
17) Chrysene-d12	21.644	240	17398	0.400	ng/u1	# 0.00
23) Perylene-d12	24.147	264	14353	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.359	96	40788	4.142	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.462	152	10665	0.257	ng/u1	-0.01
18) Fluoranthene-d10	19.478	212	24207	0.374	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.397	88	6336	0.627	ng/u1	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092722\
Data File : BN021925.D
Acq On : 26 Sep 2022 19:36
Operator : CG/JU
Sample : N4662-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXZ24

Quant Time: Sep 27 00:26:02 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

