

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
 Data File : BN012011.D
 Acq On : 29 Sep 2020 08:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 09:17:36 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	1852	0.40	ng	0.00
7) Naphthalene-d8	10.440	136	7763	0.40	ng	#-0.01
13) Acenaphthene-d10	14.306	164	4000	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	7693	0.40	ng	0.00
29) Chrysene-d12	21.248	240	6969	0.40	ng	#-0.01
36) Perylene-d12	23.460	264	7328	0.40	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.275	112	2567	0.40	ng	0.00
5) Phenol-d6	6.845	99	3255	0.40	ng	0.00
8) Nitrobenzene-d5	8.818	82	3234	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	4628	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.790	330	843	0.43	ng	-0.01
15) 2-Fluorobiphenyl	12.923	172	5805	0.38	ng	-0.01
27) Fluoranthene-d10	19.087	212	8429	0.37	ng	0.00
31) Terphenyl-d14	19.693	244	6187	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.253	88	1346	0.42	ng	91
3) n-Nitrosodimethylamine	3.559	42	2166	0.47	ng	92
6) bis(2-Chloroethyl)ether	7.103	93	2701	0.39	ng	92
9) Naphthalene	10.491	128	8371	0.38	ng	99
10) Hexachlorobutadiene	10.782	225	1528	0.42	ng	# 96
12) 2-Methylnaphthalene	12.113	142	5306	0.38	ng	95
16) Acenaphthylene	14.014	152	7210	0.37	ng	99
17) Acenaphthene	14.369	154	4912	0.37	ng	93
18) Fluorene	15.359	166	5920	0.37	ng	100
20) 4,6-Dinitro-2-methylph...	15.435	198	584	0.40	ng	# 1
21) 4-Bromophenyl-phenylether	16.250	248	1628	0.40	ng	# 82
22) Hexachlorobenzene	16.360	284	1950	0.41	ng	# 81
23) Atrazine	16.518	200	1574	0.40	ng	96
24) Pentachlorophenol	16.700	266	872	0.38	ng	98
25) Phenanthrene	17.090	178	8650	0.37	ng	99
26) Anthracene	17.175	178	7813	0.38	ng	99
28) Fluoranthene	19.117	202	9623	0.36	ng	# 96
30) Pyrene	19.479	202	9797	0.38	ng	100
32) Benzo(a)anthracene	21.237	228	7965	0.36	ng	98
33) Chrysene	21.280	228	9143	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.174	149	5963	0.38	ng	92
35) Indeno(1,2,3-cd)pyrene	25.667	276	10865	0.40	ng	# 95
37) Benzo(b)fluoranthene	22.799	252	8736	0.35	ng	# 83
38) Benzo(k)fluoranthene	22.844	252	9635	0.35	ng	# 85
39) Benzo(a)pyrene	23.364	252	8388	0.36	ng	# 77
40) Dibenzo(a,h)anthracene	25.685	278	8978	0.38	ng	# 87
41) Benzo(g,h,i)perylene	26.342	276	9629	0.37	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
 Data File : BN012011.D
 Acq On : 29 Sep 2020 08:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 09:17:36 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.675 min Scan# 583
 Delta R.T. -0.008 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 152 Resp: 1852
 Ion Ratio Lower Upper
 152 100
 150 156.6 118.3 177.5
 115 76.2 51.3 76.9



#2
 1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.253 min Scan# 34
 Delta R.T. -0.000 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42

Tgt Ion: 88 Resp: 1346
 Ion Ratio Lower Upper
 88 100
 58 70.7 63.3 94.9
 43 46.1 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.47 ng
RT: 3.559 min Scan# 72
Delta R.T. -0.000 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	42	Resp:	2166
Ion	Ratio	Lower	Upper
42	100		
74	73.1	64.4	96.6
44	7.7	6.7	10.1



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.275 min Scan# 285
Delta R.T. -0.008 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:	112	Resp:	2567
Ion	Ratio	Lower	Upper
112	100		
64	71.6	49.0	73.6
63	46.4	31.8	47.6









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.845 min Scan# 480
Delta R.T. -0.008 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 3255
Ion Ratio Lower Upper
99 100
42 36.0 23.4 35.0#
71 53.8 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.103 min Scan# 512
Delta R.T. -0.008 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion: 93 Resp: 2701
Ion Ratio Lower Upper
93 100
63 85.1 61.1 91.7
95 30.6 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.440 min Scan# 830
Delta R.T. -0.013 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	7763
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.5	14.3
54	11.4	5.6	8.4#
68	6.7	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.818 min Scan# 702
Delta R.T. -0.000 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:	82	Resp:	3234
Ion Ratio	Lower	Upper	
82	100		
128	36.8	34.2	51.2
54	59.1	43.1	64.7









#9
Naphthalene
Concen: 0.38 ng
RT: 10.491 min Scan# 834
Delta R.T. -0.013 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	8371
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.7	14.5
127	13.5	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.782 min Scan# 857
Delta R.T. -0.013 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:	225	Resp:	1528
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.0	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.038 min Scan# 1032
Delta R.T. -0.009 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 4628
Ion Ratio Lower Upper
152 100
151 20.1 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.113 min Scan# 1049
Delta R.T. -0.009 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:142 Resp: 5306
Ion Ratio Lower Upper
142 100
141 88.6 70.0 105.0
115 52.6 35.8 53.6





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.306 min Scan# 1279

Delta R.T. -0.000 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:164	Resp:	4000
Ion Ratio	Lower	Upper
164	100	
162	99.3	83.6 125.4
160	46.5	41.2 61.8



#14

2,4,6-Tribromophenol

Concen: 0.43 ng




RT: 15.790 min Scan# 1396


Delta R.T. -0.013 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:330	Resp:	843
Ion Ratio	Lower	Upper
330	100	
332	0.0	0.0 0.0
141	49.0	33.7 50.5









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.923 min Scan# 1170
Delta R.T. -0.013 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	5805
Ion Ratio	Lower	Upper	
172	100		
171	35.0	27.4	41.2
170	24.1	17.8	26.6



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.014 min Scan# 1256
Delta R.T. -0.013 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:	152	Resp:	7210
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.9	23.9
153	12.7	10.5	15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.369 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 4912
Ion Ratio Lower Upper
154 100
153 112.9 83.4 125.2
152 56.6 42.8 64.2



#18
Fluorene
Concen: 0.37 ng
RT: 15.359 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:166 Resp: 5920
Ion Ratio Lower Upper
166 100
165 98.7 78.8 118.2
167 13.2 11.2 16.8





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.053 min Scan# 1499

Delta R.T. -0.000 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:188 Resp: 7693

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0





80 7.9 7.7 11.5

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#20

4,6-Dinitro-2-methylphenol

Concen: 0.40 ng

RT: 15.435 min Scan# 1368

Delta R.T. -0.013 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42




Tgt Ion:198 Resp: 584


Ion Ratio Lower Upper

198 100

51 190.5 45.0 67.4#

105 57.1 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.250 min Scan# 1433
Delta R.T. -0.000 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1628
Ion Ratio Lower Upper
248 100
250 98.8 82.7 124.1
141 75.0 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.41 ng
RT: 16.360 min Scan# 1442
Delta R.T. -0.012 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:284 Resp: 1950
Ion Ratio Lower Upper
284 100
142 62.0 32.4 48.6#
249 33.0 26.8 40.2









#23
 Atrazine
 Concen: 0.40 ng
 RT: 16.518 min Scan# 1455
 Delta R.T. -0.012 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	200	Resp:	1574
Ion Ratio	Lower	Upper	
200	100		
173	27.6	18.7	28.1
215	49.2	40.2	60.4



#24
 Pentachlorophenol
 Concen: 0.38 ng
 RT: 16.700 min Scan# 1470
 Delta R.T. -0.012 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42

Tgt Ion:	266	Resp:	872
Ion Ratio	Lower	Upper	
266	100		
264	66.6	51.2	76.8
268	65.8	52.4	78.6









#25
Phenanthrene
Concen: 0.37 ng
RT: 17.090 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	178	Resp:	8650
Ion Ratio	Lower	Upper	
178	100		
176	19.3	15.1	22.7
179	15.4	12.3	18.5



#26
Anthracene
Concen: 0.38 ng
RT: 17.175 min Scan# 1509
Delta R.T. -0.012 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:	178	Resp:	7813
Ion Ratio	Lower	Upper	
178	100		
176	18.4	14.6	21.8
179	15.2	12.4	18.6









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.087 min Scan# 1794
Delta R.T. -0.005 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 8429
Ion Ratio Lower Upper
212 100
106 10.9 6.4 9.6#
104 7.0 3.9 5.9#



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.117 min Scan# 1800
Delta R.T. -0.005 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:202 Resp: 9623
Ion Ratio Lower Upper
202 100
101 10.0 5.4 8.2#
203 16.5 13.8 20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.248 min Scan# 2096

Delta R.T. -0.011 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:240 Resp: 6969

Ion Ratio Lower Upper

240 100

120 10.9 3.5 5.3#





236 26.0 20.5 30.7

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#30

Pyrene

Concen: 0.38 ng

RT: 19.479 min Scan# 1873

Delta R.T. -0.005 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42




Tgt Ion:202 Resp: 9797


Ion Ratio Lower Upper

202 100

200 20.6 16.5 24.7


203 17.7 13.9 20.9








#31
 Terphenyl-d14
 Concen: 0.39 ng
 RT: 19.693 min Scan# 1916
 Delta R.T. -0.005 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC







Tgt Ion:	244	Resp:	6187
Ion Ratio	Lower	Upper	
244	100		
212	9.9	7.3	10.9
122	12.6	7.8	11.6#



#32
 Benzo(a)anthracene
 Concen: 0.36 ng
 RT: 21.237 min Scan# 2095
 Delta R.T. -0.000 min
 Lab File: BN012011.D
 Acq: 29 Sep 2020 08:42

Tgt Ion:	228	Resp:	7965
Ion Ratio	Lower	Upper	
228	100		
226	26.7	21.0	31.4
229	21.1	16.0	24.0





#33

Chrysene

Concen: 0.38 ng

RT: 21.280 min Scan# 2099

Delta R.T. -0.011 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:228 Resp: 9143

Ion Ratio Lower Upper

228 100

226 30.4 23.3 34.9





229 20.1 15.8 23.8

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.38 ng

RT: 21.174 min Scan# 2089

Delta R.T. -0.000 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42




Tgt Ion:149 Resp: 5963


Ion Ratio Lower Upper

149 100

167 24.8 23.6 35.4

279 4.6 4.0 6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.40 ng
RT: 25.667 min Scan# 3158
Delta R.T. -0.014 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 10865
Ion Ratio Lower Upper
276 100
138 20.6 13.0 19.4#
277 24.4 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.460 min Scan# 2513
Delta R.T. -0.010 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:264 Resp: 7328
Ion Ratio Lower Upper
264 100
260 23.8 19.6 29.4
265 63.9 38.6 58.0#









#37
Benzo(b)fluoranthene
Concen: 0.35 ng
RT: 22.799 min Scan# 2320
Delta R.T. -0.007 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 8736
Ion Ratio Lower Upper
252 100
253 30.6 19.1 28.7#
125 16.8 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.844 min Scan# 2333
Delta R.T. -0.003 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Tgt Ion:252 Resp: 9635
Ion Ratio Lower Upper
252 100
253 30.5 19.3 28.9#
125 14.4 6.0 9.0#





#39

Benzo(a)pyrene

Concen: 0.36 ng





RT: 23.364 min Scan# 2485

Delta R.T. -0.007 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:	252	Resp:	8388
Ion Ratio	Lower	Upper	
252	100		
253	34.3	19.8	29.6#
125	21.1	7.3	10.9#



#40

Dibenzo(a,h)anthracene

Concen: 0.38 ng




RT: 25.685 min Scan# 3163


Delta R.T. -0.014 min

Lab File: BN012011.D

Acq: 29 Sep 2020 08:42

Tgt Ion:	278	Resp:	8978
Ion Ratio	Lower	Upper	
278	100		
139	15.8	8.6	13.0#
279	31.3	19.8	29.8#





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.342 min Scan# 3355
Delta R.T. -0.017 min
Lab File: BN012011.D
Acq: 29 Sep 2020 08:42

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	276	Resp:	9629
Ion Ratio	Lower	Upper	
276	100		
277	25.5	19.4	29.0
138	20.9	11.3	16.9#

