

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
 Data File : BN011987.D
 Acq On : 28 Sep 2020 16:25
 Operator : CG/JU
 Sample : L4118-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D7-20200918

Quant Time: Sep 28 17:53:25 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	1157	0.40	ng	# 0.00
7) Naphthalene-d8	10.453	136	4785	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	2747	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	5197	0.40	ng	0.00
29) Chrysene-d12	21.248	240	4996	0.40	ng	#-0.01
36) Perylene-d12	23.463	264	5190	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	451	0.11	ng	0.00
5) Phenol-d6	6.853	99	305	0.06	ng	0.00
8) Nitrobenzene-d5	8.818	82	1700	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	12.042	152	2415	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.803	330	451	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	3445	0.33	ng	-0.01
27) Fluoranthene-d10	19.087	212	6100	0.40	ng	0.00
31) Terphenyl-d14	19.693	244	5320	0.47	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.245	88	2504	1.24	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.174	149	920	0.08	ng	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
Data File : BN011987.D
Acq On : 28 Sep 2020 16:25
Operator : CG/JU
Sample : L4118-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D7-20200918

Quant Time: Sep 28 17:53:25 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 11:18:41 2020
Response via : Initial Calibration

