

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
 Data File : BN011979.D
 Acq On : 28 Sep 2020 10:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 28 14:16:17 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	56	0.00
2	1,4-Dioxane	0.696	0.579	16.8	52	0.00
3	n-Nitrosodimethylamine	1.001	1.267	-26.6#	83	0.00
4 S	2-Fluorophenol	1.401	1.322	5.6	59	0.00
5 S	Phenol-d6	1.767	1.708	3.3	61	0.00
6	bis(2-Chloroethyl)ether	1.511	1.457	3.6	61	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	59	0.01
8 S	Nitrobenzene-d5	0.413	0.432	-4.6	69	0.00
9	Naphthalene	1.146	1.078	5.9	61	0.00
10	Hexachlorobutadiene	0.187	0.190	-1.6	64	0.01
11 SURR	2-Methylnaphthalene-d10	0.633	0.611	3.5	63	0.00
12	2-Methylnaphthalene	0.729	0.688	5.6	61	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	64	0.00
14 S	2,4,6-Tribromophenol	0.194	0.205	-5.7	77	0.01
15 S	2-Fluorobiphenyl	1.513	1.396	7.7	64	0.00
16	Acenaphthylene	1.928	1.810	6.1	68	0.01
17	Acenaphthene	1.320	1.215	8.0	66	0.00
18	Fluorene	1.612	1.503	6.8	67	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.070	7.9	67	0.00
21	4-Bromophenyl-phenylether	0.212	0.204	3.8	66	0.00
22	Hexachlorobenzene	0.247	0.246	0.4	67	0.00
23	Atrazine	0.203	0.207	-2.0	73	0.00
24	Pentachlorophenol	0.121	0.116	4.1	68	0.01
25	Phenanthrene	1.227	1.149	6.4	65	0.00
26	Anthracene	1.075	1.009	6.1	66	0.01
27 SURR	Fluoranthene-d10	1.174	1.071	8.8	63	0.00
28	Fluoranthene	1.371	1.274	7.1	65	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	63	0.00
30	Pyrene	1.491	1.395	6.4	64	0.00
31 S	Terphenyl-d14	0.914	0.860	5.9	64	0.00
32	Benzo(a)anthracene	1.285	1.111	13.5	60	0.00
33	Chrysene	1.383	1.334	3.5	68	0.01
34	Bis(2-ethylhexyl)phthalate	0.907	0.851	6.2	68	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.510	3.8	67	0.00
36 I	Perylene-d12	1.000	1.000	0.0	67	0.00
37	Benzo(b)fluoranthene	1.368	1.190	13.0	66	0.00
38	Benzo(k)fluoranthene	1.487	1.360	8.5	65	0.00
39 C	Benzo(a)pyrene	1.261	1.129	10.5	66	0.00
40	Dibenzo(a,h)anthracene	1.305	1.190	8.8	68	0.00
41	Benzo(g,h,i)perylene	1.410	1.281	9.1	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4