

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092820\
 Data File : BN012011.D
 Acq On : 29 Sep 2020 08:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 29 09:17:36 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 11:18:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	58	0.00
2	1,4-Dioxane	0.696	0.727	-4.5	67	0.00
3	n-Nitrosodimethylamine	1.001	1.170	-16.9	79	0.00
4 S	2-Fluorophenol	1.401	1.386	1.1	64	0.00
5 S	Phenol-d6	1.767	1.758	0.5	65	0.00
6	bis(2-Chloroethyl)ether	1.511	1.458	3.5	63	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	60	0.00
8 S	Nitrobenzene-d5	0.413	0.417	-1.0	68	0.00
9	Naphthalene	1.146	1.078	5.9	62	0.00
10	Hexachlorobutadiene	0.187	0.197	-5.3	67	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.596	5.8	62	0.00
12	2-Methylnaphthalene	0.729	0.683	6.3	62	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
14 S	2,4,6-Tribromophenol	0.194	0.211	-8.8	76	0.00
15 S	2-Fluorobiphenyl	1.513	1.451	4.1	64	0.00
16	Acenaphthylene	1.928	1.803	6.5	65	0.00
17	Acenaphthene	1.320	1.228	7.0	64	0.00
18	Fluorene	1.612	1.480	8.2	62	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.076	0.0	68	0.00
21	4-Bromophenyl-phenylether	0.212	0.212	0.0	64	0.00
22	Hexachlorobenzene	0.247	0.253	-2.4	65	0.00
23	Atrazine	0.203	0.205	-1.0	68	0.00
24	Pentachlorophenol	0.121	0.113	6.6	62	0.00
25	Phenanthrene	1.227	1.124	8.4	60	0.00
26	Anthracene	1.075	1.016	5.5	62	0.00
27 SURR	Fluoranthene-d10	1.174	1.096	6.6	61	0.00
28	Fluoranthene	1.371	1.251	8.8	60	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	58	0.00
30	Pyrene	1.491	1.406	5.7	59	0.00
31 S	Terphenyl-d14	0.914	0.888	2.8	61	0.00
32	Benzo(a)anthracene	1.285	1.143	11.1	57	0.00
33	Chrysene	1.383	1.312	5.1	61	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.856	5.6	62	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.559	0.7	63	0.00
36 I	Perylene-d12	1.000	1.000	0.0	62	0.00
37	Benzo(b)fluoranthene	1.368	1.192	12.9	61	0.00
38	Benzo(k)fluoranthene	1.487	1.315	11.6	58	0.00
39 C	Benzo(a)pyrene	1.261	1.145	9.2	62	0.00
40	Dibenzo(a,h)anthracene	1.305	1.225	6.1	65	0.00
41	Benzo(g,h,i)perylene	1.410	1.314	6.8	63	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4