

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092920\
 Data File : BN012015.D
 Acq On : 29 Sep 2020 11:14
 Operator : CG/JU
 Sample : PB131893BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB131893BS

Quant Time: Sep 29 11:44:21 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 10:58:58 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	1337	0.40	ng	0.00
7) Naphthalene-d8	10.440	136	5573	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	2973	0.40	ng	0.00
19) Phenanthrene-d10	17.053	188	6013	0.40	ng	0.00
29) Chrysene-d12	21.248	240	5359	0.40	ng	# 0.00
36) Perylene-d12	23.460	264	5771	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.275	112	1803	0.38	ng	0.00
5) Phenol-d6	6.837	99	2304	0.39	ng	0.00
8) Nitrobenzene-d5	8.818	82	2401	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	3341	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.790	330	643	0.44	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	4163	0.37	ng	0.00
27) Fluoranthene-d10	19.087	212	6432	0.36	ng	0.00
31) Terphenyl-d14	19.693	244	4697	0.38	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.245	88	865	0.37	ng	# 91
3) n-Nitrosodimethylamine	3.559	42	1673	0.50	ng	# 82
6) bis(2-Chloroethyl)ether	7.103	93	1902	0.38	ng	91
9) Naphthalene	10.491	128	5993	0.38	ng	100
10) Hexachlorobutadiene	10.782	225	1106	0.42	ng	# 98
12) 2-Methylnaphthalene	12.113	142	3833	0.38	ng	# 93
16) Acenaphthylene	14.014	152	5396	0.38	ng	100
17) Acenaphthene	14.369	154	3582	0.37	ng	91
18) Fluorene	15.359	166	4445	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.435	198	453	0.40	ng	# 12
21) 4-Bromophenyl-phenylether	16.250	248	1221	0.38	ng	# 83
22) Hexachlorobenzene	16.360	284	1469	0.40	ng	# 79
23) Atrazine	16.518	200	1228	0.40	ng	# 95
24) Pentachlorophenol	16.700	266	670	0.37	ng	99
25) Phenanthrene	17.090	178	6567	0.36	ng	99
26) Anthracene	17.175	178	5731	0.35	ng	99
28) Fluoranthene	19.117	202	7437	0.36	ng	# 97
30) Pyrene	19.479	202	7672	0.38	ng	100
32) Benzo(a)anthracene	21.237	228	6161	0.36	ng	98
33) Chrysene	21.290	228	7082	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.174	149	4625	0.38	ng	91
35) Indeno(1,2,3-cd)pyrene	25.668	276	8832	0.42	ng	# 95
37) Benzo(b)fluoranthene	22.799	252	7104	0.36	ng	# 82
38) Benzo(k)fluoranthene	22.844	252	7622	0.36	ng	# 84
39) Benzo(a)pyrene	23.367	252	6754	0.37	ng	# 78
40) Dibenzo(a,h)anthracene	25.691	278	6895	0.37	ng	# 86
41) Benzo(g,h,i)perylene	26.349	276	7765	0.38	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092920\
Data File : BN012015.D
Acq On : 29 Sep 2020 11:14
Operator : CG/JU
Sample : PB131893BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB131893BS

Quant Time: Sep 29 11:44:21 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 29 10:58:58 2020
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.675 min Scan# 583
 Delta R.T. -0.000 min
 Lab File: BN012015.D
 Acq: 29 Sep 2020 11:14




Instrument :
 BNA_N
 ClientSampleId :
 PB131893BS

Tgt Ion: 152 Resp: 1337
 Ion Ratio Lower Upper
 152 100
 150 148.2 118.3 177.5
 115 75.9 51.3 76.9



#2
 1,4-Dioxane
 Concen: 0.37 ng
 RT: 3.245 min Scan# 33
 Delta R.T. -0.008 min
 Lab File: BN012015.D
 Acq: 29 Sep 2020 11:14

Tgt Ion: 88 Resp: 865
 Ion Ratio Lower Upper
 88 100
 58 82.0 63.3 94.9
 43 53.1 33.0 49.4#








#3
n-Nitrosodimethylamine
Concen: 0.50 ng
RT: 3.559 min Scan# 72
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion: 42 Resp: 1673
Ion Ratio Lower Upper
42 100
74 64.3 64.4 96.6#
44 4.3 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.38 ng
RT: 5.275 min Scan# 285
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion: 112 Resp: 1803
Ion Ratio Lower Upper
112 100
64 73.5 49.0 73.6
63 50.9 31.8 47.6#









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.837 min Scan# 479
Delta R.T. -0.008 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion: 99 Resp: 2304
Ion Ratio Lower Upper
99 100
42 37.7 23.4 35.0#
71 55.0 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.103 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion: 93 Resp: 1902
Ion Ratio Lower Upper
93 100
63 85.9 61.1 91.7
95 30.1 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.440 min Scan# 830
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:	136	Resp:	5573
Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.5	14.3
54	11.7	5.6	8.4#
68	6.8	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.42 ng
RT: 8.818 min Scan# 702
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:	82	Resp:	2401
Ion	Ratio	Lower	Upper
82	100		
128	38.3	34.2	51.2
54	61.8	43.1	64.7









#9
Naphthalene
Concen: 0.38 ng
RT: 10.491 min Scan# 834
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS

Tgt Ion:128 Resp: 5993
Ion Ratio Lower Upper
128 100
129 11.9 9.7 14.5
127 13.6 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.782 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:225 Resp: 1106
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.038 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS

Tgt Ion:152 Resp: 3341
Ion Ratio Lower Upper
152 100
151 19.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.113 min Scan# 1049
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:142 Resp: 3833
Ion Ratio Lower Upper
142 100
141 89.0 70.0 105.0
115 55.8 35.8 53.6#





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.306 min Scan# 1279

Delta R.T. 0.000 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14





Tgt Ion:164 Resp: 2973

Ion Ratio Lower Upper

164 100

162 101.5 83.6 125.4

160 47.6 41.2 61.8



#14

2,4,6-Tribromophenol

Concen: 0.44 ng

RT: 15.790 min Scan# 1396

Delta R.T. 0.000 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14




Tgt Ion:330 Resp: 643


Ion Ratio Lower Upper

330 100

332 0.0 0.0 0.0

141 53.0 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.923 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:	172	Resp:	4163
Ion Ratio	Lower	Upper	
172	100		
171	35.9	27.4	41.2
170	24.6	17.8	26.6



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.014 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:	152	Resp:	5396
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.9	23.9
153	13.3	10.5	15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.369 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:154 Resp: 3582
Ion Ratio Lower Upper
154 100
153 115.8 83.4 125.2
152 57.3 42.8 64.2



#18
Fluorene
Concen: 0.37 ng
RT: 15.359 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:166 Resp: 4445
Ion Ratio Lower Upper
166 100
165 99.1 78.8 118.2
167 13.3 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.053 min Scan# 1499
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:188 Resp: 6013
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.5 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.435 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:198 Resp: 453
Ion Ratio Lower Upper
198 100
51 144.0 45.0 67.4#
105 61.8 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.250 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:248 Resp: 1221
Ion Ratio Lower Upper
248 100
250 102.4 82.7 124.1
141 76.9 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.360 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:284 Resp: 1469
Ion Ratio Lower Upper
284 100
142 63.1 32.4 48.6#
249 34.6 26.8 40.2









#23
 Atrazine
 Concen: 0.40 ng
 RT: 16.518 min Scan# 1455
 Delta R.T. 0.000 min
 Lab File: BN012015.D
 Acq: 29 Sep 2020 11:14




Instrument :
 BNA_N
 ClientSampleId :
 PB131893BS


Tgt Ion:200 Resp: 1228
 Ion Ratio Lower Upper
 200 100
 173 28.8 18.7 28.1#
 215 48.5 40.2 60.4



#24
 Pentachlorophenol
 Concen: 0.37 ng
 RT: 16.700 min Scan# 1470
 Delta R.T. 0.000 min
 Lab File: BN012015.D
 Acq: 29 Sep 2020 11:14

Tgt Ion:266 Resp: 670
 Ion Ratio Lower Upper
 266 100
 264 64.6 51.2 76.8
 268 67.2 52.4 78.6









#25
Phenanthrene
Concen: 0.36 ng
RT: 17.090 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:178 Resp: 6567
Ion Ratio Lower Upper
178 100
176 19.4 15.1 22.7
179 15.2 12.3 18.5



#26
Anthracene
Concen: 0.35 ng
RT: 17.175 min Scan# 1509
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:178 Resp: 5731
Ion Ratio Lower Upper
178 100
176 18.4 14.6 21.8
179 15.1 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.087 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:212 Resp: 6432
Ion Ratio Lower Upper
212 100
106 11.3 6.4 9.6#
104 7.1 3.9 5.9#



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.117 min Scan# 1800
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:202 Resp: 7437
Ion Ratio Lower Upper
202 100
101 10.0 5.4 8.2#
203 16.8 13.8 20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.248 min Scan# 2096

Delta R.T. 0.000 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14

Tgt Ion:240 Resp: 5359

Ion Ratio Lower Upper

240 100

120 13.6 3.5 5.3#





236 26.9 20.5 30.7

Instrument :

BNA_N

ClientSampleId :

PB131893BS



#30

Pyrene

Concen: 0.38 ng

RT: 19.479 min Scan# 1873

Delta R.T. 0.000 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14




Tgt Ion:202 Resp: 7672


Ion Ratio Lower Upper

202 100

200 20.5 16.5 24.7

203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.693 min Scan# 1916
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS


Tgt Ion:244 Resp: 4697
Ion Ratio Lower Upper
244 100
212 10.7 7.3 10.9
122 12.8 7.8 11.6#



#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.237 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:228 Resp: 6161
Ion Ratio Lower Upper
228 100
226 27.1 21.0 31.4
229 20.6 16.0 24.0





#33

Chrysene

Concen: 0.38 ng

RT: 21.290 min Scan# 2100

Delta R.T. 0.011 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14

Tgt Ion:228 Resp: 7082

Ion Ratio Lower Upper

228 100

226 29.9 23.3 34.9





229 21.1 15.8 23.8

Instrument :

BNA_N

ClientSampleId :

PB131893BS



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.38 ng

RT: 21.174 min Scan# 2089

Delta R.T. 0.000 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14




Tgt Ion:149 Resp: 4625


Ion Ratio Lower Upper

149 100

167 24.2 23.6 35.4

279 4.6 4.0 6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.668 min Scan# 3158
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14




Instrument :
BNA_N
ClientSampleId :
PB131893BS

Tgt Ion:276 Resp: 8832
Ion Ratio Lower Upper
276 100
138 20.2 13.0 19.4#
277 24.0 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.460 min Scan# 2513
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14


Tgt Ion:264 Resp: 5771
Ion Ratio Lower Upper
264 100
260 25.7 19.6 29.4
265 69.7 38.6 58.0#








#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.799 min Scan# 2320
Delta R.T. 0.000 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Instrument :
BNA_N
ClientSampleId :
PB131893BS






Tgt Ion:252 Resp: 7104
Ion Ratio Lower Upper
252 100
253 31.2 19.1 28.7#
125 16.7 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.36 ng
RT: 22.844 min Scan# 2333
Delta R.T. 0.003 min
Lab File: BN012015.D
Acq: 29 Sep 2020 11:14

Tgt Ion:252 Resp: 7622
Ion Ratio Lower Upper
252 100
253 31.2 19.3 28.9#
125 14.9 6.0 9.0#





#39

Benzo(a)pyrene

Concen: 0.37 ng

RT: 23.367 min Scan# 2486

Delta R.T. 0.003 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14





Tgt Ion:252	Resp:	6754
Ion Ratio	Lower	Upper
252	100	
253	34.2	19.8 29.6#
125	20.4	7.3 10.9#

Instrument :

BNA_N

ClientSampleId :

PB131893BS



#40

Dibenzo(a,h)anthracene

Concen: 0.37 ng




RT: 25.691 min Scan# 3165


Delta R.T. 0.007 min

Lab File: BN012015.D

Acq: 29 Sep 2020 11:14

Tgt Ion:278	Resp:	6895
Ion Ratio	Lower	Upper
278	100	
139	15.7	8.6 13.0#
279	31.6	19.8 29.8#





#41
 Benzo(g,h,i)perylene
 Concen: 0.38 ng
 RT: 26.349 min Scan# 3357
 Delta R.T. 0.003 min
 Lab File: BN012015.D
 Acq: 29 Sep 2020 11:14

Instrument :
 BNA_N
 ClientSampleId :
 PB131893BS

Tgt Ion:	276	Resp:	7765
Ion Ratio	Lower	Upper	
276	100		
277	25.5	19.4	29.0
138	20.6	11.3	16.9#

