

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092920\
 Data File : BN012027.D
 Acq On : 29 Sep 2020 19:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 19:43:44 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 10:58:58 2020
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	1441	0.40	ng	0.00
7) Naphthalene-d8	10.440	136	5952	0.40	ng	# 0.00
13) Acenaphthene-d10	14.306	164	3107	0.40	ng	0.00
19) Phenanthrene-d10	17.041	188	6116	0.40	ng	#-0.01
29) Chrysene-d12	21.248	240	5548	0.40	ng	# 0.00
36) Perylene-d12	23.460	264	6063	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.275	112	1955	0.39	ng	0.00
5) Phenol-d6	6.845	99	2473	0.39	ng	0.00
8) Nitrobenzene-d5	8.818	82	2471	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	3490	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	684	0.45	ng	0.00
15) 2-Fluorobiphenyl	12.923	172	4420	0.38	ng	0.00
27) Fluoranthene-d10	19.082	212	6662	0.37	ng	0.00
31) Terphenyl-d14	19.688	244	4925	0.39	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.245	88	860	0.34	ng	# 87
3) n-Nitrosodimethylamine	3.560	42	1793	0.50	ng	# 86
6) bis(2-Chloroethyl)ether	7.103	93	1958	0.36	ng	93
9) Naphthalene	10.491	128	6330	0.37	ng	99
10) Hexachlorobutadiene	10.782	225	1208	0.43	ng	# 99
12) 2-Methylnaphthalene	12.113	142	3978	0.37	ng	# 93
16) Acenaphthylene	14.014	152	5677	0.38	ng	99
17) Acenaphthene	14.357	154	3753	0.37	ng	92
18) Fluorene	15.359	166	4622	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.435	198	464	0.40	ng	# 13
21) 4-Bromophenyl-phenylether	16.250	248	1282	0.40	ng	# 85
22) Hexachlorobenzene	16.360	284	1534	0.41	ng	# 79
23) Atrazine	16.518	200	1220	0.39	ng	# 95
24) Pentachlorophenol	16.701	266	733	0.40	ng	97
25) Phenanthrene	17.090	178	6732	0.36	ng	100
26) Anthracene	17.175	178	5835	0.36	ng	99
28) Fluoranthene	19.112	202	7537	0.36	ng	# 97
30) Pyrene	19.479	202	7762	0.38	ng	100
32) Benzo(a)anthracene	21.227	228	6411	0.36	ng	98
33) Chrysene	21.280	228	7315	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.174	149	5040	0.40	ng	# 90
35) Indeno(1,2,3-cd)pyrene	25.664	276	8905	0.41	ng	97
37) Benzo(b)fluoranthene	22.796	252	7068	0.34	ng	# 77
38) Benzo(k)fluoranthene	22.837	252	7904	0.35	ng	# 85
39) Benzo(a)pyrene	23.361	252	6961	0.36	ng	# 68
40) Dibenzo(a,h)anthracene	25.685	278	7412	0.37	ng	# 86
41) Benzo(g,h,i)perylene	26.342	276	8060	0.38	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092920\
 Data File : BN012027.D
 Acq On : 29 Sep 2020 19:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 19:43:44 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 10:58:58 2020
 Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.675 min Scan# 583
 Delta R.T. -0.000 min
 Lab File: BN012027.D
 Acq: 29 Sep 2020 19:13




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:152 Resp: 1441
 Ion Ratio Lower Upper
 152 100
 150 150.0 118.3 177.5
 115 74.7 51.3 76.9



#2
 1,4-Dioxane
 Concen: 0.34 ng
 RT: 3.245 min Scan# 33
 Delta R.T. -0.008 min
 Lab File: BN012027.D
 Acq: 29 Sep 2020 19:13

Tgt Ion: 88 Resp: 860
 Ion Ratio Lower Upper
 88 100
 58 84.5 63.3 94.9
 43 56.6 33.0 49.4#








#3
n-Nitrosodimethylamine
Concen: 0.50 ng
RT: 3.560 min Scan# 72
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 42 Resp: 1793
Ion Ratio Lower Upper
42 100
74 67.5 64.4 96.6
44 3.7 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.275 min Scan# 285
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion: 112 Resp: 1955
Ion Ratio Lower Upper
112 100
64 71.3 49.0 73.6
63 48.5 31.8 47.6#









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.845 min Scan# 480
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	2473
Ion	Ratio	Lower	Upper
99	100		
42	37.3	23.4	35.0#
71	53.8	38.1	57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.103 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:	93	Resp:	1958
Ion	Ratio	Lower	Upper
93	100		
63	83.4	61.1	91.7
95	30.3	25.5	38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.440 min Scan# 830
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	5952
Ion Ratio	Lower	Upper	
136	100		
137	12.4	9.5	14.3
54	11.7	5.6	8.4#
68	6.1	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.818 min Scan# 702
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:	82	Resp:	2471
Ion Ratio	Lower	Upper	
82	100		
128	37.6	34.2	51.2
54	64.1	43.1	64.7









#9
Naphthalene
Concen: 0.37 ng
RT: 10.491 min Scan# 834
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	128	Resp:	6330
Ion Ratio	Lower	Upper	
128	100		
129	11.9	9.7	14.5
127	14.2	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.43 ng
RT: 10.782 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:	225	Resp:	1208
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.7	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.038 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 3490
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.113 min Scan# 1049
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:142 Resp: 3978
Ion Ratio Lower Upper
142 100
141 89.1 70.0 105.0
115 56.6 35.8 53.6#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.306 min Scan# 1279
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 3107
Ion Ratio Lower Upper
164 100
162 100.9 83.6 125.4
160 46.1 41.2 61.8



#14
2,4,6-Tribromophenol
Concen: 0.45 ng
RT: 15.791 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:330 Resp: 684
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 52.0 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.923 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	4420
Ion Ratio	Lower	Upper	
172	100		
171	34.8	27.4	41.2
170	24.1	17.8	26.6



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.014 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:	152	Resp:	5677
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.9	23.9
153	12.7	10.5	15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.357 min Scan# 1283
Delta R.T. -0.013 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 3753
Ion Ratio Lower Upper
154 100
153 114.1 83.4 125.2
152 56.3 42.8 64.2



#18
Fluorene
Concen: 0.37 ng
RT: 15.359 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:166 Resp: 4622
Ion Ratio Lower Upper
166 100
165 98.9 78.8 118.2
167 13.6 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.041 min Scan# 1498
Delta R.T. -0.012 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 6116
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.9 7.7 11.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.435 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:198 Resp: 464
Ion Ratio Lower Upper
198 100
51 139.4 45.0 67.4#
105 66.0 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.250 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1282
Ion Ratio Lower Upper
248 100
250 103.0 82.7 124.1
141 74.9 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.41 ng
RT: 16.360 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:284 Resp: 1534
Ion Ratio Lower Upper
284 100
142 64.1 32.4 48.6#
249 32.7 26.8 40.2





#23

Atrazine

Concen: 0.39 ng

RT: 16.518 min Scan# 1455

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	200	Resp:	1220
Ion Ratio	Lower	Upper	
200	100		
173	29.2	18.7	28.1#
215	51.9	40.2	60.4



#24

Pentachlorophenol

Concen: 0.40 ng




RT: 16.701 min Scan# 1470


Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13

Tgt Ion:	266	Resp:	733
Ion Ratio	Lower	Upper	
266	100		
264	60.0	51.2	76.8
268	67.0	52.4	78.6





#25

Phenanthrene

Concen: 0.36 ng

RT: 17.090 min Scan# 1502

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:178	Resp:	6732
Ion Ratio	Lower	Upper
178	100	
176	19.0	15.1 22.7
179	15.1	12.3 18.5



#26

Anthracene

Concen: 0.36 ng




RT: 17.175 min Scan# 1509


Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13

Tgt Ion:178	Resp:	5835
Ion Ratio	Lower	Upper
178	100	
176	18.4	14.6 21.8
179	15.8	12.4 18.6









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.082 min Scan# 1793
Delta R.T. -0.005 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 6662
Ion Ratio Lower Upper
212 100
106 10.4 6.4 9.6#
104 6.3 3.9 5.9#



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.112 min Scan# 1799
Delta R.T. -0.005 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:202 Resp: 7537
Ion Ratio Lower Upper
202 100
101 9.5 5.4 8.2#
203 16.8 13.8 20.6





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.248 min Scan# 2096

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13

Tgt Ion:240 Resp: 5548

Ion Ratio Lower Upper

240 100

120 11.5 3.5 5.3#





236 27.0 20.5 30.7

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#30

Pyrene

Concen: 0.38 ng

RT: 19.479 min Scan# 1873

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13




Tgt Ion:202 Resp: 7762


Ion Ratio Lower Upper

202 100

200 20.4 16.5 24.7

203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.39 ng
RT: 19.688 min Scan# 1915
Delta R.T. -0.005 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 4925
Ion Ratio Lower Upper
244 100
212 10.9 7.3 10.9#
122 13.8 7.8 11.6#



#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.227 min Scan# 2094
Delta R.T. -0.010 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:228 Resp: 6411
Ion Ratio Lower Upper
228 100
226 27.6 21.0 31.4
229 19.6 16.0 24.0





#33

Chrysene

Concen: 0.38 ng

RT: 21.280 min Scan# 2099

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13

Tgt Ion:228 Resp: 7315

Ion Ratio Lower Upper

228 100

226 30.1 23.3 34.9





229 20.2 15.8 23.8

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.40 ng

RT: 21.174 min Scan# 2089

Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13




Tgt Ion:149 Resp: 5040


Ion Ratio Lower Upper

149 100

167 23.3 23.6 35.4#

279 4.9 4.0 6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.41 ng
RT: 25.664 min Scan# 3157
Delta R.T. -0.003 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:276 Resp: 8905
Ion Ratio Lower Upper
276 100
138 18.9 13.0 19.4
277 24.2 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.460 min Scan# 2513
Delta R.T. 0.000 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13


Tgt Ion:264 Resp: 6063
Ion Ratio Lower Upper
264 100
260 24.6 19.6 29.4
265 67.1 38.6 58.0#







#37
Benzo(b)fluoranthene
Concen: 0.34 ng
RT: 22.796 min Scan# 2319
Delta R.T. -0.003 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC






Tgt Ion:252 Resp: 7068
Ion Ratio Lower Upper
252 100
253 33.0 19.1 28.7#
125 20.1 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.837 min Scan# 2331
Delta R.T. -0.003 min
Lab File: BN012027.D
Acq: 29 Sep 2020 19:13

Tgt Ion:252 Resp: 7904
Ion Ratio Lower Upper
252 100
253 31.2 19.3 28.9#
125 14.0 6.0 9.0#





#39

Benzo(a)pyrene

Concen: 0.36 ng

RT: 23.361 min Scan# 2484

Delta R.T. -0.003 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13





Tgt Ion:252	Resp:	6961
Ion Ratio	Lower	Upper
252	100	
253	36.1	19.8 29.6#
125	29.9	7.3 10.9#

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#40

Dibenzo(a,h)anthracene

Concen: 0.37 ng




RT: 25.685 min Scan# 3163


Delta R.T. 0.000 min

Lab File: BN012027.D

Acq: 29 Sep 2020 19:13

Tgt Ion:278	Resp:	7412
Ion Ratio	Lower	Upper
278	100	
139	16.1	8.6 13.0#
279	32.2	19.8 29.8#





#41
 Benzo(g,h,i)perylene
 Concen: 0.38 ng
 RT: 26.342 min Scan# 3355
 Delta R.T. -0.003 min
 Lab File: BN012027.D
 Acq: 29 Sep 2020 19:13

Instrument :
 BNA_N
ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	8060
Ion Ratio	Lower	Upper	
276	100		
277	26.2	19.4	29.0
138	19.5	11.3	16.9#

