

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092920\
 Data File : BN012013.D
 Acq On : 29 Sep 2020 09:58
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 29 11:01:19 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 10:58:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	49#	0.00
2	1,4-Dioxane	0.696	0.608	12.6	48#	0.00
3	n-Nitrosodimethylamine	1.001	1.270	-26.9#	73	0.00
4 S	2-Fluorophenol	1.401	1.374	1.9	53	0.00
5 S	Phenol-d6	1.767	1.746	1.2	54	0.00
6	bis(2-Chloroethyl)ether	1.511	1.416	6.3	52	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	51	0.00
8 S	Nitrobenzene-d5	0.413	0.427	-3.4	60	0.00
9	Naphthalene	1.146	1.081	5.7	53	0.00
10	Hexachlorobutadiene	0.187	0.194	-3.7	57	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.612	3.3	55	0.00
12	2-Methylnaphthalene	0.729	0.694	4.8	54	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	56	0.00
14 S	2,4,6-Tribromophenol	0.194	0.208	-7.2	68	0.00
15 S	2-Fluorobiphenyl	1.513	1.410	6.8	56	0.00
16	Acenaphthylene	1.928	1.803	6.5	59	0.00
17	Acenaphthene	1.320	1.205	8.7	57	0.00
18	Fluorene	1.612	1.475	8.5	57	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	54	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.076	0.0	63	0.00
21	4-Bromophenyl-phenylether	0.212	0.209	1.4	59	0.00
22	Hexachlorobenzene	0.247	0.257	-4.0	61	0.00
23	Atrazine	0.203	0.215	-5.9	66	0.00
24	Pentachlorophenol	0.121	0.114	5.8	58	0.00
25	Phenanthrene	1.227	1.130	7.9	56	0.00
26	Anthracene	1.075	0.995	7.4	57	0.00
27 SURR	Fluoranthene-d10	1.174	1.096	6.6	56	0.00
28	Fluoranthene	1.371	1.252	8.7	55	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	54	0.00
30	Pyrene	1.491	1.417	5.0	55	0.00
31 S	Terphenyl-d14	0.914	0.886	3.1	57	0.00
32	Benzo(a)anthracene	1.285	1.167	9.2	54	0.00
33	Chrysene	1.383	1.313	5.1	57	0.00
34	Bis(2-ethylhexyl)phthalate	0.907	0.879	3.1	60	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.675	-6.7	63	0.00
36 I	Perylene-d12	1.000	1.000	0.0	60	0.00
37	Benzo(b)fluoranthene	1.368	1.225	10.5	60	0.00
38	Benzo(k)fluoranthene	1.487	1.349	9.3	58	0.00
39 C	Benzo(a)pyrene	1.261	1.154	8.5	60	0.00
40	Dibenzo(a,h)anthracene	1.305	1.264	3.1	65	0.00
41	Benzo(g,h,i)perylene	1.410	1.336	5.2	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4