

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092921\
 Data File : BN016571.D
 Acq On : 29 Sep 2021 10:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 29 11:39:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 10:56:17 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.320	20.0	85	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	100	0.00
4 S	2-Fluorophenol	0.400	0.349	12.8	98	0.00
5 S	Phenol-d6	0.400	0.343	14.2	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.324	19.0	109	0.00
9	Naphthalene	0.400	0.383	4.3	104	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	106	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.299	25.3#	98	0.00
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	107	0.00
16	Acenaphthylene	0.400	0.327	18.3	98	0.00
17	Acenaphthene	0.400	0.375	6.3	106	0.00
18	Fluorene	0.400	0.368	8.0	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.415	-3.7	176	0.00
21	4-Bromophenyl-phenylether	0.400	0.367	8.3	107	0.00
22	Hexachlorobenzene	0.400	0.386	3.5	108	0.00
23	Atrazine	0.400	0.334	16.5	100	0.00
24	Pentachlorophenol	0.400	0.392	2.0	130	0.00
25	Phenanthrene	0.400	0.390	2.5	110	0.00
26	Anthracene	0.400	0.347	13.3	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.357	10.8	102	0.00
28	Fluoranthene	0.400	0.377	5.8	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
30	Pyrene	0.400	0.394	1.5	106	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	113	0.00
32	Benzo(a)anthracene	0.400	0.352	12.0	99	0.00
33	Chrysene	0.400	0.422	-5.5	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.322	19.5	92	0.00
35 I	Perylene-d12	0.400	0.400	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	118	0.00
37	Benzo(b)fluoranthene	0.400	0.389	2.8	109	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	114	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	105	0.00
40	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	121	0.00
41	Benzo(g,h,i)perylene	0.400	0.394	1.5	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4